

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031425\
 Data File : PP070502.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Mar 2025 08:31
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 14 13:24:27 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 12 02:42:43 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.520	3.826	80389632	52888592	51.685	52.768
2) SA Decachlor...	10.242	8.873	57005404	48081671	53.660	49.145
Target Compounds						
21) L5 AR-1248-1	5.672	4.912	16639028	12130239	487.149	516.718
22) L5 AR-1248-2	5.943	5.150	22777776	15670645	536.613	500.604
23) L5 AR-1248-3	6.146	5.192	23742872	16463035	514.487	518.985
24) L5 AR-1248-4	6.545	5.365	29340319	20023814	505.944	513.785
25) L5 AR-1248-5	6.584	5.759	27257454	19050910	507.707	500.199



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031425\
Data File : PP070502.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Mar 2025 08:31
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 14 13:24:27 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031125.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 12 02:42:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.520 min
Delta R.T.: -0.004 min
Response: 80389632
Conc: 51.68 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.826 min
Delta R.T.: -0.003 min
Response: 52888592
Conc: 52.77 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.242 min
Delta R.T.: -0.010 min
Response: 57005404
Conc: 53.66 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.873 min
Delta R.T.: -0.008 min
Response: 48081671
Conc: 49.15 ng/ml



#21 AR-1248-1


R.T.: 5.672 min
Delta R.T.: -0.004 min
Response: 16639028
Conc: 487.15 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.912 min
Delta R.T.: -0.002 min
Response: 12130239
Conc: 516.72 ng/ml




#22 AR-1248-2

R.T.: 5.943 min
Delta R.T.: -0.005 min
Response: 22777776
Conc: 536.61 ng/ml



#22 AR-1248-2


R.T.: 5.150 min
Delta R.T.: -0.003 min
Response: 15670645
Conc: 500.60 ng/ml



#23 AR-1248-3


R.T.: 6.146 min
Delta R.T.: -0.005 min
Response: 23742872
Conc: 514.49 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.192 min
Delta R.T.: -0.003 min
Response: 16463035
Conc: 518.99 ng/ml




#24 AR-1248-4

R.T.: 6.545 min
Delta R.T.: -0.004 min
Response: 29340319
Conc: 505.94 ng/ml



#24 AR-1248-4


R.T.: 5.365 min
Delta R.T.: -0.003 min
Response: 20023814
Conc: 513.78 ng/ml



#25 AR-1248-5

R.T.: 6.584 min
Delta R.T.: -0.005 min
Response: 27257454
Conc: 507.71 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.759 min
Delta R.T.: -0.004 min
Response: 19050910
Conc: 500.20 ng/ml