

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031425\
 Data File : PP070516.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Mar 2025 12:52
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 14 13:51:01 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 12 02:42:43 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.523	3.826	77032070	49870587	49.526	49.757
2) SA Decachlor...	10.242	8.872	52905219	46714089	49.801	47.747
Target Compounds						
16) L4 AR-1242-1	5.674	4.913	20417662	14986642	459.355	489.577
17) L4 AR-1242-2	5.696	4.931	29824821	21049444	485.898	482.812
18) L4 AR-1242-3	5.759	5.109	18882004	11807931	508.932	495.459
19) L4 AR-1242-4	5.856	5.193	15444195	11157633	480.632	506.187
20) L4 AR-1242-5	6.586	5.717	16541628	14536670	532.784	520.634



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031425\
Data File : PP070516.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Mar 2025 12:52
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 14 13:51:01 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031125.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 12 02:42:43 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.523 min
Delta R.T.: 0.000 min
Response: 77032070
Conc: 49.53 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.826 min
Delta R.T.: -0.002 min
Response: 49870587
Conc: 49.76 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.242 min
Delta R.T.: -0.009 min
Response: 52905219
Conc: 49.80 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.872 min
Delta R.T.: -0.009 min
Response: 46714089
Conc: 47.75 ng/ml



#16 AR-1242-1


R.T.: 5.674 min
Delta R.T.: -0.004 min
Response: 20417662
Conc: 459.36 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.913 min
Delta R.T.: -0.004 min
Response: 14986642
Conc: 489.58 ng/ml




#17 AR-1242-2

R.T.: 5.696 min
Delta R.T.: -0.004 min
Response: 29824821
Conc: 485.90 ng/ml



#17 AR-1242-2


R.T.: 4.931 min
Delta R.T.: -0.004 min
Response: 21049444
Conc: 482.81 ng/ml



#18 AR-1242-3


R.T.: 5.759 min
Delta R.T.: -0.003 min
Response: 18882004
Conc: 508.93 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 5.109 min
Delta R.T.: -0.004 min
Response: 11807931
Conc: 495.46 ng/ml




#19 AR-1242-4

R.T.: 5.856 min
Delta R.T.: -0.003 min
Response: 15444195
Conc: 480.63 ng/ml



#19 AR-1242-4


R.T.: 5.193 min
Delta R.T.: -0.004 min
Response: 11157633
Conc: 506.19 ng/ml



#20 AR-1242-5

R.T.: 6.586 min
Delta R.T.: -0.004 min
Response: 16541628
Conc: 532.78 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.717 min
Delta R.T.: -0.005 min
Response: 14536670
Conc: 520.63 ng/ml