

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031519\
 Data File : PP029347.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 16 Mar 2019 00:15
 Operator : EI\MA
 Sample : K1087-40
 Misc : hexane
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 TP-20

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 16 01:06:59 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP022119.M
 Quant Title : 8080.M
 QLast Update : Fri Feb 22 02:24:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.930	12.124	23550098	30046153	783.531	791.678
Target Compounds							
3) T	4-Nitroph...	11.511	0.000	1474458	0	21.524	N.D. #
5) T	DICAMBA	12.330	0.000	177492	0	1.335	N.D. #
8) T	DICHLORPROP	13.444	13.551	600779	714017	19.471	21.119
9) T	2,4-D	13.780	14.037	971200	720137	24.771	15.882 #
10) T	Pentachlo...	14.207	14.605	2864176	3228363	4.395	6.273 #
11) T	2,4,5-TP ...	0.000	15.152	0	5375053	N.D.	26.425 #
12) T	2,4,5-T	0.000	15.869	0	12815946	N.D.	77.404 #
13) T	2,4-DB	16.240	16.452	18236227	28592228	954.660	1269.266 #
14) T	DINOSEB	17.707	16.913	2042994	26251521	14.029	200.283 #
15) T	Picloram	17.483	0.000	4165043	0	14.979	N.D. #
16) T	DCPA	0.000	18.151	0	20528699	N.D.	88.335 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031519\
Data File : PP029347.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 16 Mar 2019 00:15
Operator : EI\MA
Sample : K1087-40
Misc : hexane
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
TP-20

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 16 01:06:59 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP022119.M
Quant Title : 8080.M
QLast Update : Fri Feb 22 02:24:04 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#3 4-Nitrophenol


R.T.: 11.511 min
Delta R.T.: 0.007 min
Response: 1474458
Conc: 21.52 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-20




#3 4-Nitrophenol

R.T.: 0.000 min
Exp R.T.: 11.273 min
Response: 0
Conc: N.D.




#4 2,4-DCAA

R.T.: 11.930 min
Delta R.T.: -0.011 min
Response: 23550098
Conc: 783.53 ng/ml



#4 2,4-DCAA


R.T.: 12.124 min
Delta R.T.: -0.004 min
Response: 30046153
Conc: 791.68 ng/ml



#5 DICAMBA


R.T.: 12.330 min
Delta R.T.: 0.068 min
Response: 177492
Conc: 1.33 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-20




#5 DICAMBA

R.T.: 0.000 min
Exp R.T.: 12.441 min
Response: 0
Conc: N.D.




#8 DICHLORPROP

R.T.: 13.444 min
Delta R.T.: -0.008 min
Response: 600779
Conc: 19.47 ng/ml



#8 DICHLORPROP


R.T.: 13.551 min
Delta R.T.: 0.000 min
Response: 714017
Conc: 21.12 ng/ml



#9 2,4-D


R.T.: 13.780 min
Delta R.T.: -0.020 min
Response: 971200
Conc: 24.77 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-20




#9 2,4-D

R.T.: 14.037 min
Delta R.T.: -0.002 min
Response: 720137
Conc: 15.88 ng/ml




#10 Pentachlorophenol

R.T.: 14.207 min
Delta R.T.: -0.010 min
Response: 2864176
Conc: 4.40 ng/ml



#10 Pentachlorophenol


R.T.: 14.605 min
Delta R.T.: -0.040 min
Response: 3228363
Conc: 6.27 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 0.000 min
Exp R.T.: 15.075 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :
TP-20




#11 2,4,5-TP (SILVEX)

R.T.: 15.152 min
Delta R.T.: -0.064 min
Response: 5375053
Conc: 26.43 ng/ml




#12 2,4,5-T

R.T.: 0.000 min
Exp R.T.: 15.464 min
Response: 0
Conc: N.D.



#12 2,4,5-T


R.T.: 15.869 min
Delta R.T.: 0.005 min
Response: 12815946
Conc: 77.40 ng/ml



#13 2,4-DB


R.T.: 16.240 min
Delta R.T.: 0.017 min
Response: 18236227
Conc: 954.66 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-20




#13 2,4-DB

R.T.: 16.452 min
Delta R.T.: -0.008 min
Response: 28592228
Conc: 1269.27 ng/ml




#14 DINOSEB

R.T.: 17.707 min
Delta R.T.: 0.021 min
Response: 2042994
Conc: 14.03 ng/ml



#14 DINOSEB


R.T.: 16.913 min
Delta R.T.: 0.064 min
Response: 26251521
Conc: 200.28 ng/ml



#15 Picloram


R.T.: 17.483 min
Delta R.T.: 0.008 min
Response: 4165043
Conc: 14.98 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-20




#15 Picloram

R.T.: 0.000 min
Exp R.T. : 18.245 min
Response: 0
Conc: N.D.



#16 DCPA

R.T.: 0.000 min
Exp R.T. : 18.069 min
Response: 0
Conc: N.D.



#16 DCPA

R.T.: 18.151 min
Delta R.T.: 0.033 min
Response: 20528699
Conc: 88.33 ng/ml