

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031522\
 Data File : PP044746.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Mar 2022 17:58
 Operator : YP\AJ
 Sample : N1942-05
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 16 08:01:29 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 03 02:21:53 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.906	3.162	31238496	27402126	15.077	15.663
2)	SA Decachlor...	10.035	8.520	24581466	27424756	17.860	17.878
Target Compounds							
3)	L1 AR-1016-1	5.170	4.320	147759	128810	2.176	2.317
4)	L1 AR-1016-2	5.187	4.324	57428	85115	0.567	1.098 #
5)	L1 AR-1016-3	5.259	4.499f	61431	617798	0.979	14.402 #
6)	L1 AR-1016-4	5.366	4.553	18689	484641	0.360	14.379 #
7)	L1 AR-1016-5	5.682	4.790	246130	1444992	4.887	34.739 #
8)	L2 AR-1221-1	4.125	3.384	27567287	8704	1115.650	0.379 #
9)	L2 AR-1221-2	4.231	3.455f	502827	26042821	26.800	1507.794 #
10)	L2 AR-1221-3	4.312	3.562	394067	575939	7.043	11.582 #
11)	L3 AR-1232-1	4.312	3.562	394067	575939	8.587	13.859 #
12)	L3 AR-1232-2	4.872	4.324	275399	85115	12.238	2.464 #
13)	L3 AR-1232-3	5.187	4.499f	57428	617798	1.327	33.113 #
14)	L3 AR-1232-4	5.366	4.617	18689	1521917	0.931	92.299 #
15)	L3 AR-1232-5	5.456	4.790	83527	1444992	5.505	79.822 #
16)	L4 AR-1242-1	5.170	4.320	147759	128810	2.715	2.809
17)	L4 AR-1242-2	5.187	4.324	57428	85115	0.721	1.347 #
18)	L4 AR-1242-3	5.259	4.499f	61431	617798	1.301	17.296 #
19)	L4 AR-1242-4	5.366	4.617	18689	1521917	0.483	43.991 #
20)	L4 AR-1242-5	6.158	5.160	879496	967368	20.629	22.903
21)	L5 AR-1248-1	5.170	4.320	147759	128810	3.574	3.632
22)	L5 AR-1248-2	5.456	4.553	83527	484641	1.489	10.298 #
23)	L5 AR-1248-3	5.682	4.617	246130	1521917	3.604	30.881 #
24)	L5 AR-1248-4	6.113	4.790	783263	1444992	10.741	25.566 #
25)	L5 AR-1248-5	6.158	5.213	879496	353813	11.974	6.308 #
26)	L6 AR-1254-1	6.113f	5.160	783263	967368	10.444	11.701
27)	L6 AR-1254-2	6.326	5.336	255446	260583	2.273	3.622 #
28)	L6 AR-1254-3	6.742	5.746	555558	129505	4.748	1.124 #
29)	L6 AR-1254-4	7.050	6.024f	56458	101689	0.702	1.351 #
30)	L6 AR-1254-5	7.506	6.457	319989	31971	3.676	0.293 #
31)	L7 AR-1260-1	6.907	5.904	23678	48020	0.269	0.599 #
32)	L7 AR-1260-2	7.193	6.111	50048	55817	0.521	0.567
33)	L7 AR-1260-3	7.591	6.264	60854	46539	0.938	0.502 #
34)	L7 AR-1260-4	7.821f	6.789	1029441	5806	13.199	0.081 #
35)	L7 AR-1260-5	8.174	7.058	366751	9770	2.617	0.059 #
36)	L8 AR-1262-1	7.591	6.501	60854	131787	0.602	1.270 #
37)	L8 AR-1262-2	8.174	6.789	366751	5806	2.209	0.061 #
38)	L8 AR-1262-3	8.506	7.350	57338	200944	0.512	2.547 #
39)	L8 AR-1262-4	8.604	7.426	143005	27297	2.881	0.189 #
40)	L8 AR-1262-5	9.266f	7.969	41055	39155	0.719	0.547
41)	L9 AR-1268-1	8.506	7.363	57338	39827	0.294	0.175 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031522\
 Data File : PP044746.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Mar 2022 17:58
 Operator : YP\AJ
 Sample : N1942-05
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 16 08:01:29 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 03 02:21:53 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.604	7.426	143005	27297	0.785	0.133 #
43) L9	AR-1268-3	8.832	7.651	66958	8877	0.434	0.051 #
44) L9	AR-1268-4	9.266f	7.969	41055	39155	0.656	0.483 #
45) L9	AR-1268-5	9.713	8.276	68408	11245	0.146	0.021 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031522\
Data File : PP044746.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Mar 2022 17:58
Operator : YP\AJ
Sample : N1942-05
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 16 08:01:29 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030222.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 03 02:21:53 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.906 min
Delta R.T.: -0.006 min
Response: 31238496
Conc: 15.08 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.162 min
Delta R.T.: -0.004 min
Response: 27402126
Conc: 15.66 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.035 min
Delta R.T.: -0.010 min
Response: 24581466
Conc: 17.86 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.520 min
Delta R.T.: -0.010 min
Response: 27424756
Conc: 17.88 ng/ml



#3 AR-1016-1


R.T.: 5.170 min
Delta R.T.: 0.005 min
Response: 147759
Conc: 2.18 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 4.320 min
Delta R.T.: 0.007 min
Response: 128810
Conc: 2.32 ng/ml




#4 AR-1016-2

R.T.: 5.187 min
Delta R.T.: -0.002 min
Response: 57428
Conc: 0.57 ng/ml



#4 AR-1016-2


R.T.: 4.324 min
Delta R.T.: -0.007 min
Response: 85115
Conc: 1.10 ng/ml



#5 AR-1016-3


R.T.: 5.259 min
Delta R.T.: 0.004 min
Response: 61431
Conc: 0.98 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 4.499 min
Delta R.T.: -0.017 min
Response: 617798
Conc: 14.40 ng/ml




#6 AR-1016-4

R.T.: 5.366 min
Delta R.T.: 0.006 min
Response: 18689
Conc: 0.36 ng/ml



#6 AR-1016-4


R.T.: 4.553 min
Delta R.T.: -0.012 min
Response: 484641
Conc: 14.38 ng/ml



#7 AR-1016-5


R.T.: 5.682 min
Delta R.T.: 0.001 min
Response: 246130
Conc: 4.89 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 4.790 min
Delta R.T.: 0.000 min
Response: 1444992
Conc: 34.74 ng/ml




#8 AR-1221-1

R.T.: 4.125 min
Delta R.T.: -0.014 min
Response: 27567287
Conc: 1115.65 ng/ml



#8 AR-1221-1


R.T.: 3.384 min
Delta R.T.: -0.010 min
Response: 8704
Conc: 0.38 ng/ml



#9 AR-1221-2


R.T.: 4.231 min
Delta R.T.: 0.000 min
Response: 502827
Conc: 26.80 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.455 min
Delta R.T.: -0.029 min
Response: 26042821
Conc: 1507.79 ng/ml




#10 AR-1221-3

R.T.: 4.312 min
Delta R.T.: 0.001 min
Response: 394067
Conc: 7.04 ng/ml



#10 AR-1221-3


R.T.: 3.562 min
Delta R.T.: -0.001 min
Response: 575939
Conc: 11.58 ng/ml



#11 AR-1232-1


R.T.: 4.312 min
Delta R.T.: 0.000 min
Response: 394067
Conc: 8.59 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 3.562 min
Delta R.T.: 0.000 min
Response: 575939
Conc: 13.86 ng/ml




#12 AR-1232-2

R.T.: 4.872 min
Delta R.T.: -0.003 min
Response: 275399
Conc: 12.24 ng/ml



#12 AR-1232-2


R.T.: 4.324 min
Delta R.T.: -0.007 min
Response: 85115
Conc: 2.46 ng/ml



#13 AR-1232-3


R.T.: 5.187 min
Delta R.T.: -0.002 min
Response: 57428
Conc: 1.33 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 4.499 min
Delta R.T.: -0.017 min
Response: 617798
Conc: 33.11 ng/ml




#14 AR-1232-4

R.T.: 5.366 min
Delta R.T.: 0.005 min
Response: 18689
Conc: 0.93 ng/ml



#14 AR-1232-4


R.T.: 4.617 min
Delta R.T.: 0.009 min
Response: 1521917
Conc: 92.30 ng/ml



#15 AR-1232-5


R.T.: 5.456 min
Delta R.T.: -0.007 min
Response: 83527
Conc: 5.51 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 4.790 min
Delta R.T.: 0.000 min
Response: 1444992
Conc: 79.82 ng/ml




#16 AR-1242-1

R.T.: 5.170 min
Delta R.T.: 0.004 min
Response: 147759
Conc: 2.71 ng/ml



#16 AR-1242-1


R.T.: 4.320 min
Delta R.T.: 0.008 min
Response: 128810
Conc: 2.81 ng/ml



#17 AR-1242-2


R.T.: 5.187 min
Delta R.T.: -0.002 min
Response: 57428
Conc: 0.72 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 4.324 min
Delta R.T.: -0.007 min
Response: 85115
Conc: 1.35 ng/ml




#18 AR-1242-3

R.T.: 5.259 min
Delta R.T.: 0.004 min
Response: 61431
Conc: 1.30 ng/ml



#18 AR-1242-3


R.T.: 4.499 min
Delta R.T.: -0.017 min
Response: 617798
Conc: 17.30 ng/ml



#19 AR-1242-4


R.T.: 5.366 min
Delta R.T.: 0.006 min
Response: 18689
Conc: 0.48 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 4.617 min
Delta R.T.: 0.009 min
Response: 1521917
Conc: 43.99 ng/ml




#20 AR-1242-5

R.T.: 6.158 min
Delta R.T.: -0.003 min
Response: 879496
Conc: 20.63 ng/ml



#20 AR-1242-5


R.T.: 5.160 min
Delta R.T.: -0.007 min
Response: 967368
Conc: 22.90 ng/ml



#21 AR-1248-1


R.T.: 5.170 min
Delta R.T.: 0.004 min
Response: 147759
Conc: 3.57 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.320 min
Delta R.T.: 0.007 min
Response: 128810
Conc: 3.63 ng/ml




#22 AR-1248-2

R.T.: 5.456 min
Delta R.T.: -0.006 min
Response: 83527
Conc: 1.49 ng/ml



#22 AR-1248-2


R.T.: 4.553 min
Delta R.T.: -0.012 min
Response: 484641
Conc: 10.30 ng/ml



#23 AR-1248-3


R.T.: 5.682 min
Delta R.T.: 0.001 min
Response: 246130
Conc: 3.60 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 4.617 min
Delta R.T.: 0.009 min
Response: 1521917
Conc: 30.88 ng/ml




#24 AR-1248-4

R.T.: 6.113 min
Delta R.T.: -0.007 min
Response: 783263
Conc: 10.74 ng/ml



#24 AR-1248-4


R.T.: 4.790 min
Delta R.T.: 0.000 min
Response: 1444992
Conc: 25.57 ng/ml



#25 AR-1248-5


R.T.: 6.158 min
Delta R.T.: -0.003 min
Response: 879496
Conc: 11.97 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.213 min
Delta R.T.: 0.003 min
Response: 353813
Conc: 6.31 ng/ml




#26 AR-1254-1

R.T.: 6.113 min
Delta R.T.: 0.022 min
Response: 783263
Conc: 10.44 ng/ml



#26 AR-1254-1


R.T.: 5.160 min
Delta R.T.: -0.007 min
Response: 967368
Conc: 11.70 ng/ml



#27 AR-1254-2


R.T.: 6.326 min
Delta R.T.: -0.005 min
Response: 255446
Conc: 2.27 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 5.336 min
Delta R.T.: 0.008 min
Response: 260583
Conc: 3.62 ng/ml




#28 AR-1254-3

R.T.: 6.742 min
Delta R.T.: 0.011 min
Response: 555558
Conc: 4.75 ng/ml



#28 AR-1254-3


R.T.: 5.746 min
Delta R.T.: -0.013 min
Response: 129505
Conc: 1.12 ng/ml



#29 AR-1254-4


R.T.: 7.050 min
Delta R.T.: 0.005 min
Response: 56458
Conc: 0.70 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 6.024 min
Delta R.T.: 0.015 min
Response: 101689
Conc: 1.35 ng/ml




#30 AR-1254-5

R.T.: 7.506 min
Delta R.T.: 0.000 min
Response: 319989
Conc: 3.68 ng/ml



#30 AR-1254-5


R.T.: 6.457 min
Delta R.T.: -0.003 min
Response: 31971
Conc: 0.29 ng/ml



#31 AR-1260-1


R.T.: 6.907 min
Delta R.T.: -0.004 min
Response: 23678
Conc: 0.27 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 5.904 min
Delta R.T.: 0.004 min
Response: 48020
Conc: 0.60 ng/ml




#32 AR-1260-2

R.T.: 7.193 min
Delta R.T.: -0.002 min
Response: 50048
Conc: 0.52 ng/ml



#32 AR-1260-2


R.T.: 6.111 min
Delta R.T.: 0.004 min
Response: 55817
Conc: 0.57 ng/ml



#33 AR-1260-3


R.T.: 7.591 min
Delta R.T.: 0.002 min
Response: 60854
Conc: 0.94 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.264 min
Delta R.T.: -0.005 min
Response: 46539
Conc: 0.50 ng/ml




#34 AR-1260-4

R.T.: 7.821 min
Delta R.T.: -0.016 min
Response: 1029441
Conc: 13.20 ng/ml



#34 AR-1260-4


R.T.: 6.789 min
Delta R.T.: 0.006 min
Response: 5806
Conc: 0.08 ng/ml



#35 AR-1260-5


R.T.: 8.174 min
Delta R.T.: -0.006 min
Response: 366751
Conc: 2.62 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.058 min
Delta R.T.: 0.008 min
Response: 9770
Conc: 0.06 ng/ml




#36 AR-1262-1

R.T.: 7.591 min
Delta R.T.: 0.000 min
Response: 60854
Conc: 0.60 ng/ml



#36 AR-1262-1


R.T.: 6.501 min
Delta R.T.: 0.000 min
Response: 131787
Conc: 1.27 ng/ml



#37 AR-1262-2


R.T.: 8.174 min
Delta R.T.: -0.006 min
Response: 366751
Conc: 2.21 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 6.789 min
Delta R.T.: 0.007 min
Response: 5806
Conc: 0.06 ng/ml




#38 AR-1262-3

R.T.: 8.506 min
Delta R.T.: -0.006 min
Response: 57338
Conc: 0.51 ng/ml



#38 AR-1262-3


R.T.: 7.350 min
Delta R.T.: -0.006 min
Response: 200944
Conc: 2.55 ng/ml



#39 AR-1262-4


R.T.: 8.604 min
Delta R.T.: 0.000 min
Response: 143005
Conc: 2.88 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 7.426 min
Delta R.T.: 0.000 min
Response: 27297
Conc: 0.19 ng/ml




#40 AR-1262-5

R.T.: 9.266 min
Delta R.T.: -0.021 min
Response: 41055
Conc: 0.72 ng/ml



#40 AR-1262-5


R.T.: 7.969 min
Delta R.T.: 0.000 min
Response: 39155
Conc: 0.55 ng/ml



#41 AR-1268-1


R.T.: 8.506 min
Delta R.T.: -0.003 min
Response: 57338
Conc: 0.29 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.363 min
Delta R.T.: 0.006 min
Response: 39827
Conc: 0.18 ng/ml




#42 AR-1268-2

R.T.: 8.604 min
Delta R.T.: -0.004 min
Response: 143005
Conc: 0.79 ng/ml



#42 AR-1268-2


R.T.: 7.426 min
Delta R.T.: 0.000 min
Response: 27297
Conc: 0.13 ng/ml



#43 AR-1268-3


R.T.: 8.832 min
Delta R.T.: -0.013 min
Response: 66958
Conc: 0.43 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.651 min
Delta R.T.: 0.002 min
Response: 8877
Conc: 0.05 ng/ml




#44 AR-1268-4

R.T.: 9.266 min
Delta R.T.: -0.021 min
Response: 41055
Conc: 0.66 ng/ml



#44 AR-1268-4


R.T.: 7.969 min
Delta R.T.: 0.000 min
Response: 39155
Conc: 0.48 ng/ml



#45 AR-1268-5

R.T.: 9.713 min
Delta R.T.: 0.003 min
Response: 68408
Conc: 0.15 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.276 min
Delta R.T.: 0.002 min
Response: 11245
Conc: 0.02 ng/ml