

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031522\
 Data File : PP044737.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Mar 2022 15:41
 Operator : YP\AJ
 Sample : N1914-04
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 TP-2

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 16 07:58:09 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 03 02:21:53 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.906	3.162	43941484	34385222	21.207	19.655
2) SA Decachlor...	10.034	8.519	22795463	25951779	16.562	16.918
Target Compounds						
26) L6 AR-1254-1	6.081	5.155	50847079	62395841	677.980	754.725
27) L6 AR-1254-2	6.321	5.316	98907609	74175926	880.244	1031.100
28) L6 AR-1254-3	6.720	5.747	148.1E6	164.3E6	1265.453	1426.488
29) L6 AR-1254-4	7.034	5.997	115.6E6	119.9E6	1438.189	1593.300
30) L6 AR-1254-5	7.494	6.448	159.1E6	216.9E6	1827.220	1990.520



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031522\
Data File : PP044737.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Mar 2022 15:41
Operator : YP\AJ
Sample : N1914-04
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
TP-2

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 16 07:58:09 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030222.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 03 02:21:53 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.906 min
Delta R.T.: -0.006 min
Response: 43941484
Conc: 21.21 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-2




#1 Tetrachloro-m-xylene

R.T.: 3.162 min
Delta R.T.: -0.005 min
Response: 34385222
Conc: 19.65 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.034 min
Delta R.T.: -0.011 min
Response: 22795463
Conc: 16.56 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.519 min
Delta R.T.: -0.011 min
Response: 25951779
Conc: 16.92 ng/ml



#26 AR-1254-1


R.T.: 6.081 min
Delta R.T.: -0.010 min
Response: 50847079
Conc: 677.98 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-2




#26 AR-1254-1

R.T.: 5.155 min
Delta R.T.: -0.011 min
Response: 62395841
Conc: 754.73 ng/ml




#27 AR-1254-2

R.T.: 6.321 min
Delta R.T.: -0.010 min
Response: 98907609
Conc: 880.24 ng/ml



#27 AR-1254-2


R.T.: 5.316 min
Delta R.T.: -0.011 min
Response: 74175926
Conc: 1031.10 ng/ml



#28 AR-1254-3


R.T.: 6.720 min
Delta R.T.: -0.010 min
Response: 148078550
Conc: 1265.45 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-2




#28 AR-1254-3

R.T.: 5.747 min
Delta R.T.: -0.012 min
Response: 164349396
Conc: 1426.49 ng/ml




#29 AR-1254-4

R.T.: 7.034 min
Delta R.T.: -0.010 min
Response: 115632005
Conc: 1438.19 ng/ml



#29 AR-1254-4


R.T.: 5.997 min
Delta R.T.: -0.012 min
Response: 119948711
Conc: 1593.30 ng/ml



#30 AR-1254-5

R.T.: 7.494 min
Delta R.T.: -0.011 min
Response: 159051145
Conc: 1827.22 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-2



#30 AR-1254-5

R.T.: 6.448 min
Delta R.T.: -0.013 min
Response: 216927809
Conc: 1990.52 ng/ml