

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031522\
 Data File : PP044774.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Mar 2022 03:00
 Operator : YP\AJ
 Sample : N1939-05 10X
 Misc :
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 PCB-031022-SR-05

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 16 08:13:54 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030222.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 03 02:21:53 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.905	3.161	5985866	5057892	2.889	2.891
2) SA Decachlor...	10.033	8.519	3844756	4757712	2.793	3.102

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031522\
Data File : PP044774.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Mar 2022 03:00
Operator : YP\AJ
Sample : N1939-05 10X
Misc :
ALS Vial : 45 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PCB-031022-SR-05

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 16 08:13:54 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP030222.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 03 02:21:53 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.905 min
Delta R.T.: -0.007 min
Response: 5985866
Conc: 2.89 ng/ml

Instrument :
ECD_P
ClientSampleId :
PCB-031022-SR-05




#1 Tetrachloro-m-xylene

R.T.: 3.161 min
Delta R.T.: -0.006 min
Response: 5057892
Conc: 2.89 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.033 min
Delta R.T.: -0.012 min
Response: 3844756
Conc: 2.79 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.519 min
Delta R.T.: -0.011 min
Response: 4757712
Conc: 3.10 ng/ml