

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031524\
 Data File : PP063499.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Mar 2024 21:01
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 03/18/2024
 Supervised By :Ankita Jodhani 03/18/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 15 22:40:41 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP022624.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Feb 27 01:30:59 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.490	3.640	98887431	102.5E6	44.005	44.560
2) SA Decachlor...	10.364	8.683	98933631	110.1E6	51.055	53.982
Target Compounds						
21) L5 AR-1248-1	5.673	4.733	20617532	20141837	491.791	492.935m
22) L5 AR-1248-2	5.950	4.972	32152050	30655589	471.983	458.823
23) L5 AR-1248-3	6.156	5.014	34961074	32210136	465.178	461.948
24) L5 AR-1248-4	6.563	5.187	34874114	37699646	516.844	467.564
25) L5 AR-1248-5	6.602	5.581	35202021	33034781	518.183	531.737

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031524\
Data File : PP063499.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Mar 2024 21:01
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1248CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/18/2024
Supervised By :Ankita Jodhani 03/18/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 15 22:40:41 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP022624.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Feb 27 01:30:59 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.490 min
Delta R.T.: 0.000 min
Response: 98887431
Conc: 44.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/18/2024
Supervised By :Ankita Jodhani 03/18/2024




#1 Tetrachloro-m-xylene

R.T.: 3.640 min
Delta R.T.: 0.000 min
Response: 102510472
Conc: 44.56 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.364 min
Delta R.T.: 0.002 min
Response: 98933631
Conc: 51.06 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.683 min
Delta R.T.: -0.002 min
Response: 110131472
Conc: 53.98 ng/ml




#21 AR-1248-1

R.T.: 5.673 min
Delta R.T.: 0.000 min
Response: 20617532
Conc: 491.79 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/18/2024
Supervised By :Ankita Jodhani 03/18/2024




#21 AR-1248-1

R.T.: 4.733 min
Delta R.T.: -0.001 min
Response: 20141837
Conc: 492.93 ng/ml m




#22 AR-1248-2

R.T.: 5.950 min
Delta R.T.: 0.000 min
Response: 32152050
Conc: 471.98 ng/ml



#22 AR-1248-2

R.T.: 4.972 min
Delta R.T.: 0.000 min
Response: 30655589
Conc: 458.82 ng/ml




#23 AR-1248-3

R.T.: 6.156 min
Delta R.T.: 0.000 min
Response: 34961074
Conc: 465.18 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/18/2024
Supervised By :Ankita Jodhani 03/18/2024




#23 AR-1248-3

R.T.: 5.014 min
Delta R.T.: -0.001 min
Response: 32210136
Conc: 461.95 ng/ml




#24 AR-1248-4

R.T.: 6.563 min
Delta R.T.: 0.000 min
Response: 34874114
Conc: 516.84 ng/ml



#24 AR-1248-4

R.T.: 5.187 min
Delta R.T.: 0.000 min
Response: 37699646
Conc: 467.56 ng/ml




#25 AR-1248-5

R.T.: 6.602 min
Delta R.T.: 0.000 min
Response: 35202021
Conc: 518.18 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/18/2024
Supervised By :Ankita Jodhani 03/18/2024



#25 AR-1248-5

R.T.: 5.581 min
Delta R.T.: -0.001 min
Response: 33034781
Conc: 531.74 ng/ml