

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031623\
 Data File : PP056534.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Mar 2023 22:24
 Operator : YP\AJ
 Sample : AR1248ICV500
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 17 02:59:00 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Mar 17 02:58:04 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.439	3.671	107.3E6	98700311	53.060	51.339
2) SA Decachlor...	10.278	8.773	71559140	76217317	52.518	51.969
Target Compounds						
21) L5 AR-1248-1	5.614	4.773	20063419	19684785	521.374	502.935
22) L5 AR-1248-2	5.889	5.014	30866200	30976554	524.803	503.868
23) L5 AR-1248-3	6.096	5.057	33818556	32345554	524.009	508.922
24) L5 AR-1248-4	6.501	5.231	32170915	37034440	516.030	504.984
25) L5 AR-1248-5	6.540	5.629	31461858	30300313	522.611	504.560



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031623\
Data File : PP056534.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Mar 2023 22:24
Operator : YP\AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 17 02:59:00 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031623.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 17 02:58:04 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.439 min
Delta R.T.: 0.000 min
Response: 107336604
Conc: 53.06 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.671 min
Delta R.T.: 0.000 min
Response: 98700311
Conc: 51.34 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.278 min
Delta R.T.: -0.002 min
Response: 71559140
Conc: 52.52 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.773 min
Delta R.T.: -0.003 min
Response: 76217317
Conc: 51.97 ng/ml



#21 AR-1248-1


R.T.: 5.614 min
Delta R.T.: 0.000 min
Response: 20063419
Conc: 521.37 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.773 min
Delta R.T.: -0.002 min
Response: 19684785
Conc: 502.94 ng/ml




#22 AR-1248-2

R.T.: 5.889 min
Delta R.T.: 0.000 min
Response: 30866200
Conc: 524.80 ng/ml



#22 AR-1248-2


R.T.: 5.014 min
Delta R.T.: -0.001 min
Response: 30976554
Conc: 503.87 ng/ml



#23 AR-1248-3


R.T.: 6.096 min
Delta R.T.: 0.000 min
Response: 33818556
Conc: 524.01 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 5.057 min
Delta R.T.: -0.002 min
Response: 32345554
Conc: 508.92 ng/ml




#24 AR-1248-4

R.T.: 6.501 min
Delta R.T.: 0.000 min
Response: 32170915
Conc: 516.03 ng/ml



#24 AR-1248-4


R.T.: 5.231 min
Delta R.T.: -0.002 min
Response: 37034440
Conc: 504.98 ng/ml



#25 AR-1248-5

R.T.: 6.540 min
Delta R.T.: 0.000 min
Response: 31461858
Conc: 522.61 ng/ml

Instrument :
ECD_P
ClientSampleId :



#25 AR-1248-5

R.T.: 5.629 min
Delta R.T.: -0.002 min
Response: 30300313
Conc: 504.56 ng/ml