

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031819\
 Data File : PP029375.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 18 Mar 2019 7:49 pm
 Operator : EI\MA
 Sample : HSTDCCC750
 Misc : hexane
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Sohil
 3/19/2019 1:58:57 PM

Integration File signal 1: sample.E
 Integration File signal 2: events2.e
 Quant Time: Mar 19 01:50:40 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP022119.M
 Quant Title : 8080.M
 QLast Update : Fri Feb 22 02:24:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal #1 Phase : Rxi-1ms Signal #2 Phase:
 Signal #1 Info : 20M x 0.18mm x 0. Signal #2 Info :

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.927	12.125	21423311	29621169	712.771m	780.480
Target Compounds							
1) T	Dalapon	3.243	2.958	40591303	34325807	692.938	663.324
2) T	3,5-DICHL...	10.339	10.211	42152810	36126338	683.358	712.791
3) T	4-Nitroph...	11.492	11.268	46188708	16662676	674.251m	695.791m
5) T	DICAMBA	12.251	12.437	92692182	91930966	697.097	718.209
8) T	DICHLORPROP	13.441	13.547	22653332	25382690	734.173	750.772m
9) T	2,4-D	13.790	14.034	28903450	33117825	737.196	730.375m
10) T	Pentachlo...	14.209	14.641	467.0E6	375.2E6	716.634	728.985m
11) T	2,4,5-TP ...	15.066	15.214	174.6E6	147.0E6	735.784	722.562
12) T	2,4,5-T	15.454	15.862	161.1E6	118.3E6	763.263	714.611
13) T	2,4-DB	16.212	16.458	15143468	16942209	792.755	752.098
14) T	DINOSEB	17.676	16.848	117.8E6	102.5E6	808.846	782.217
15) T	Picloram	17.464	18.245	224.9E6	181.5E6	808.669	806.690
16) T	DCPA	18.059	18.119	220.5E6	183.2E6	780.577	788.342

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031819\
Data File : PP029375.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 18 Mar 2019 7:49 pm
Operator : EI\MA
Sample : HSTDCCC750
Misc : hexane
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Sohil
3/19/2019 1:58:57 PM

Integration File signal 1: sample.E
Integration File signal 2: events2.e
Quant Time: Mar 19 01:50:40 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP022119.M
Quant Title : 8080.M
QLast Update : Fri Feb 22 02:24:04 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
Signal #1 Phase : Rxi-1ms Signal #2 Phase:
Signal #1 Info : 20M x 0.18mm x 0. Signal #2 Info :

