

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP031821.M
 Title : GC EXTRACTABLES
 Last Update : Fri Mar 19 05:03:42 2021
 Response Via : Initial Calibration

Calibration Files

500 =PP034138.D 1000 =PP034136.D 750 =PP034137.D
 250 =PP034139.D 50 =PP034140.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	8.489	8.078	8.268	8.482	8.852	8.434	E4	3.43
2) SA	Decachlorobiphenyl	7.634	7.137	7.326	7.721	8.051	7.574	E4	4.69
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1	4.502	4.195	4.340	4.574	4.693	4.461	E3	4.39
37) L8	AR-1262-2	7.661	7.306	7.453	7.690	7.789	7.580	E3	2.58
38) L8	AR-1262-3	5.441	5.131	5.266	5.444	5.373	5.331	E3	2.50
39) L8	AR-1262-4	4.428	4.149	4.252	4.412	4.308	4.310	E3	2.69
40) L8	AR-1262-5	2.915	2.776	2.769	2.902	2.838	2.840	E3	2.40
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

500 =PP034138.D 1000 =PP034136.D 750 =PP034137.D
 250 =PP034139.D 50 =PP034140.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	4.428	4.112	4.244	4.528	4.834	4.429	E4	6.28
2) SA	Decachlorobiphenyl	2.536	2.351	2.435	2.615	2.806	2.548	E4	6.87
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP031821.M
 Title : GC EXTRACTABLES
 Last Update : Fri Mar 19 05:03:42 2021
 Response Via : Initial Calibration

Calibration Files

500 =PP034138.D 1000 =PP034136.D 750 =PP034137.D
 250 =PP034139.D 50 =PP034140.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1	1.036	0.922	0.971	1.058	1.063	1.010 E3	6.08
37) L8	AR-1262-2	3.208	3.023	3.108	3.303	3.496	3.228 E3	5.68
38) L8	AR-1262-3	1.346	1.230	1.281	1.411	1.583	1.370 E3	10.03
39) L8	AR-1262-4	2.413	2.233	2.309	2.484	2.669	2.422 E3	6.97
40) L8	AR-1262-5	1.072	1.000	1.032	1.101	1.051	1.051 E3	3.64
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range