

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP031821.M
 Title : GC EXTRACTABLES
 Last Update : Fri Mar 19 05:18:29 2021
 Response Via : Initial Calibration

Calibration Files

500 =PP034143.D 1000 =PP034141.D 750 =PP034142.D
 250 =PP034144.D 50 =PP034145.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	8.290	8.108	8.201	8.345	8.630	8.315	E4	2.38
2) SA	Decachlorobiphenyl	1.281	1.235	1.244	1.299	1.347	1.281	E5	3.53
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1	0.993	0.971	0.977	0.999	1.016	0.992	E4	1.80
42) L9	AR-1268-2	9.374	9.173	9.173	9.405	9.528	9.331	E3	1.66
43) L9	AR-1268-3	8.985	8.736	8.819	9.039	8.319	8.780	E3	3.25
44) L9	AR-1268-4	3.231	3.169	3.119	3.100	3.170	3.158	E3	1.62
45) L9	AR-1268-5	2.834	2.770	2.788	2.784	2.682	2.772	E4	2.00

Signal #2 Calibration Files

500 =PP034143.D 1000 =PP034141.D 750 =PP034142.D
 250 =PP034144.D 50 =PP034145.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	4.315	4.123	4.210	4.437	4.681	4.353	E4	5.00
2) SA	Decachlorobiphenyl	4.247	4.038	4.090	4.372	4.699	4.289	E4	6.16
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP031821.M
 Title : GC EXTRACTABLES
 Last Update : Fri Mar 19 05:18:29 2021
 Response Via : Initial Calibration

Calibration Files

500 =PP034143.D 1000 =PP034141.D 750 =PP034142.D
 250 =PP034144.D 50 =PP034145.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1	3.891	3.732	3.787	4.033	4.439	3.976 E3	7.11
42) L9	AR-1268-2	3.670	3.509	3.564	3.777	4.036	3.711 E3	5.62
43) L9	AR-1268-3	3.378	3.187	3.246	3.484	3.714	3.402 E3	6.15
44) L9	AR-1268-4	1.170	1.171	1.183	1.205	1.158	1.178 E3	1.50
45) L9	AR-1268-5	9.370	9.110	9.158	9.499	9.480	9.323 E3	1.94

(#) = Out of Range