

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031821\
 Data File : PP034117.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 18 Mar 2021 15:32
 Operator : DD\AJ
 Sample : AR1660ICC250
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 19 03:37:34 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031821.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Mar 19 03:28:07 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.852	4.263	2123959	1143239	25.589	26.351
2) SA Decachlor...	10.784	9.583	2026197	685345	26.327	26.802
Target Compounds						
3) L1 AR-1016-1	6.166	5.520	659976	305216	272.629	274.710
4) L1 AR-1016-2	6.190	5.540	962642	445374	264.466	272.402
5) L1 AR-1016-3	6.255	5.733	604672	240007	267.723	269.329
6) L1 AR-1016-4	6.362	5.782	485648	196333	264.058	272.052
7) L1 AR-1016-5	6.676	6.012	506833	251355	271.468	270.997
31) L7 AR-1260-1	7.848	7.103	824200	412846	264.871	276.590
32) L7 AR-1260-2	8.114	7.297	966596	468151	264.308	272.524
33) L7 AR-1260-3	8.478	7.453	788134	444941	261.562	272.161
34) L7 AR-1260-4	8.708	7.934	886655	374144	259.754	274.730
35) L7 AR-1260-5	9.027	8.177	1732496	798319	256.146	265.259



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031821\
Data File : PP034117.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 18 Mar 2021 15:32
Operator : DD\AJ
Sample : AR1660ICC250
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :
AR1660ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 03:37:34 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031821.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 19 03:28:07 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.852 min
Delta R.T.: 0.000 min
Response: 2123959
Conc: 25.59 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#1 Tetrachloro-m-xylene

R.T.: 4.263 min
Delta R.T.: 0.000 min
Response: 1143239
Conc: 26.35 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.784 min
Delta R.T.: 0.000 min
Response: 2026197
Conc: 26.33 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.583 min
Delta R.T.: 0.000 min
Response: 685345
Conc: 26.80 ng/ml



#3 AR-1016-1


R.T.: 6.166 min
Delta R.T.: 0.000 min
Response: 659976
Conc: 272.63 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#3 AR-1016-1

R.T.: 5.520 min
Delta R.T.: 0.000 min
Response: 305216
Conc: 274.71 ng/ml




#4 AR-1016-2

R.T.: 6.190 min
Delta R.T.: 0.000 min
Response: 962642
Conc: 264.47 ng/ml



#4 AR-1016-2


R.T.: 5.540 min
Delta R.T.: 0.000 min
Response: 445374
Conc: 272.40 ng/ml



#5 AR-1016-3


R.T.: 6.255 min
Delta R.T.: 0.000 min
Response: 604672
Conc: 267.72 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#5 AR-1016-3

R.T.: 5.733 min
Delta R.T.: 0.000 min
Response: 240007
Conc: 269.33 ng/ml




#6 AR-1016-4

R.T.: 6.362 min
Delta R.T.: 0.000 min
Response: 485648
Conc: 264.06 ng/ml



#6 AR-1016-4


R.T.: 5.782 min
Delta R.T.: 0.000 min
Response: 196333
Conc: 272.05 ng/ml



#7 AR-1016-5


R.T.: 6.676 min
Delta R.T.: 0.000 min
Response: 506833
Conc: 271.47 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#7 AR-1016-5

R.T.: 6.012 min
Delta R.T.: 0.000 min
Response: 251355
Conc: 271.00 ng/ml




#31 AR-1260-1

R.T.: 7.848 min
Delta R.T.: 0.000 min
Response: 824200
Conc: 264.87 ng/ml



#31 AR-1260-1


R.T.: 7.103 min
Delta R.T.: 0.000 min
Response: 412846
Conc: 276.59 ng/ml



#32 AR-1260-2


R.T.: 8.114 min
Delta R.T.: 0.000 min
Response: 966596
Conc: 264.31 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#32 AR-1260-2

R.T.: 7.297 min
Delta R.T.: 0.000 min
Response: 468151
Conc: 272.52 ng/ml




#33 AR-1260-3

R.T.: 8.478 min
Delta R.T.: 0.000 min
Response: 788134
Conc: 261.56 ng/ml



#33 AR-1260-3


R.T.: 7.453 min
Delta R.T.: 0.000 min
Response: 444941
Conc: 272.16 ng/ml



#34 AR-1260-4


R.T.: 8.708 min
Delta R.T.: 0.000 min
Response: 886655
Conc: 259.75 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#34 AR-1260-4

R.T.: 7.934 min
Delta R.T.: 0.000 min
Response: 374144
Conc: 274.73 ng/ml



#35 AR-1260-5

R.T.: 9.027 min
Delta R.T.: 0.000 min
Response: 1732496
Conc: 256.15 ng/ml



#35 AR-1260-5

R.T.: 8.177 min
Delta R.T.: 0.000 min
Response: 798319
Conc: 265.26 ng/ml