

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031821\
 Data File : PP034132.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 18 Mar 2021 19:46
 Operator : DD\AJ
 Sample : AR1254ICC750
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 19 04:38:36 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031821.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Mar 19 04:33:45 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.853	4.263	6142477	3154571	75.365	75.382
2) SA Decachlor...	10.787	9.584	5391221	1776041	74.566	74.876
Target Compounds						
26) L6 AR-1254-1	7.073	6.390	1940748	1194316	747.100	749.217
27) L6 AR-1254-2	7.303	6.547	2968971	1030717	748.780	749.320
28) L6 AR-1254-3	7.683	6.968	3167631	1645276	748.576	745.919
29) L6 AR-1254-4	7.979	7.204	2101933	870580	747.558	748.924
30) L6 AR-1254-5	8.406	7.632	2477094	1347353	749.074	747.027



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031821\
Data File : PP034132.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 18 Mar 2021 19:46
Operator : DD\AJ
Sample : AR1254ICC750
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 04:38:36 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031821.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 19 04:33:45 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.853 min
Delta R.T.: 0.000 min
Response: 6142477
Conc: 75.37 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750




#1 Tetrachloro-m-xylene

R.T.: 4.263 min
Delta R.T.: 0.000 min
Response: 3154571
Conc: 75.38 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.787 min
Delta R.T.: 0.000 min
Response: 5391221
Conc: 74.57 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.584 min
Delta R.T.: 0.000 min
Response: 1776041
Conc: 74.88 ng/ml



#26 AR-1254-1


R.T.: 7.073 min
Delta R.T.: 0.000 min
Response: 1940748
Conc: 747.10 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750




#26 AR-1254-1

R.T.: 6.390 min
Delta R.T.: 0.000 min
Response: 1194316
Conc: 749.22 ng/ml




#27 AR-1254-2

R.T.: 7.303 min
Delta R.T.: 0.000 min
Response: 2968971
Conc: 748.78 ng/ml



#27 AR-1254-2


R.T.: 6.547 min
Delta R.T.: 0.000 min
Response: 1030717
Conc: 749.32 ng/ml



#28 AR-1254-3


R.T.: 7.683 min
Delta R.T.: 0.000 min
Response: 3167631
Conc: 748.58 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750




#28 AR-1254-3

R.T.: 6.968 min
Delta R.T.: 0.000 min
Response: 1645276
Conc: 745.92 ng/ml




#29 AR-1254-4

R.T.: 7.979 min
Delta R.T.: 0.000 min
Response: 2101933
Conc: 747.56 ng/ml



#29 AR-1254-4


R.T.: 7.204 min
Delta R.T.: 0.000 min
Response: 870580
Conc: 748.92 ng/ml



#30 AR-1254-5

R.T.: 8.406 min
Delta R.T.: 0.000 min
Response: 2477094
Conc: 749.07 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750



#30 AR-1254-5

R.T.: 7.632 min
Delta R.T.: 0.000 min
Response: 1347353
Conc: 747.03 ng/ml