

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031821\
 Data File : PP034125.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 18 Mar 2021 17:48
 Operator : DD\AJ
 Sample : AR1242ICC050
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 19 03:56:41 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031821.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Mar 19 03:46:41 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.852	4.262	438345	242289	5.197	5.440
2) SA Decachlor...	10.783	9.583	404471	142191	5.276	5.487
Target Compounds						
16) L4 AR-1242-1	6.167	5.519	115785	49660	58.105	54.521
17) L4 AR-1242-2	6.190	5.540	169488	69698	57.003	53.099
18) L4 AR-1242-3	6.255	5.733	103178	38423	55.447	52.844
19) L4 AR-1242-4	6.362	5.827	82726	39215	54.876	54.133
20) L4 AR-1242-5	7.142	6.389	80568	39929	52.847	50.875



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031821\
Data File : PP034125.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 18 Mar 2021 17:48
Operator : DD\AJ
Sample : AR1242ICC050
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 03:56:41 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031821.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 19 03:46:41 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.852 min
Delta R.T.: 0.000 min
Response: 438345
Conc: 5.20 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC050




#1 Tetrachloro-m-xylene

R.T.: 4.262 min
Delta R.T.: 0.000 min
Response: 242289
Conc: 5.44 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.783 min
Delta R.T.: 0.000 min
Response: 404471
Conc: 5.28 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.583 min
Delta R.T.: 0.000 min
Response: 142191
Conc: 5.49 ng/ml



#16 AR-1242-1


R.T.: 6.167 min
Delta R.T.: 0.000 min
Response: 115785
Conc: 58.11 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC050




#16 AR-1242-1

R.T.: 5.519 min
Delta R.T.: 0.000 min
Response: 49660
Conc: 54.52 ng/ml




#17 AR-1242-2

R.T.: 6.190 min
Delta R.T.: 0.000 min
Response: 169488
Conc: 57.00 ng/ml



#17 AR-1242-2


R.T.: 5.540 min
Delta R.T.: 0.000 min
Response: 69698
Conc: 53.10 ng/ml



#18 AR-1242-3


R.T.: 6.255 min
Delta R.T.: 0.000 min
Response: 103178
Conc: 55.45 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC050




#18 AR-1242-3

R.T.: 5.733 min
Delta R.T.: 0.000 min
Response: 38423
Conc: 52.84 ng/ml




#19 AR-1242-4

R.T.: 6.362 min
Delta R.T.: 0.000 min
Response: 82726
Conc: 54.88 ng/ml



#19 AR-1242-4


R.T.: 5.827 min
Delta R.T.: 0.000 min
Response: 39215
Conc: 54.13 ng/ml



#20 AR-1242-5

R.T.: 7.142 min
Delta R.T.: 0.000 min
Response: 80568
Conc: 52.85 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC050



#20 AR-1242-5

R.T.: 6.389 min
Delta R.T.: 0.000 min
Response: 39929
Conc: 50.88 ng/ml