

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031822\
 Data File : PP044871.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Mar 2022 19:26
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 18 00:54:22 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 17 04:32:40 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.898	3.156	99954195	72147878	54.858	54.747
2) SA Decachlor...	10.022	8.513	60803954	60147165	47.235	46.289
Target Compounds						
3) L1 AR-1016-1	5.149	4.298	30665441	28010415	522.591	552.961
4) L1 AR-1016-2	5.172	4.317	44753423	38918423	500.621	539.871
5) L1 AR-1016-3	5.238	4.501	26720976	21500663	513.439	546.465
6) L1 AR-1016-4	5.344	4.550	21990245	17508278	513.912	556.110
7) L1 AR-1016-5	5.664	4.775	21690255	21579641	494.251	540.637
31) L7 AR-1260-1	6.893	5.882	32476407	36321191	471.727	515.846
32) L7 AR-1260-2	7.176	6.089	36790161	42858656	473.859	479.248
33) L7 AR-1260-3	7.571	6.251	28725367	39592715	456.493	464.871
34) L7 AR-1260-4	7.819	6.763	32847033	32547040	447.777	484.280
35) L7 AR-1260-5	8.161	7.030	63699384	73691338	454.887	471.604



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031822\
Data File : PP044871.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Mar 2022 19:26
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 18 00:54:22 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031722.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 17 04:32:40 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.898 min
Delta R.T.: 0.000 min
Response: 99954195
Conc: 54.86 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.156 min
Delta R.T.: 0.000 min
Response: 72147878
Conc: 54.75 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.022 min
Delta R.T.: -0.001 min
Response: 60803954
Conc: 47.23 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.513 min
Delta R.T.: -0.003 min
Response: 60147165
Conc: 46.29 ng/ml



#3 AR-1016-1


R.T.: 5.149 min
Delta R.T.: -0.001 min
Response: 30665441
Conc: 522.59 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.298 min
Delta R.T.: 0.000 min
Response: 28010415
Conc: 552.96 ng/ml




#4 AR-1016-2

R.T.: 5.172 min
Delta R.T.: 0.000 min
Response: 44753423
Conc: 500.62 ng/ml



#4 AR-1016-2


R.T.: 4.317 min
Delta R.T.: 0.000 min
Response: 38918423
Conc: 539.87 ng/ml



#5 AR-1016-3


R.T.: 5.238 min
Delta R.T.: 0.000 min
Response: 26720976
Conc: 513.44 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.501 min
Delta R.T.: -0.001 min
Response: 21500663
Conc: 546.46 ng/ml




#6 AR-1016-4

R.T.: 5.344 min
Delta R.T.: 0.000 min
Response: 21990245
Conc: 513.91 ng/ml



#6 AR-1016-4


R.T.: 4.550 min
Delta R.T.: -0.001 min
Response: 17508278
Conc: 556.11 ng/ml



#7 AR-1016-5


R.T.: 5.664 min
Delta R.T.: 0.000 min
Response: 21690255
Conc: 494.25 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 4.775 min
Delta R.T.: 0.000 min
Response: 21579641
Conc: 540.64 ng/ml




#31 AR-1260-1

R.T.: 6.893 min
Delta R.T.: -0.001 min
Response: 32476407
Conc: 471.73 ng/ml



#31 AR-1260-1


R.T.: 5.882 min
Delta R.T.: -0.001 min
Response: 36321191
Conc: 515.85 ng/ml



#32 AR-1260-2


R.T.: 7.176 min
Delta R.T.: 0.000 min
Response: 36790161
Conc: 473.86 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.089 min
Delta R.T.: -0.002 min
Response: 42858656
Conc: 479.25 ng/ml




#33 AR-1260-3

R.T.: 7.571 min
Delta R.T.: 0.000 min
Response: 28725367
Conc: 456.49 ng/ml



#33 AR-1260-3


R.T.: 6.251 min
Delta R.T.: -0.002 min
Response: 39592715
Conc: 464.87 ng/ml



#34 AR-1260-4


R.T.: 7.819 min
Delta R.T.: 0.000 min
Response: 32847033
Conc: 447.78 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 6.763 min
Delta R.T.: -0.002 min
Response: 32547040
Conc: 484.28 ng/ml



#35 AR-1260-5

R.T.: 8.161 min
Delta R.T.: 0.000 min
Response: 63699384
Conc: 454.89 ng/ml



#35 AR-1260-5

R.T.: 7.030 min
Delta R.T.: -0.002 min
Response: 73691338
Conc: 471.60 ng/ml