

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_P\Data\PP031919\  
 Data File : PP029425.D  
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH  
 Acq On : 20 Mar 2019 1:34  
 Operator : EI\MA  
 Sample : HSTDCCC750  
 Misc : hexane  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 ECD\_P  
 ClientSampleId :  
 HSTDCCC750

Manual Integrations  
 APPROVED

Sohil  
 3/20/2019 4:06:24 PM

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Mar 20 03:38:19 2019  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_P\methods\PP022119.M  
 Quant Title : 8080.M  
 QLast Update : Fri Feb 22 02:24:04 2019  
 Response via : Initial Calibration  
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl  
 Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2  
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

|                             | Compound     | RT#1   | RT#2   | Resp#1   | Resp#2   | ng/ml    | ng/ml    |
|-----------------------------|--------------|--------|--------|----------|----------|----------|----------|
| -----                       |              |        |        |          |          |          |          |
| System Monitoring Compounds |              |        |        |          |          |          |          |
| 4) S                        | 2,4-DCAA     | 11.927 | 12.123 | 20964189 | 28219484 | 697.496  | 743.547  |
| Target Compounds            |              |        |        |          |          |          |          |
| 1) T                        | Dalapon      | 3.241  | 2.956  | 39892788 | 33526857 | 681.014  | 647.884  |
| 2) T                        | 3,5-DICHL... | 10.336 | 10.210 | 41370773 | 34382119 | 670.680  | 678.376  |
| 3) T                        | 4-Nitroph... | 11.491 | 11.268 | 46561270 | 16675975 | 679.690  | 696.346  |
| 5) T                        | DICAMBA      | 12.249 | 12.435 | 92024650 | 87081658 | 692.077  | 680.324  |
| 8) T                        | DICHLORPROP  | 13.438 | 13.546 | 22660263 | 24002499 | 734.398  | 709.948  |
| 9) T                        | 2,4-D        | 13.787 | 14.032 | 28851055 | 32078853 | 735.860  | 707.462  |
| 10) T                       | Pentachlo... | 14.207 | 14.641 | 463.7E6  | 355.8E6  | 711.636  | 691.248  |
| 11) T                       | 2,4,5-TP ... | 15.063 | 15.211 | 175.4E6  | 141.8E6  | 739.529  | 697.131  |
| 12) T                       | 2,4,5-T      | 15.451 | 15.860 | 158.6E6  | 117.4E6  | 751.200  | 709.188  |
| 13) T                       | 2,4-DB       | 16.207 | 16.456 | 14312622 | 18071297 | 749.261m | 802.221  |
| 14) T                       | DINOSEB      | 17.671 | 16.846 | 115.4E6  | 103.0E6  | 792.755m | 786.068  |
| 15) T                       | Picloram     | 17.460 | 18.242 | 226.8E6  | 181.9E6  | 815.508  | 808.263m |
| 16) T                       | DCPA         | 18.055 | 18.117 | 223.5E6  | 184.9E6  | 791.102  | 795.654  |
| -----                       |              |        |        |          |          |          |          |

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_P\Data\PP031919\  
Data File : PP029425.D  
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH  
Acq On : 20 Mar 2019 1:34  
Operator : EI\MA  
Sample : HSTDCCC750  
Misc : hexane  
ALS Vial : 3 Sample Multiplier: 1

Instrument :  
ECD\_P  
ClientSampleId :  
HSTDCCC750

Manual Integrations  
APPROVED

Sohil  
3/20/2019 4:06:24 PM

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Mar 20 03:38:19 2019  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_P\methods\PP022119.M  
Quant Title : 8080.M  
QLast Update : Fri Feb 22 02:24:04 2019  
Response via : Initial Calibration  
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl  
Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2  
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

