

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031921\
 Data File : PP034171.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 19 Mar 2021 14:16
 Operator : DD\AJ
 Sample : M1720-06
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 20 02:40:11 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031821.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Mar 20 01:42:16 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.850	4.261	1815376	948682	21.610	21.438
2)	SA Decachlor...	10.776	9.580	1711168	588270	21.805	22.292
Target Compounds							
8)	L2 AR-1221-1	5.101	0.000	23575	0	25.963	N.D. #
9)	L2 AR-1221-2	5.200	4.616	22374	8759	32.510	24.699
24)	L5 AR-1248-4	7.105	0.000	13259	0	5.132	N.D. #
28)	L6 AR-1254-3	7.696	0.000	17220	0	3.965	N.D. #
30)	L6 AR-1254-5	8.408	7.627	331747	15176	99.378	8.219 #
31)	L7 AR-1260-1	7.848	7.105	3724	28840	1.168	18.500 #
32)	L7 AR-1260-2	8.108	7.309	86231	32571	22.946	18.203
34)	L7 AR-1260-4	8.681f	0.000	82138	0	23.592	N.D. #
36)	L8 AR-1262-1	0.000	7.627	0	15176	N.D.	15.029 #
39)	L8 AR-1262-4	9.446f	0.000	12295	0	2.853	N.D. #
42)	L9 AR-1268-2	9.446f	0.000	12295	0	1.318	N.D. #
43)	L9 AR-1268-3	0.000	8.743	0	10444	N.D.	3.065 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031921\
Data File : PP034171.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 19 Mar 2021 14:16
Operator : DD\AJ
Sample : M1720-06
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 20 02:40:11 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031821.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Mar 20 01:42:16 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.850 min
Delta R.T.: -0.002 min
Response: 1815376
Conc: 21.61 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.261 min
Delta R.T.: -0.001 min
Response: 948682
Conc: 21.44 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.776 min
Delta R.T.: -0.008 min
Response: 1711168
Conc: 21.80 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.580 min
Delta R.T.: -0.005 min
Response: 588270
Conc: 22.29 ng/ml



#8 AR-1221-1


R.T.: 5.101 min
Delta R.T.: -0.002 min
Response: 23575
Conc: 25.96 ng/ml

Instrument :
ECD_P
ClientSampleId :




#8 AR-1221-1

R.T.: 0.000 min
Exp R.T. : 4.519 min
Response: 0
Conc: N.D.




#9 AR-1221-2

R.T.: 5.200 min
Delta R.T.: 0.000 min
Response: 22374
Conc: 32.51 ng/ml



#9 AR-1221-2


R.T.: 4.616 min
Delta R.T.: -0.003 min
Response: 8759
Conc: 24.70 ng/ml



#24 AR-1248-4


R.T.: 7.105 min
Delta R.T.: 0.001 min
Response: 13259
Conc: 5.13 ng/ml

Instrument :
ECD_P
ClientSampleId :




#24 AR-1248-4

R.T.: 0.000 min
Exp R.T. : 6.012 min
Response: 0
Conc: N.D.




#28 AR-1254-3

R.T.: 7.696 min
Delta R.T.: 0.014 min
Response: 17220
Conc: 3.96 ng/ml



#28 AR-1254-3


R.T.: 0.000 min
Exp R.T. : 6.968 min
Response: 0
Conc: N.D.



#30 AR-1254-5


R.T.: 8.408 min
Delta R.T.: 0.003 min
Response: 331747
Conc: 99.38 ng/ml

Instrument :
ECD_P
ClientSampleId :




#30 AR-1254-5

R.T.: 7.627 min
Delta R.T.: -0.004 min
Response: 15176
Conc: 8.22 ng/ml




#31 AR-1260-1

R.T.: 7.848 min
Delta R.T.: 0.000 min
Response: 3724
Conc: 1.17 ng/ml



#31 AR-1260-1


R.T.: 7.105 min
Delta R.T.: 0.002 min
Response: 28840
Conc: 18.50 ng/ml



#32 AR-1260-2


R.T.: 8.108 min
Delta R.T.: -0.006 min
Response: 86231
Conc: 22.95 ng/ml

Instrument :
ECD_P
ClientSampleId :




#32 AR-1260-2

R.T.: 7.309 min
Delta R.T.: 0.012 min
Response: 32571
Conc: 18.20 ng/ml




#34 AR-1260-4

R.T.: 8.681 min
Delta R.T.: -0.027 min
Response: 82138
Conc: 23.59 ng/ml



#34 AR-1260-4


R.T.: 0.000 min
Exp R.T.: 7.934 min
Response: 0
Conc: N.D.



#36 AR-1262-1


R.T.: 0.000 min
Exp R.T.: 8.479 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#36 AR-1262-1

R.T.: 7.627 min
Delta R.T.: -0.003 min
Response: 15176
Conc: 15.03 ng/ml




#39 AR-1262-4

R.T.: 9.446 min
Delta R.T.: 0.026 min
Response: 12295
Conc: 2.85 ng/ml



#39 AR-1262-4


R.T.: 0.000 min
Exp R.T.: 8.527 min
Response: 0
Conc: N.D.



#42 AR-1268-2


R.T.: 9.446 min
Delta R.T.: 0.025 min
Response: 12295
Conc: 1.32 ng/ml

Instrument :
ECD_P
ClientSampleId :




#42 AR-1268-2

R.T.: 0.000 min
Exp R.T. : 8.527 min
Response: 0
Conc: N.D.



#43 AR-1268-3

R.T.: 0.000 min
Exp R.T. : 9.639 min
Response: 0
Conc: N.D.



#43 AR-1268-3

R.T.: 8.743 min
Delta R.T.: 0.008 min
Response: 10444
Conc: 3.07 ng/ml