

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031924\
 Data File : PP063551.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Mar 2024 15:57
 Operator : YP\AJ
 Sample : AR1660ICC1000
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 19 03:35:27 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031924.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Mar 19 03:34:36 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.490	3.640	233.4E6	247.9E6	95.278	100.417
2) SA Decachlor...	10.367	8.683	238.5E6	254.9E6	93.271	94.242
Target Compounds						
3) L1 AR-1016-1	5.674	4.733	77484508	78145055	928.663	931.841
4) L1 AR-1016-2	5.696	4.752	112.2E6	112.2E6	919.562	952.259
5) L1 AR-1016-3	5.759	4.930	72780847	55133100	897.453	872.330
6) L1 AR-1016-4	5.859	4.972	58974642	49728649	916.882	895.804
7) L1 AR-1016-5	6.157	5.186	59206810	64649291	903.791	917.505
31) L7 AR-1260-1	7.294	6.226	113.4E6	131.9E6	914.119	924.511
32) L7 AR-1260-2	7.553	6.415	130.8E6	158.5E6	931.590	938.864
33) L7 AR-1260-3	7.915	6.569	92493116	151.4E6	937.448	937.039
34) L7 AR-1260-4	8.144	7.044	111.5E6	114.2E6	940.786	932.579
35) L7 AR-1260-5	8.475	7.287	203.7E6	257.4E6	969.580	958.018



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031924\
Data File : PP063551.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2024 15:57
Operator : YP\AJ
Sample : AR1660ICC1000
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 03:35:27 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031924.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 19 03:34:36 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.490 min
Delta R.T.: 0.000 min
Response: 233412493
Conc: 95.28 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.640 min
Delta R.T.: 0.000 min
Response: 247934111
Conc: 100.42 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.367 min
Delta R.T.: 0.001 min
Response: 238540670
Conc: 93.27 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.683 min
Delta R.T.: 0.000 min
Response: 254941711
Conc: 94.24 ng/ml



#3 AR-1016-1


R.T.: 5.674 min
Delta R.T.: 0.000 min
Response: 77484508
Conc: 928.66 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000




#3 AR-1016-1

R.T.: 4.733 min
Delta R.T.: 0.000 min
Response: 78145055
Conc: 931.84 ng/ml




#4 AR-1016-2

R.T.: 5.696 min
Delta R.T.: 0.000 min
Response: 112156676
Conc: 919.56 ng/ml



#4 AR-1016-2


R.T.: 4.752 min
Delta R.T.: 0.000 min
Response: 112152695
Conc: 952.26 ng/ml



#5 AR-1016-3


R.T.: 5.759 min
Delta R.T.: 0.000 min
Response: 72780847
Conc: 897.45 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000




#5 AR-1016-3

R.T.: 4.930 min
Delta R.T.: 0.000 min
Response: 55133100
Conc: 872.33 ng/ml




#6 AR-1016-4

R.T.: 5.859 min
Delta R.T.: 0.000 min
Response: 58974642
Conc: 916.88 ng/ml



#6 AR-1016-4


R.T.: 4.972 min
Delta R.T.: 0.000 min
Response: 49728649
Conc: 895.80 ng/ml



#7 AR-1016-5


R.T.: 6.157 min
Delta R.T.: 0.000 min
Response: 59206810
Conc: 903.79 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000




#7 AR-1016-5

R.T.: 5.186 min
Delta R.T.: 0.000 min
Response: 64649291
Conc: 917.50 ng/ml




#31 AR-1260-1

R.T.: 7.294 min
Delta R.T.: 0.000 min
Response: 113428897
Conc: 914.12 ng/ml



#31 AR-1260-1


R.T.: 6.226 min
Delta R.T.: 0.000 min
Response: 131937737
Conc: 924.51 ng/ml



#32 AR-1260-2


R.T.: 7.553 min
Delta R.T.: 0.000 min
Response: 130809958
Conc: 931.59 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000




#32 AR-1260-2

R.T.: 6.415 min
Delta R.T.: 0.000 min
Response: 158454772
Conc: 938.86 ng/ml




#33 AR-1260-3

R.T.: 7.915 min
Delta R.T.: 0.000 min
Response: 92493116
Conc: 937.45 ng/ml



#33 AR-1260-3


R.T.: 6.569 min
Delta R.T.: 0.000 min
Response: 151366848
Conc: 937.04 ng/ml



#34 AR-1260-4


R.T.: 8.144 min
Delta R.T.: 0.000 min
Response: 111496123
Conc: 940.79 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000




#34 AR-1260-4

R.T.: 7.044 min
Delta R.T.: 0.000 min
Response: 114181392
Conc: 932.58 ng/ml



#35 AR-1260-5

R.T.: 8.475 min
Delta R.T.: 0.000 min
Response: 203708057
Conc: 969.58 ng/ml



#35 AR-1260-5

R.T.: 7.287 min
Delta R.T.: 0.000 min
Response: 257357590
Conc: 958.02 ng/ml