

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031924\
 Data File : PP063570.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Mar 2024 21:06
 Operator : YP\AJ
 Sample : AR1254ICC500
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 19 05:20:49 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031924.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Mar 19 05:19:48 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.489	3.638	115.9E6	121.2E6	50.000	50.000
2) SA Decachlor...	10.364	8.681	122.6E6	132.3E6	50.000	50.000
Target Compounds						
26) L6 AR-1254-1	6.536	5.539	45797741	62285618	500.000	500.000
27) L6 AR-1254-2	6.757	5.687	67971496	57533782	500.000	500.000
28) L6 AR-1254-3	7.126	6.093	69174830	91225250	500.000	500.000
29) L6 AR-1254-4	7.414	6.323	45415453	47315943	500.000	500.000
30) L6 AR-1254-5	7.836	6.743	59119486	85678965	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031924\
Data File : PP063570.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2024 21:06
Operator : YP\AJ
Sample : AR1254ICC500
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 05:20:49 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031924.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 19 05:19:48 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.489 min
Delta R.T.: 0.000 min
Response: 115896288
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.638 min
Delta R.T.: 0.000 min
Response: 121215616
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.364 min
Delta R.T.: 0.000 min
Response: 122634224
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.681 min
Delta R.T.: 0.000 min
Response: 132311299
Conc: 50.00 ng/ml



#26 AR-1254-1


R.T.: 6.536 min
Delta R.T.: 0.000 min
Response: 45797741
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500




#26 AR-1254-1

R.T.: 5.539 min
Delta R.T.: 0.000 min
Response: 62285618
Conc: 500.00 ng/ml




#27 AR-1254-2

R.T.: 6.757 min
Delta R.T.: 0.000 min
Response: 67971496
Conc: 500.00 ng/ml



#27 AR-1254-2


R.T.: 5.687 min
Delta R.T.: 0.000 min
Response: 57533782
Conc: 500.00 ng/ml



#28 AR-1254-3


R.T.: 7.126 min
Delta R.T.: 0.000 min
Response: 69174830
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500




#28 AR-1254-3

R.T.: 6.093 min
Delta R.T.: 0.000 min
Response: 91225250
Conc: 500.00 ng/ml




#29 AR-1254-4

R.T.: 7.414 min
Delta R.T.: 0.000 min
Response: 45415453
Conc: 500.00 ng/ml



#29 AR-1254-4


R.T.: 6.323 min
Delta R.T.: 0.000 min
Response: 47315943
Conc: 500.00 ng/ml



#30 AR-1254-5

R.T.: 7.836 min
Delta R.T.: 0.000 min
Response: 59119486
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254\ICC500



#30 AR-1254-5

R.T.: 6.743 min
Delta R.T.: 0.000 min
Response: 85678965
Conc: 500.00 ng/ml