

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031924\
 Data File : PP063580.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Mar 2024 23:49
 Operator : YP\AJ
 Sample : AR1242ICV500
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP031924AR1242

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 19 04:44:37 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031924.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Mar 19 04:42:27 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.489	3.638	119.2E6	124.7E6	51.123	51.030
2) SA Decachlor...	10.366	8.681	122.8E6	131.8E6	50.839	51.868
Target Compounds						
16) L4 AR-1242-1	5.672	4.732	30872138	30246785	506.299	508.660
17) L4 AR-1242-2	5.695	4.750	44129245	42029646	513.490	505.932
18) L4 AR-1242-3	5.758	4.927	29156104	23213475	511.814	506.738
19) L4 AR-1242-4	5.857	5.012	23644395	25431150	513.461	508.577
20) L4 AR-1242-5	6.602	5.538	23534771	30425445	527.977	508.634



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031924\
Data File : PP063580.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2024 23:49
Operator : YP\AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP031924AR1242

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 04:44:37 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031924.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 19 04:42:27 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.489 min
Delta R.T.: 0.000 min
Response: 119209659
Conc: 51.12 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP031924AR1242




#1 Tetrachloro-m-xylene

R.T.: 3.638 min
Delta R.T.: 0.000 min
Response: 124678452
Conc: 51.03 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.366 min
Delta R.T.: 0.000 min
Response: 122803937
Conc: 50.84 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.681 min
Delta R.T.: -0.002 min
Response: 131796002
Conc: 51.87 ng/ml



#16 AR-1242-1


R.T.: 5.672 min
Delta R.T.: -0.001 min
Response: 30872138
Conc: 506.30 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP031924AR1242




#16 AR-1242-1

R.T.: 4.732 min
Delta R.T.: -0.002 min
Response: 30246785
Conc: 508.66 ng/ml




#17 AR-1242-2

R.T.: 5.695 min
Delta R.T.: -0.002 min
Response: 44129245
Conc: 513.49 ng/ml



#17 AR-1242-2


R.T.: 4.750 min
Delta R.T.: -0.002 min
Response: 42029646
Conc: 505.93 ng/ml



#18 AR-1242-3


R.T.: 5.758 min
Delta R.T.: -0.001 min
Response: 29156104
Conc: 511.81 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP031924AR1242




#18 AR-1242-3

R.T.: 4.927 min
Delta R.T.: -0.002 min
Response: 23213475
Conc: 506.74 ng/ml




#19 AR-1242-4

R.T.: 5.857 min
Delta R.T.: -0.001 min
Response: 23644395
Conc: 513.46 ng/ml



#19 AR-1242-4


R.T.: 5.012 min
Delta R.T.: -0.001 min
Response: 25431150
Conc: 508.58 ng/ml



#20 AR-1242-5

R.T.: 6.602 min
Delta R.T.: -0.001 min
Response: 23534771
Conc: 527.98 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP031924AR1242



#20 AR-1242-5

R.T.: 5.538 min
Delta R.T.: -0.002 min
Response: 30425445
Conc: 508.63 ng/ml