

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031924\
 Data File : PP063584.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Mar 2024 04:21
 Operator : YP\AJ
 Sample : PP031924ICV500
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP031924

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 19 05:59:03 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031924.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Mar 19 05:56:49 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.489	3.638	128.1E6	127.1E6	53.678	53.246
2) SA Decachlor...	10.369	8.681	133.7E6	143.1E6	52.660	52.950
Target Compounds						
3) L1 AR-1016-1	5.674	4.732	43259476	42927941	519.127	519.674
4) L1 AR-1016-2	5.696	4.750	63613319	60546689	525.914	522.282
5) L1 AR-1016-3	5.759	4.928	41882976	32623616	524.443	524.159
6) L1 AR-1016-4	5.859	4.970	33369182	26908489	533.110	491.004
7) L1 AR-1016-5	6.157	5.185	34375188	37227087	533.490	535.541
31) L7 AR-1260-1	7.295	6.224	65102638	74931788	515.451	524.336
32) L7 AR-1260-2	7.553	6.412	74324555	89353367	520.624	528.580
33) L7 AR-1260-3	7.916	6.566	53196379	83995460	529.975	522.594
34) L7 AR-1260-4	8.144	7.042	63602982	63361283	528.697	519.826
35) L7 AR-1260-5	8.475	7.284	110.7E6	142.5E6	533.099	534.234



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031924\
Data File : PP063584.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Mar 2024 04:21
Operator : YP\AJ
Sample : PP031924ICV500
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP031924

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 05:59:03 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031924.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 19 05:56:49 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.489 min
Delta R.T.: 0.000 min
Response: 128083663
Conc: 53.68 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP031924




#1 Tetrachloro-m-xylene

R.T.: 3.638 min
Delta R.T.: 0.000 min
Response: 127126095
Conc: 53.25 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.369 min
Delta R.T.: 0.002 min
Response: 133703765
Conc: 52.66 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.681 min
Delta R.T.: -0.003 min
Response: 143098698
Conc: 52.95 ng/ml



#3 AR-1016-1


R.T.: 5.674 min
Delta R.T.: 0.000 min
Response: 43259476
Conc: 519.13 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP031924




#3 AR-1016-1

R.T.: 4.732 min
Delta R.T.: -0.001 min
Response: 42927941
Conc: 519.67 ng/ml




#4 AR-1016-2

R.T.: 5.696 min
Delta R.T.: 0.000 min
Response: 63613319
Conc: 525.91 ng/ml



#4 AR-1016-2


R.T.: 4.750 min
Delta R.T.: -0.001 min
Response: 60546689
Conc: 522.28 ng/ml



#5 AR-1016-3


R.T.: 5.759 min
Delta R.T.: 0.000 min
Response: 41882976
Conc: 524.44 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP031924




#5 AR-1016-3

R.T.: 4.928 min
Delta R.T.: -0.001 min
Response: 32623616
Conc: 524.16 ng/ml




#6 AR-1016-4

R.T.: 5.859 min
Delta R.T.: 0.000 min
Response: 33369182
Conc: 533.11 ng/ml



#6 AR-1016-4


R.T.: 4.970 min
Delta R.T.: -0.002 min
Response: 26908489
Conc: 491.00 ng/ml



#7 AR-1016-5


R.T.: 6.157 min
Delta R.T.: 0.000 min
Response: 34375188
Conc: 533.49 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP031924




#7 AR-1016-5

R.T.: 5.185 min
Delta R.T.: -0.002 min
Response: 37227087
Conc: 535.54 ng/ml




#31 AR-1260-1

R.T.: 7.295 min
Delta R.T.: 0.000 min
Response: 65102638
Conc: 515.45 ng/ml



#31 AR-1260-1


R.T.: 6.224 min
Delta R.T.: -0.002 min
Response: 74931788
Conc: 524.34 ng/ml



#32 AR-1260-2


R.T.: 7.553 min
Delta R.T.: 0.000 min
Response: 74324555
Conc: 520.62 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP031924




#32 AR-1260-2

R.T.: 6.412 min
Delta R.T.: -0.003 min
Response: 89353367
Conc: 528.58 ng/ml




#33 AR-1260-3

R.T.: 7.916 min
Delta R.T.: 0.001 min
Response: 53196379
Conc: 529.97 ng/ml



#33 AR-1260-3


R.T.: 6.566 min
Delta R.T.: -0.003 min
Response: 83995460
Conc: 522.59 ng/ml



#34 AR-1260-4


R.T.: 8.144 min
Delta R.T.: 0.000 min
Response: 63602982
Conc: 528.70 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP031924




#34 AR-1260-4

R.T.: 7.042 min
Delta R.T.: -0.002 min
Response: 63361283
Conc: 519.83 ng/ml



#35 AR-1260-5

R.T.: 8.475 min
Delta R.T.: 0.000 min
Response: 110664830
Conc: 533.10 ng/ml



#35 AR-1260-5

R.T.: 7.284 min
Delta R.T.: -0.002 min
Response: 142535083
Conc: 534.23 ng/ml