

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031925\
 Data File : PP070697.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Mar 2025 14:32
 Operator : YP\AJ
 Sample : Q1598-01 10X
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ETGI-354

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 03/20/2025
 Supervised By :mohammad ahmed 03/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 20 01:23:55 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 19 05:25:31 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.512	3.821	5992877	1878750	3.853m	1.874 #
2) SA Decachlor...	10.237	8.865	2432262	2089171	2.290m	2.135
Target Compounds						
31) L7 AR-1260-1	7.263	6.396	2796156	2075470	47.604m	41.689
32) L7 AR-1260-2	7.517	6.583	8137252	3313649	99.183m	50.140 #
33) L7 AR-1260-3	7.876	6.736	2748922	2644830	41.804m	46.240m
34) L7 AR-1260-4	8.100	7.210	2879512	2184884	44.068m	42.704
35) L7 AR-1260-5	8.422	7.450	4944704	5396196	36.178	41.901

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP031925\
Data File : PP070697.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Mar 2025 14:32
Operator : YP\AJ
Sample : Q1598-01 10X
Misc :
ALS Vial : 16 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
ETGI-354


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/20/2025
Supervised By :mohammad ahmed 03/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 20 01:23:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031125.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 05:25:31 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm





#1 Tetrachloro-m-xylene

R.T.: 4.512 min
Delta R.T.: -0.012 min
Response: 5992877
Conc: 3.85 ng/ml

Instrument :


ECD_P

ClientSampleId :

ETGI-354


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/20/2025
Supervised By :mohammad ahmed 03/24/2025




#1 Tetrachloro-m-xylene

R.T.: 3.821 min
Delta R.T.: -0.007 min
Response: 1878750
Conc: 1.87 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.237 min
Delta R.T.: -0.015 min
Response: 2432262
Conc: 2.29 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.865 min
Delta R.T.: -0.016 min
Response: 2089171
Conc: 2.14 ng/ml




#31 AR-1260-1

R.T.: 7.263 min
Delta R.T.: -0.008 min
Response: 2796156
Conc: 47.60 ng/ml

Instrument :
ECD_P
ClientSampleId :
ETGI-354


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/20/2025
Supervised By :mohammad ahmed 03/24/2025




#31 AR-1260-1

R.T.: 6.396 min
Delta R.T.: -0.011 min
Response: 2075470
Conc: 41.69 ng/ml




#32 AR-1260-2

R.T.: 7.517 min
Delta R.T.: -0.008 min
Response: 8137252
Conc: 99.18 ng/ml m



#32 AR-1260-2

R.T.: 6.583 min
Delta R.T.: -0.012 min
Response: 3313649
Conc: 50.14 ng/ml




#33 AR-1260-3

R.T.: 7.876 min
Delta R.T.: -0.007 min
Response: 2748922
Conc: 41.80 ng/ml

Instrument :
ECD_P
ClientSampleId :
ETGI-354


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/20/2025
Supervised By :mohammad ahmed 03/24/2025




#33 AR-1260-3

R.T.: 6.736 min
Delta R.T.: -0.013 min
Response: 2644830
Conc: 46.24 ng/ml




#34 AR-1260-4

R.T.: 8.100 min
Delta R.T.: -0.007 min
Response: 2879512
Conc: 44.07 ng/ml



#34 AR-1260-4

R.T.: 7.210 min
Delta R.T.: -0.011 min
Response: 2184884
Conc: 42.70 ng/ml




#35 AR-1260-5

R.T.: 8.422 min
Delta R.T.: -0.007 min
Response: 4944704
Conc: 36.18 ng/ml

Instrument :
ECD_P
ClientSampleId :
ETGI-354

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/20/2025
Supervised By :mohammad ahmed 03/24/2025



#35 AR-1260-5

R.T.: 7.450 min
Delta R.T.: -0.012 min
Response: 5396196
Conc: 41.90 ng/ml