

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032024\
 Data File : PP063588.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Mar 2024 09:33
 Operator : YP\AJ
 Sample : PB159644BL
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 PB159644BL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 20 01:23:39 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031924.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Mar 19 06:23:02 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.488	3.638	45292245	45294032	18.981	18.971
2) SA Decachlor...	10.364	8.678	48621694	51083298	19.150	18.902

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032024\
Data File : PP063588.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Mar 2024 09:33
Operator : YP\AJ
Sample : PB159644BL
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PB159644BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 20 01:23:39 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031924.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 19 06:23:02 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.488 min
Delta R.T.: 0.000 min
Response: 45292245
Conc: 18.98 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB159644BL




#1 Tetrachloro-m-xylene

R.T.: 3.638 min
Delta R.T.: -0.001 min
Response: 45294032
Conc: 18.97 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.364 min
Delta R.T.: -0.002 min
Response: 48621694
Conc: 19.15 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.678 min
Delta R.T.: -0.005 min
Response: 51083298
Conc: 18.90 ng/ml