

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032119\
 Data File : PP029458.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 21 Mar 2019 20:51
 Operator : EI\MA
 Sample : HSTDICV750
 Misc : hexane
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP032119

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 22 09:24:39 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032119.M
 Quant Title : 8080.M
 QLast Update : Fri Mar 22 09:23:06 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 T	Dalapon	67.039	64.130 E3	4.3	98	0.00
2 T	3,5-DICHLORO BENZOIC ACID	63.171	63.859 E3	-1.1	106	0.00
3 T	4-Nitrophenol	65.513	62.590 E3	4.5	101	0.00
4 S	2,4-DCAA	32.715	36.340 E3	-11.1	108	0.00
5 T	DICAMBA	147.757	151.889 E3	-2.8	107	0.00
8 T	DICHLORPROP	34.231	36.988 E3	-8.1	110	0.00
9 T	2,4-D	38.673	40.931 E3	-5.8	109	0.00
10 T	Pentachlorophenol	681.489	718.414 E3	-5.4	110	0.00
11 T	2,4,5-TP (SILVEX)	243.204	253.489 E3	-4.2	110	0.00
12 T	2,4,5-T	226.520	230.025 E3	-1.5	108	0.00
13 T	2,4-DB	20.794	20.419 E3	1.8	105	0.00
14 T	DINOSEB	168.545	169.377 E3	-0.5	106	0.00
15 T	Picloram	327.617	312.267 E3	4.7	98	0.00
16 T	DCPA	317.265	318.803 E3	-0.5	107	0.00

Signal #2

1 T	Dalapon	54.715	53.652 E3	1.9	98	0.00
2 T	3,5-DICHLORO BENZOIC ACID	54.377	52.882 E3	2.7	97	0.00
3 T	4-Nitrophenol	25.324	24.294 E3	4.1	96	0.00
4 S	2,4-DCAA	42.941	42.053 E3	2.1	97	0.00
5 T	DICAMBA	146.693	144.774 E3	1.3	98	0.00
8 T	DICHLORPROP	40.497	40.172 E3	0.8	98	0.00
9 T	2,4-D	47.175	46.106 E3	2.3	96	0.00
10 T	Pentachlorophenol	547.155	551.735 E3	-0.8	100	0.00
11 T	2,4,5-TP (SILVEX)	206.987	206.232 E3	0.4	99	0.00
12 T	2,4,5-T	165.776	164.620 E3	0.7	97	0.00
13 T	2,4-DB	23.721	23.580 E3	0.6	97	0.00
14 T	DINOSEB	151.082	148.851 E3	1.5	98	0.00
15 T	Picloram	238.348	236.863 E3	0.6	98	0.00
16 T	DCPA	262.368	258.404 E3	1.5	98	0.00

Evaluate Continuing Calibration Report - Not Found

6 T	MCP	0.000	0.000	0.0	0#	-12.60#
7 T	MCPA	0.000	0.000	0.0	0#	-12.84#

Signal #2

6 T	MCP	0.000	0.000	0.0	0#	-12.63#
7 T	MCPA	0.000	0.000	0.0	0#	-13.02#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032119\
Data File : PP029458.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 21 Mar 2019 20:51
Operator : EI\MA
Sample : HSTDICV750
Misc : hexane
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP032119

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 22 09:24:39 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032119.M
Quant Title : 8080.M
QLast Update : Fri Mar 22 09:23:06 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)

(#) = Out of Range

SPCC's out = 0 CCC's out = 0