

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032122\
 Data File : PP044939.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Mar 2022 20:55
 Operator : YP\AJ
 Sample : N2033-03
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 VNJ-237

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 03/22/2022
 Supervised By :mohammad ahmed 03/23/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 22 06:05:37 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 17 04:32:40 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.897	3.158	24636500	21407520	13.521m	16.244
2) SA Decachlor...	10.021	8.513	21189308	20303862	16.461m	15.626m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032122\
Data File : PP044939.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Mar 2022 20:55
Operator : YP\AJ
Sample : N2033-03
Misc :
ALS Vial : 25 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
VNJ-237


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/22/2022
Supervised By :mohammad ahmed 03/23/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 22 06:05:37 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031722.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 17 04:32:40 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.897 min
Delta R.T.: 0.000 min
Response: 24636500
Conc: 13.52 ng/ml

Instrument :
ECD_P
ClientSampleId :
VNJ-237


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/22/2022
Supervised By :mohammad ahmed 03/23/2022




#1 Tetrachloro-m-xylene

R.T.: 3.158 min
Delta R.T.: 0.003 min
Response: 21407520
Conc: 16.24 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.021 min
Delta R.T.: -0.002 min
Response: 21189308
Conc: 16.46 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.513 min
Delta R.T.: -0.002 min
Response: 20303862
Conc: 15.63 ng/ml m