

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032124\
 Data File : PP063626.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Mar 2024 13:53
 Operator : YP\AJ
 Sample : P1805-01 10X
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 KMH304I-1-1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 20 14:52:31 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031924.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Mar 19 06:23:02 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.489	3.639	2909768	2919760	1.219	1.223
2)	SA Decachlor...	10.368	8.680	3935766	3348176	1.550	1.239
Target Compounds							
28)	L6 AR-1254-3	7.127	0.000	394930	0	2.822	N.D. #
32)	L7 AR-1260-2	0.000	6.396f	0	529405	N.D.	3.132 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032124\
Data File : PP063626.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Mar 2024 13:53
Operator : YP\AJ
Sample : P1805-01 10X
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
KMH304I-1-1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 20 14:52:31 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031924.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 19 06:23:02 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.489 min
Delta R.T.: 0.000 min
Response: 2909768
Conc: 1.22 ng/ml

Instrument :
ECD_P
ClientSampleId :
KMH304I-1-1




#1 Tetrachloro-m-xylene

R.T.: 3.639 min
Delta R.T.: 0.000 min
Response: 2919760
Conc: 1.22 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.368 min
Delta R.T.: 0.002 min
Response: 3935766
Conc: 1.55 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.680 min
Delta R.T.: -0.004 min
Response: 3348176
Conc: 1.24 ng/ml



#28 AR-1254-3


R.T.: 7.127 min
Delta R.T.: 0.000 min
Response: 394930
Conc: 2.82 ng/ml

Instrument :
ECD_P
ClientSampleId :
KMH304I-1-1




#28 AR-1254-3

R.T.: 0.000 min
Exp R.T. : 6.093 min
Response: 0
Conc: N.D.



#32 AR-1260-2

R.T.: 0.000 min
Exp R.T. : 7.552 min
Response: 0
Conc: N.D.



#32 AR-1260-2

R.T.: 6.396 min
Delta R.T.: -0.019 min
Response: 529405
Conc: 3.13 ng/ml