

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032321\
 Data File : PP034298.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 23 Mar 2021 19:40
 Operator : DD\AJ
 Sample : M1763-01
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 24 03:06:16 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031821.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Mar 20 01:42:16 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.848	4.261	839276	689902	9.991	15.590 #
2)	SA Decachlor...	10.772	9.574	762766	283653	9.720	10.749
Target Compounds							
3)	L1 AR-1016-1	6.150f	5.520	609023	9567	242.137	8.353 #
4)	L1 AR-1016-2	0.000	5.552	0	619573	N.D.	368.063 #
5)	L1 AR-1016-3	6.253	5.732	187011	341027	79.452	377.901 #
6)	L1 AR-1016-4	6.350	0.000	36164	0	19.013	N.D. #
7)	L1 AR-1016-5	6.657f	6.002	39218	91975	20.308	97.361 #
8)	L2 AR-1221-1	5.095	4.500f	715251	117049	787.713	249.527 #
9)	L2 AR-1221-2	5.203	4.633	511538	1120610	743.279	3159.988 #
10)	L2 AR-1221-3	5.298	4.706	1695683	113724	817.974	106.630 #
11)	L3 AR-1232-1	5.298	4.706	1695683	113724	1034.574	130.945 #
12)	L3 AR-1232-2	5.884	5.552	826456	619573	966.327	849.426
13)	L3 AR-1232-3	0.000	5.732	0	341027	N.D.	873.454 #
14)	L3 AR-1232-4	6.350	0.000	36164	0	46.343	N.D. #
15)	L3 AR-1232-5	6.455	6.002	55575	91975	101.501	251.483 #
16)	L4 AR-1242-1	6.150f	5.520	609023	9567	305.631	10.503 #
17)	L4 AR-1242-2	0.000	5.552	0	619573	N.D.	472.014 #
18)	L4 AR-1242-3	6.253	5.732	187011	341027	100.499	469.020 #
19)	L4 AR-1242-4	6.350	0.000	36164	0	23.989	N.D. #
20)	L4 AR-1242-5	7.131	6.374f	412803	22919	270.772	29.201 #
21)	L5 AR-1248-1	6.150f	5.520	609023	9567	407.338	13.908 #
22)	L5 AR-1248-2	6.455	0.000	55575	0	26.484	N.D. #
23)	L5 AR-1248-3	6.657f	0.000	39218	0	15.400	N.D. #
24)	L5 AR-1248-4	7.106	6.002	129745	91975	50.217	74.948 #
25)	L5 AR-1248-5	7.131	6.420	412803	170345	159.072	168.717
26)	L6 AR-1254-1	7.062	6.374f	147025	22919	53.718	13.667 #
27)	L6 AR-1254-2	7.297	6.549	519100	13645	124.662	9.409 #
28)	L6 AR-1254-3	7.687	6.975	89370	166491	20.577	72.700 #
30)	L6 AR-1254-5	8.408	7.644	448872	183100	134.463	99.163 #
31)	L7 AR-1260-1	7.849	7.103	343948	32440	107.910	20.809 #
32)	L7 AR-1260-2	8.111	7.308	36304	28025	9.660	15.663 #
33)	L7 AR-1260-3	8.475	7.438f	127317	221259	41.717	130.759 #
34)	L7 AR-1260-4	8.706	7.931	219082	17829	62.926	12.522 #
35)	L7 AR-1260-5	9.025	8.179	129264	75636	18.882	24.358 #
36)	L8 AR-1262-1	8.475	7.644	127317	183100	28.540	181.326 #
37)	L8 AR-1262-2	9.025	8.179	129264	75636	17.053	23.433 #
38)	L8 AR-1262-3	9.324	8.481f	71888	23238	13.484	16.961 #
39)	L8 AR-1262-4	9.434	8.524	18032	42212	4.184	17.432 #
40)	L8 AR-1262-5	10.058	9.028	391846	12534	137.971	11.921 #
41)	L9 AR-1268-1	9.324	8.481f	71888	23238	7.218	5.871
42)	L9 AR-1268-2	9.434	8.524	18032	42212	1.933	11.435 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032321\
 Data File : PP034298.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 23 Mar 2021 19:40
 Operator : DD\AJ
 Sample : M1763-01
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 24 03:06:16 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031821.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Mar 20 01:42:16 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
43) L9	AR-1268-3	0.000	8.752f	0	54503	N.D.	15.997 #
44) L9	AR-1268-4	10.058	9.028	391846	12534	125.873	10.415 #
45) L9	AR-1268-5	10.473	9.330	88181	29973	3.191	3.215



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032321\
Data File : PP034298.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 23 Mar 2021 19:40
Operator : DD\AJ
Sample : M1763-01
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 24 03:06:16 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031821.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Mar 20 01:42:16 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.848 min
Delta R.T.: -0.004 min
Response: 839276
Conc: 9.99 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.261 min
Delta R.T.: -0.002 min
Response: 689902
Conc: 15.59 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.772 min
Delta R.T.: -0.012 min
Response: 762766
Conc: 9.72 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.574 min
Delta R.T.: -0.011 min
Response: 283653
Conc: 10.75 ng/ml



#3 AR-1016-1


R.T.: 6.150 min
Delta R.T.: -0.017 min
Response: 609023
Conc: 242.14 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 5.520 min
Delta R.T.: 0.000 min
Response: 9567
Conc: 8.35 ng/ml




#4 AR-1016-2

R.T.: 0.000 min
Exp R.T. : 6.190 min
Response: 0
Conc: N.D.



#4 AR-1016-2


R.T.: 5.552 min
Delta R.T.: 0.012 min
Response: 619573
Conc: 368.06 ng/ml



#5 AR-1016-3


R.T.: 6.253 min
Delta R.T.: -0.003 min
Response: 187011
Conc: 79.45 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 5.732 min
Delta R.T.: -0.002 min
Response: 341027
Conc: 377.90 ng/ml




#6 AR-1016-4

R.T.: 6.350 min
Delta R.T.: -0.012 min
Response: 36164
Conc: 19.01 ng/ml



#6 AR-1016-4


R.T.: 0.000 min
Exp R.T. : 5.782 min
Response: 0
Conc: N.D.



#7 AR-1016-5


R.T.: 6.657 min
Delta R.T.: -0.019 min
Response: 39218
Conc: 20.31 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 6.002 min
Delta R.T.: -0.011 min
Response: 91975
Conc: 97.36 ng/ml




#8 AR-1221-1

R.T.: 5.095 min
Delta R.T.: -0.007 min
Response: 715251
Conc: 787.71 ng/ml



#8 AR-1221-1


R.T.: 4.500 min
Delta R.T.: -0.019 min
Response: 117049
Conc: 249.53 ng/ml



#9 AR-1221-2


R.T.: 5.203 min
Delta R.T.: 0.003 min
Response: 511538
Conc: 743.28 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 4.633 min
Delta R.T.: 0.014 min
Response: 1120610
Conc: 3159.99 ng/ml




#10 AR-1221-3

R.T.: 5.298 min
Delta R.T.: 0.011 min
Response: 1695683
Conc: 817.97 ng/ml



#10 AR-1221-3


R.T.: 4.706 min
Delta R.T.: -0.003 min
Response: 113724
Conc: 106.63 ng/ml



#11 AR-1232-1


R.T.: 5.298 min
Delta R.T.: 0.011 min
Response: 1695683
Conc: 1034.57 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 4.706 min
Delta R.T.: -0.003 min
Response: 113724
Conc: 130.94 ng/ml




#12 AR-1232-2

R.T.: 5.884 min
Delta R.T.: 0.012 min
Response: 826456
Conc: 966.33 ng/ml



#12 AR-1232-2


R.T.: 5.552 min
Delta R.T.: 0.012 min
Response: 619573
Conc: 849.43 ng/ml



#13 AR-1232-3


R.T.: 0.000 min
Exp R.T.: 6.190 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 5.732 min
Delta R.T.: -0.001 min
Response: 341027
Conc: 873.45 ng/ml




#14 AR-1232-4

R.T.: 6.350 min
Delta R.T.: -0.012 min
Response: 36164
Conc: 46.34 ng/ml



#14 AR-1232-4


R.T.: 0.000 min
Exp R.T.: 5.827 min
Response: 0
Conc: N.D.



#15 AR-1232-5


R.T.: 6.455 min
Delta R.T.: -0.005 min
Response: 55575
Conc: 101.50 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 6.002 min
Delta R.T.: -0.011 min
Response: 91975
Conc: 251.48 ng/ml




#16 AR-1242-1

R.T.: 6.150 min
Delta R.T.: -0.017 min
Response: 609023
Conc: 305.63 ng/ml



#16 AR-1242-1


R.T.: 5.520 min
Delta R.T.: 0.000 min
Response: 9567
Conc: 10.50 ng/ml



#17 AR-1242-2


R.T.: 0.000 min
Exp R.T.: 6.190 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 5.552 min
Delta R.T.: 0.013 min
Response: 619573
Conc: 472.01 ng/ml




#18 AR-1242-3

R.T.: 6.253 min
Delta R.T.: -0.002 min
Response: 187011
Conc: 100.50 ng/ml



#18 AR-1242-3


R.T.: 5.732 min
Delta R.T.: 0.000 min
Response: 341027
Conc: 469.02 ng/ml



#19 AR-1242-4


R.T.: 6.350 min
Delta R.T.: -0.011 min
Response: 36164
Conc: 23.99 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 0.000 min
Exp R.T. : 5.827 min
Response: 0
Conc: N.D.




#20 AR-1242-5

R.T.: 7.131 min
Delta R.T.: -0.010 min
Response: 412803
Conc: 270.77 ng/ml



#20 AR-1242-5


R.T.: 6.374 min
Delta R.T.: -0.015 min
Response: 22919
Conc: 29.20 ng/ml



#21 AR-1248-1


R.T.: 6.150 min
Delta R.T.: -0.017 min
Response: 609023
Conc: 407.34 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 5.520 min
Delta R.T.: 0.000 min
Response: 9567
Conc: 13.91 ng/ml




#22 AR-1248-2

R.T.: 6.455 min
Delta R.T.: -0.005 min
Response: 55575
Conc: 26.48 ng/ml



#22 AR-1248-2


R.T.: 0.000 min
Exp R.T.: 5.781 min
Response: 0
Conc: N.D.



#23 AR-1248-3


R.T.: 6.657 min
Delta R.T.: -0.019 min
Response: 39218
Conc: 15.40 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 0.000 min
Exp R.T. : 5.827 min
Response: 0
Conc: N.D.




#24 AR-1248-4

R.T.: 7.106 min
Delta R.T.: 0.003 min
Response: 129745
Conc: 50.22 ng/ml



#24 AR-1248-4


R.T.: 6.002 min
Delta R.T.: -0.010 min
Response: 91975
Conc: 74.95 ng/ml



#25 AR-1248-5


R.T.: 7.131 min
Delta R.T.: -0.011 min
Response: 412803
Conc: 159.07 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 6.420 min
Delta R.T.: -0.014 min
Response: 170345
Conc: 168.72 ng/ml




#26 AR-1254-1

R.T.: 7.062 min
Delta R.T.: -0.010 min
Response: 147025
Conc: 53.72 ng/ml



#26 AR-1254-1


R.T.: 6.374 min
Delta R.T.: -0.016 min
Response: 22919
Conc: 13.67 ng/ml



#27 AR-1254-2


R.T.: 7.297 min
Delta R.T.: -0.004 min
Response: 519100
Conc: 124.66 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 6.549 min
Delta R.T.: 0.002 min
Response: 13645
Conc: 9.41 ng/ml




#28 AR-1254-3

R.T.: 7.687 min
Delta R.T.: 0.005 min
Response: 89370
Conc: 20.58 ng/ml



#28 AR-1254-3


R.T.: 6.975 min
Delta R.T.: 0.007 min
Response: 166491
Conc: 72.70 ng/ml



#30 AR-1254-5


R.T.: 8.408 min
Delta R.T.: 0.003 min
Response: 448872
Conc: 134.46 ng/ml

Instrument :
ECD_P
ClientSampleId :




#30 AR-1254-5

R.T.: 7.644 min
Delta R.T.: 0.012 min
Response: 183100
Conc: 99.16 ng/ml




#31 AR-1260-1

R.T.: 7.849 min
Delta R.T.: 0.000 min
Response: 343948
Conc: 107.91 ng/ml



#31 AR-1260-1


R.T.: 7.103 min
Delta R.T.: 0.000 min
Response: 32440
Conc: 20.81 ng/ml



#32 AR-1260-2


R.T.: 8.111 min
Delta R.T.: -0.003 min
Response: 36304
Conc: 9.66 ng/ml

Instrument :
ECD_P
ClientSampleId :




#32 AR-1260-2

R.T.: 7.308 min
Delta R.T.: 0.011 min
Response: 28025
Conc: 15.66 ng/ml




#33 AR-1260-3

R.T.: 8.475 min
Delta R.T.: -0.003 min
Response: 127317
Conc: 41.72 ng/ml



#33 AR-1260-3


R.T.: 7.438 min
Delta R.T.: -0.016 min
Response: 221259
Conc: 130.76 ng/ml



#34 AR-1260-4


R.T.: 8.706 min
Delta R.T.: -0.002 min
Response: 219082
Conc: 62.93 ng/ml

Instrument :
ECD_P
ClientSampleId :




#34 AR-1260-4

R.T.: 7.931 min
Delta R.T.: -0.004 min
Response: 17829
Conc: 12.52 ng/ml




#35 AR-1260-5

R.T.: 9.025 min
Delta R.T.: -0.003 min
Response: 129264
Conc: 18.88 ng/ml



#35 AR-1260-5


R.T.: 8.179 min
Delta R.T.: 0.001 min
Response: 75636
Conc: 24.36 ng/ml



#36 AR-1262-1


R.T.: 8.475 min
Delta R.T.: -0.004 min
Response: 127317
Conc: 28.54 ng/ml

Instrument :
ECD_P
ClientSampleId :




#36 AR-1262-1

R.T.: 7.644 min
Delta R.T.: 0.014 min
Response: 183100
Conc: 181.33 ng/ml




#37 AR-1262-2

R.T.: 9.025 min
Delta R.T.: -0.003 min
Response: 129264
Conc: 17.05 ng/ml



#37 AR-1262-2


R.T.: 8.179 min
Delta R.T.: 0.002 min
Response: 75636
Conc: 23.43 ng/ml



#38 AR-1262-3


R.T.: 9.324 min
Delta R.T.: -0.006 min
Response: 71888
Conc: 13.48 ng/ml

Instrument :
ECD_P
ClientSampleId :




#38 AR-1262-3

R.T.: 8.481 min
Delta R.T.: 0.019 min
Response: 23238
Conc: 16.96 ng/ml




#39 AR-1262-4

R.T.: 9.434 min
Delta R.T.: 0.014 min
Response: 18032
Conc: 4.18 ng/ml



#39 AR-1262-4


R.T.: 8.524 min
Delta R.T.: -0.002 min
Response: 42212
Conc: 17.43 ng/ml



#40 AR-1262-5


R.T.: 10.058 min
Delta R.T.: -0.008 min
Response: 391846
Conc: 137.97 ng/ml

Instrument :
ECD_P
ClientSampleId :




#40 AR-1262-5

R.T.: 9.028 min
Delta R.T.: 0.004 min
Response: 12534
Conc: 11.92 ng/ml




#41 AR-1268-1

R.T.: 9.324 min
Delta R.T.: -0.004 min
Response: 71888
Conc: 7.22 ng/ml



#41 AR-1268-1


R.T.: 8.481 min
Delta R.T.: 0.019 min
Response: 23238
Conc: 5.87 ng/ml



#42 AR-1268-2


R.T.: 9.434 min
Delta R.T.: 0.013 min
Response: 18032
Conc: 1.93 ng/ml

Instrument :
ECD_P
ClientSampleId :




#42 AR-1268-2

R.T.: 8.524 min
Delta R.T.: -0.002 min
Response: 42212
Conc: 11.44 ng/ml




#43 AR-1268-3

R.T.: 0.000 min
Exp R.T. : 9.639 min
Response: 0
Conc: N.D.



#43 AR-1268-3


R.T.: 8.752 min
Delta R.T.: 0.018 min
Response: 54503
Conc: 16.00 ng/ml



#44 AR-1268-4


R.T.: 10.058 min
Delta R.T.: -0.005 min
Response: 391846
Conc: 125.87 ng/ml

Instrument :
ECD_P
ClientSampleId :




#44 AR-1268-4

R.T.: 9.028 min
Delta R.T.: 0.004 min
Response: 12534
Conc: 10.42 ng/ml



#45 AR-1268-5

R.T.: 10.473 min
Delta R.T.: 0.012 min
Response: 88181
Conc: 3.19 ng/ml



#45 AR-1268-5

R.T.: 9.330 min
Delta R.T.: 0.008 min
Response: 29973
Conc: 3.21 ng/ml