

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032321\
 Data File : PP034299.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 23 Mar 2021 20:14
 Operator : DD\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 24 03:06:42 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031821.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Mar 20 01:42:16 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.849	4.260	3817049	2267886	45.438	51.249
2)	SA Decachlor...	10.772	9.574	3841136	1394884	48.946	52.858
Target Compounds							
3)	L1 AR-1016-1	6.161	5.516	1144265	607856	454.940	530.699
4)	L1 AR-1016-2	6.185	5.536	1714723	906327	452.894	538.412
5)	L1 AR-1016-3	6.250	5.728	1045600	493751	444.226	547.138
6)	L1 AR-1016-4	6.357	5.778	875827	384458	460.471	523.558
7)	L1 AR-1016-5	6.670	6.008	860655	444382	445.672	470.402
8)	L2 AR-1221-1	5.094	4.515	117127	72134	128.993	153.777
9)	L2 AR-1221-2	5.196	4.616	177838	102840	258.404	289.995
10)	L2 AR-1221-3	5.282	4.705	632330	380091	305.027	356.381
11)	L3 AR-1232-1	5.282	4.705	632330	380091	385.799	437.648
12)	L3 AR-1232-2	5.867	5.536	863129	906327	1009.206	1242.562
13)	L3 AR-1232-3	6.185	5.728	1714723	493751	1102.510	1264.617
14)	L3 AR-1232-4	6.357	5.823	875827	476712	1122.344	1373.548
15)	L3 AR-1232-5	6.455	6.008	680261	444382	1242.412	1215.049
16)	L4 AR-1242-1	6.161	5.516	1144265	607856	574.235	667.353
17)	L4 AR-1242-2	6.185	5.536	1714723	906327	576.708	690.474
18)	L4 AR-1242-3	6.250	5.728	1045600	493751	561.899	679.063
19)	L4 AR-1242-4	6.357	5.823	875827	476712	580.974	658.063
20)	L4 AR-1242-5	7.137	6.385	201937	346490	132.458	441.474 #
21)	L5 AR-1248-1	6.161	5.516	1144265	607856	765.328	883.649
22)	L5 AR-1248-2	6.455	5.778	680261	384458	324.175	385.000
23)	L5 AR-1248-3	6.670	5.823	860655	476712	337.954	452.044 #
24)	L5 AR-1248-4	7.097	6.008	200347	444382	77.543	362.116 #
25)	L5 AR-1248-5	7.137	6.427	201937	57442	77.816	56.893 #
26)	L6 AR-1254-1	7.067	6.385	592720	346490	216.560	206.627
27)	L6 AR-1254-2	7.295	6.542	687113	310567	165.011	214.157 #
28)	L6 AR-1254-3	7.675	6.979	423603	595063	97.532	259.839 #
29)	L6 AR-1254-4	7.970	7.198	299946	82379	103.318	68.893 #
30)	L6 AR-1254-5	8.397	7.624	2128000	1007831	637.460	545.819
31)	L7 AR-1260-1	7.842	7.097	1562139	777206	490.106	498.546
32)	L7 AR-1260-2	8.108	7.291	1814256	940780	482.777	525.781
33)	L7 AR-1260-3	8.472	7.448	1431296	864696	468.983	511.017
34)	L7 AR-1260-4	8.702	7.927	1770619	711527	508.572	499.754
35)	L7 AR-1260-5	9.020	8.170	3598542	1680802	525.660	541.288
36)	L8 AR-1262-1	8.472	7.624	1431296	1007831	320.848	998.070 #
37)	L8 AR-1262-2	9.020	8.170	3598542	1680802	474.746	520.736
38)	L8 AR-1262-3	9.342	8.456	2184724	354138	409.796	258.482 #
39)	L8 AR-1262-4	9.412	8.519	650242	1213729	150.881	501.220 #
40)	L8 AR-1262-5	10.056	9.016	1591500	466551	560.375	443.764
41)	L9 AR-1268-1	9.342	8.456	2184724	354138	219.376	89.469 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032321\
 Data File : PP034299.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 23 Mar 2021 20:14
 Operator : DD\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 24 03:06:42 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031821.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Mar 20 01:42:16 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	9.412	8.519	650242	1213729	69.690	328.799 #
43) L9	AR-1268-3	9.628	8.728	97602	32920	10.975	9.662
44) L9	AR-1268-4	10.056	9.016	1591500	466551	511.239	387.700
45) L9	AR-1268-5	10.452	9.313	315928	123032	11.433	13.196



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032321\
Data File : PP034299.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 23 Mar 2021 20:14
Operator : DD\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 24 03:06:42 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031821.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Mar 20 01:42:16 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.849 min
Delta R.T.: -0.003 min
Response: 3817049
Conc: 45.44 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.260 min
Delta R.T.: -0.003 min
Response: 2267886
Conc: 51.25 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.772 min
Delta R.T.: -0.012 min
Response: 3841136
Conc: 48.95 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.574 min
Delta R.T.: -0.011 min
Response: 1394884
Conc: 52.86 ng/ml



#3 AR-1016-1


R.T.: 6.161 min
Delta R.T.: -0.005 min
Response: 1144265
Conc: 454.94 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 5.516 min
Delta R.T.: -0.005 min
Response: 607856
Conc: 530.70 ng/ml




#4 AR-1016-2

R.T.: 6.185 min
Delta R.T.: -0.005 min
Response: 1714723
Conc: 452.89 ng/ml



#4 AR-1016-2


R.T.: 5.536 min
Delta R.T.: -0.005 min
Response: 906327
Conc: 538.41 ng/ml



#5 AR-1016-3


R.T.: 6.250 min
Delta R.T.: -0.006 min
Response: 1045600
Conc: 444.23 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 5.728 min
Delta R.T.: -0.005 min
Response: 493751
Conc: 547.14 ng/ml




#6 AR-1016-4

R.T.: 6.357 min
Delta R.T.: -0.004 min
Response: 875827
Conc: 460.47 ng/ml



#6 AR-1016-4


R.T.: 5.778 min
Delta R.T.: -0.005 min
Response: 384458
Conc: 523.56 ng/ml



#7 AR-1016-5


R.T.: 6.670 min
Delta R.T.: -0.006 min
Response: 860655
Conc: 445.67 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 6.008 min
Delta R.T.: -0.004 min
Response: 444382
Conc: 470.40 ng/ml




#8 AR-1221-1

R.T.: 5.094 min
Delta R.T.: -0.008 min
Response: 117127
Conc: 128.99 ng/ml



#8 AR-1221-1


R.T.: 4.515 min
Delta R.T.: -0.005 min
Response: 72134
Conc: 153.78 ng/ml



#9 AR-1221-2


R.T.: 5.196 min
Delta R.T.: -0.004 min
Response: 177838
Conc: 258.40 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 4.616 min
Delta R.T.: -0.003 min
Response: 102840
Conc: 290.00 ng/ml




#10 AR-1221-3

R.T.: 5.282 min
Delta R.T.: -0.004 min
Response: 632330
Conc: 305.03 ng/ml



#10 AR-1221-3


R.T.: 4.705 min
Delta R.T.: -0.003 min
Response: 380091
Conc: 356.38 ng/ml



#11 AR-1232-1


R.T.: 5.282 min
Delta R.T.: -0.005 min
Response: 632330
Conc: 385.80 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 4.705 min
Delta R.T.: -0.004 min
Response: 380091
Conc: 437.65 ng/ml




#12 AR-1232-2

R.T.: 5.867 min
Delta R.T.: -0.005 min
Response: 863129
Conc: 1009.21 ng/ml



#12 AR-1232-2


R.T.: 5.536 min
Delta R.T.: -0.004 min
Response: 906327
Conc: 1242.56 ng/ml



#13 AR-1232-3


R.T.: 6.185 min
Delta R.T.: -0.005 min
Response: 1714723
Conc: 1102.51 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 5.728 min
Delta R.T.: -0.005 min
Response: 493751
Conc: 1264.62 ng/ml




#14 AR-1232-4

R.T.: 6.357 min
Delta R.T.: -0.005 min
Response: 875827
Conc: 1122.34 ng/ml



#14 AR-1232-4


R.T.: 5.823 min
Delta R.T.: -0.004 min
Response: 476712
Conc: 1373.55 ng/ml



#15 AR-1232-5


R.T.: 6.455 min
Delta R.T.: -0.005 min
Response: 680261
Conc: 1242.41 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 6.008 min
Delta R.T.: -0.004 min
Response: 444382
Conc: 1215.05 ng/ml




#16 AR-1242-1

R.T.: 6.161 min
Delta R.T.: -0.005 min
Response: 1144265
Conc: 574.24 ng/ml



#16 AR-1242-1


R.T.: 5.516 min
Delta R.T.: -0.004 min
Response: 607856
Conc: 667.35 ng/ml



#17 AR-1242-2


R.T.: 6.185 min
Delta R.T.: -0.005 min
Response: 1714723
Conc: 576.71 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 5.536 min
Delta R.T.: -0.004 min
Response: 906327
Conc: 690.47 ng/ml




#18 AR-1242-3

R.T.: 6.250 min
Delta R.T.: -0.005 min
Response: 1045600
Conc: 561.90 ng/ml



#18 AR-1242-3


R.T.: 5.728 min
Delta R.T.: -0.004 min
Response: 493751
Conc: 679.06 ng/ml



#19 AR-1242-4


R.T.: 6.357 min
Delta R.T.: -0.004 min
Response: 875827
Conc: 580.97 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 5.823 min
Delta R.T.: -0.004 min
Response: 476712
Conc: 658.06 ng/ml




#20 AR-1242-5

R.T.: 7.137 min
Delta R.T.: -0.005 min
Response: 201937
Conc: 132.46 ng/ml



#20 AR-1242-5


R.T.: 6.385 min
Delta R.T.: -0.005 min
Response: 346490
Conc: 441.47 ng/ml



#21 AR-1248-1


R.T.: 6.161 min
Delta R.T.: -0.005 min
Response: 1144265
Conc: 765.33 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 5.516 min
Delta R.T.: -0.004 min
Response: 607856
Conc: 883.65 ng/ml




#22 AR-1248-2

R.T.: 6.455 min
Delta R.T.: -0.005 min
Response: 680261
Conc: 324.17 ng/ml



#22 AR-1248-2


R.T.: 5.778 min
Delta R.T.: -0.004 min
Response: 384458
Conc: 385.00 ng/ml



#23 AR-1248-3


R.T.: 6.670 min
Delta R.T.: -0.005 min
Response: 860655
Conc: 337.95 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 5.823 min
Delta R.T.: -0.004 min
Response: 476712
Conc: 452.04 ng/ml




#24 AR-1248-4

R.T.: 7.097 min
Delta R.T.: -0.006 min
Response: 200347
Conc: 77.54 ng/ml



#24 AR-1248-4


R.T.: 6.008 min
Delta R.T.: -0.003 min
Response: 444382
Conc: 362.12 ng/ml



#25 AR-1248-5


R.T.: 7.137 min
Delta R.T.: -0.006 min
Response: 201937
Conc: 77.82 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 6.427 min
Delta R.T.: -0.006 min
Response: 57442
Conc: 56.89 ng/ml




#26 AR-1254-1

R.T.: 7.067 min
Delta R.T.: -0.005 min
Response: 592720
Conc: 216.56 ng/ml



#26 AR-1254-1


R.T.: 6.385 min
Delta R.T.: -0.005 min
Response: 346490
Conc: 206.63 ng/ml



#27 AR-1254-2


R.T.: 7.295 min
Delta R.T.: -0.007 min
Response: 687113
Conc: 165.01 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 6.542 min
Delta R.T.: -0.005 min
Response: 310567
Conc: 214.16 ng/ml




#28 AR-1254-3

R.T.: 7.675 min
Delta R.T.: -0.007 min
Response: 423603
Conc: 97.53 ng/ml



#28 AR-1254-3


R.T.: 6.979 min
Delta R.T.: 0.011 min
Response: 595063
Conc: 259.84 ng/ml



#29 AR-1254-4


R.T.: 7.970 min
Delta R.T.: -0.007 min
Response: 299946
Conc: 103.32 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 7.198 min
Delta R.T.: -0.005 min
Response: 82379
Conc: 68.89 ng/ml




#30 AR-1254-5

R.T.: 8.397 min
Delta R.T.: -0.008 min
Response: 212800
Conc: 637.46 ng/ml



#30 AR-1254-5


R.T.: 7.624 min
Delta R.T.: -0.007 min
Response: 1007831
Conc: 545.82 ng/ml



#31 AR-1260-1


R.T.: 7.842 min
Delta R.T.: -0.007 min
Response: 1562139
Conc: 490.11 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 7.097 min
Delta R.T.: -0.006 min
Response: 777206
Conc: 498.55 ng/ml




#32 AR-1260-2

R.T.: 8.108 min
Delta R.T.: -0.007 min
Response: 1814256
Conc: 482.78 ng/ml



#32 AR-1260-2


R.T.: 7.291 min
Delta R.T.: -0.006 min
Response: 940780
Conc: 525.78 ng/ml



#33 AR-1260-3


R.T.: 8.472 min
Delta R.T.: -0.006 min
Response: 1431296
Conc: 468.98 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 7.448 min
Delta R.T.: -0.006 min
Response: 864696
Conc: 511.02 ng/ml




#34 AR-1260-4

R.T.: 8.702 min
Delta R.T.: -0.007 min
Response: 1770619
Conc: 508.57 ng/ml



#34 AR-1260-4


R.T.: 7.927 min
Delta R.T.: -0.007 min
Response: 711527
Conc: 499.75 ng/ml



#35 AR-1260-5


R.T.: 9.020 min
Delta R.T.: -0.008 min
Response: 3598542
Conc: 525.66 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 8.170 min
Delta R.T.: -0.007 min
Response: 1680802
Conc: 541.29 ng/ml




#36 AR-1262-1

R.T.: 8.472 min
Delta R.T.: -0.007 min
Response: 1431296
Conc: 320.85 ng/ml



#36 AR-1262-1


R.T.: 7.624 min
Delta R.T.: -0.006 min
Response: 1007831
Conc: 998.07 ng/ml



#37 AR-1262-2


R.T.: 9.020 min
Delta R.T.: -0.008 min
Response: 3598542
Conc: 474.75 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 8.170 min
Delta R.T.: -0.006 min
Response: 1680802
Conc: 520.74 ng/ml




#38 AR-1262-3

R.T.: 9.342 min
Delta R.T.: 0.011 min
Response: 2184724
Conc: 409.80 ng/ml



#38 AR-1262-3


R.T.: 8.456 min
Delta R.T.: -0.007 min
Response: 354138
Conc: 258.48 ng/ml



#39 AR-1262-4


R.T.: 9.412 min
Delta R.T.: -0.009 min
Response: 650242
Conc: 150.88 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 8.519 min
Delta R.T.: -0.008 min
Response: 1213729
Conc: 501.22 ng/ml




#40 AR-1262-5

R.T.: 10.056 min
Delta R.T.: -0.010 min
Response: 1591500
Conc: 560.37 ng/ml



#40 AR-1262-5


R.T.: 9.016 min
Delta R.T.: -0.007 min
Response: 466551
Conc: 443.76 ng/ml



#41 AR-1268-1


R.T.: 9.342 min
Delta R.T.: 0.014 min
Response: 2184724
Conc: 219.38 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 8.456 min
Delta R.T.: -0.006 min
Response: 354138
Conc: 89.47 ng/ml




#42 AR-1268-2

R.T.: 9.412 min
Delta R.T.: -0.010 min
Response: 650242
Conc: 69.69 ng/ml



#42 AR-1268-2


R.T.: 8.519 min
Delta R.T.: -0.008 min
Response: 1213729
Conc: 328.80 ng/ml



#43 AR-1268-3


R.T.: 9.628 min
Delta R.T.: -0.012 min
Response: 97602
Conc: 10.97 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 8.728 min
Delta R.T.: -0.006 min
Response: 32920
Conc: 9.66 ng/ml




#44 AR-1268-4

R.T.: 10.056 min
Delta R.T.: -0.008 min
Response: 1591500
Conc: 511.24 ng/ml



#44 AR-1268-4


R.T.: 9.016 min
Delta R.T.: -0.007 min
Response: 466551
Conc: 387.70 ng/ml



#45 AR-1268-5

R.T.: 10.452 min
Delta R.T.: -0.009 min
Response: 315928
Conc: 11.43 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 9.313 min
Delta R.T.: -0.008 min
Response: 123032
Conc: 13.20 ng/ml