

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032322\
 Data File : PP045019.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Mar 2022 12:21
 Operator : YP\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 23 13:23:20 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 17 04:32:40 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.900	3.154	46445617	35388852	25.491	26.854
2) SA Decachlor...	10.029	8.510	34918010	34858286	27.126	26.827
Target Compounds						
3) L1 AR-1016-1	5.147	4.303	52840	8170	0.900	0.161 #
4) L1 AR-1016-2	5.173	4.319	77136	32285	0.863	0.448 #
5) L1 AR-1016-3	5.248	4.501	117638	14888	2.260	0.378 #
6) L1 AR-1016-4	5.345	4.555	21190	16280	0.495	0.517
7) L1 AR-1016-5	5.658	4.775	64286	18145	1.465	0.455 #
8) L2 AR-1221-1	4.148	3.375	37835	83621	1.843	5.634 #
9) L2 AR-1221-2	4.246	3.466	158545	621631	10.371	50.177 #
10) L2 AR-1221-3	4.318	3.552	141439	17063	2.929	0.440 #
11) L3 AR-1232-1	4.298	3.552	104691	17063	2.327	0.475 #
12) L3 AR-1232-2	4.859	4.319	110041	32285	5.237	0.934 #
13) L3 AR-1232-3	5.173	4.501	77136	14888	1.849	0.795 #
14) L3 AR-1232-4	5.345	4.589	21190	35326	1.060	2.134 #
15) L3 AR-1232-5	5.454	4.775	133523	18145	9.449	0.981 #
16) L4 AR-1242-1	5.147	4.303	52840	8170	1.152	0.205 #
17) L4 AR-1242-2	5.173	4.319	77136	32285	1.101	0.569 #
18) L4 AR-1242-3	5.248	4.501	117638	14888	2.864	0.477 #
19) L4 AR-1242-4	5.345	4.589	21190	35326	0.616	1.166 #
20) L4 AR-1242-5	6.150	5.148	328547	188200	8.985	4.731 #
21) L5 AR-1248-1	5.147	4.303	52840	8170	1.502	0.259 #
22) L5 AR-1248-2	5.454	4.547	133523	42647	2.823	0.975 #
23) L5 AR-1248-3	5.658	4.589	64286	35326	1.088	0.773 #
24) L5 AR-1248-4	6.109	4.775	188211	18145	2.966	0.336 #
25) L5 AR-1248-5	6.150	5.196	328547	44482	5.266	0.792 #
26) L6 AR-1254-1	6.079	5.148	551752	188200	8.310	2.176 #
27) L6 AR-1254-2	6.315	5.310	267877	72556	2.681	0.975 #
28) L6 AR-1254-3	6.718	5.739	285240	87426	2.789	0.761 #
29) L6 AR-1254-4	7.040	5.990	244756	90334	3.365	1.212 #
30) L6 AR-1254-5	7.493	6.445	462722	59067	5.983	0.574 #
31) L7 AR-1260-1	6.887	5.884	34792	28932	0.505	0.411
32) L7 AR-1260-2	7.185	6.086	111063	57795	1.430	0.646 #
33) L7 AR-1260-3	7.563	6.247	119053	617481	1.892	7.250 #
34) L7 AR-1260-4	7.810	6.758	1225951	36533	16.712	0.544 #
35) L7 AR-1260-5	8.165	7.030	310346	22340	2.216	0.143 #
36) L8 AR-1262-1	7.563	6.485	119053	21340	1.256	0.207 #
37) L8 AR-1262-2	8.165	6.758	310346	36533	1.928	0.396 #
38) L8 AR-1262-3	8.489	7.333	186756	46479	1.731	0.639 #
39) L8 AR-1262-4	8.579	7.409	95767	73777	1.872	0.549 #
40) L8 AR-1262-5	9.265	7.946	42137	57308	0.754	0.870
41) L9 AR-1268-1	8.489	7.345	186756	41445	0.970	0.192 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032322\
Data File : PP045019.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Mar 2022 12:21
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 23 13:23:20 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031722.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 17 04:32:40 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42)	L9 AR-1268-2	8.579	7.409	95767	73777	0.549	0.377 #
43)	L9 AR-1268-3	8.824	7.627	76749	85262	0.503	0.518
44)	L9 AR-1268-4	9.265	7.946	42137	57308	0.673	0.748
45)	L9 AR-1268-5	9.692	8.248	131155	251971	0.272	0.489 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032322\
Data File : PP045019.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Mar 2022 12:21
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 23 13:23:20 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031722.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 17 04:32:40 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.900 min
Delta R.T.: 0.003 min
Response: 46445617
Conc: 25.49 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.154 min
Delta R.T.: 0.000 min
Response: 35388852
Conc: 26.85 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.029 min
Delta R.T.: 0.006 min
Response: 34918010
Conc: 27.13 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.510 min
Delta R.T.: -0.006 min
Response: 34858286
Conc: 26.83 ng/ml



#3 AR-1016-1


R.T.: 5.147 min
Delta R.T.: -0.003 min
Response: 52840
Conc: 0.90 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 4.303 min
Delta R.T.: 0.004 min
Response: 8170
Conc: 0.16 ng/ml




#4 AR-1016-2

R.T.: 5.173 min
Delta R.T.: 0.000 min
Response: 77136
Conc: 0.86 ng/ml



#4 AR-1016-2


R.T.: 4.319 min
Delta R.T.: 0.000 min
Response: 32285
Conc: 0.45 ng/ml



#5 AR-1016-3


R.T.: 5.248 min
Delta R.T.: 0.009 min
Response: 117638
Conc: 2.26 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 4.501 min
Delta R.T.: -0.002 min
Response: 14888
Conc: 0.38 ng/ml




#6 AR-1016-4

R.T.: 5.345 min
Delta R.T.: 0.001 min
Response: 21190
Conc: 0.50 ng/ml



#6 AR-1016-4


R.T.: 4.555 min
Delta R.T.: 0.003 min
Response: 16280
Conc: 0.52 ng/ml



#7 AR-1016-5


R.T.: 5.658 min
Delta R.T.: -0.007 min
Response: 64286
Conc: 1.46 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 4.775 min
Delta R.T.: 0.000 min
Response: 18145
Conc: 0.45 ng/ml




#8 AR-1221-1

R.T.: 4.148 min
Delta R.T.: 0.003 min
Response: 37835
Conc: 1.84 ng/ml



#8 AR-1221-1


R.T.: 3.375 min
Delta R.T.: -0.007 min
Response: 83621
Conc: 5.63 ng/ml



#9 AR-1221-2


R.T.: 4.246 min
Delta R.T.: 0.008 min
Response: 158545
Conc: 10.37 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.466 min
Delta R.T.: -0.005 min
Response: 621631
Conc: 50.18 ng/ml




#10 AR-1221-3

R.T.: 4.318 min
Delta R.T.: 0.000 min
Response: 141439
Conc: 2.93 ng/ml



#10 AR-1221-3


R.T.: 3.552 min
Delta R.T.: 0.001 min
Response: 17063
Conc: 0.44 ng/ml



#11 AR-1232-1


R.T.: 4.298 min
Delta R.T.: 0.001 min
Response: 104691
Conc: 2.33 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 3.552 min
Delta R.T.: 0.002 min
Response: 17063
Conc: 0.47 ng/ml




#12 AR-1232-2

R.T.: 4.859 min
Delta R.T.: 0.000 min
Response: 110041
Conc: 5.24 ng/ml



#12 AR-1232-2


R.T.: 4.319 min
Delta R.T.: 0.001 min
Response: 32285
Conc: 0.93 ng/ml



#13 AR-1232-3


R.T.: 5.173 min
Delta R.T.: 0.000 min
Response: 77136
Conc: 1.85 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 4.501 min
Delta R.T.: -0.002 min
Response: 14888
Conc: 0.79 ng/ml




#14 AR-1232-4

R.T.: 5.345 min
Delta R.T.: 0.000 min
Response: 21190
Conc: 1.06 ng/ml



#14 AR-1232-4


R.T.: 4.589 min
Delta R.T.: -0.004 min
Response: 35326
Conc: 2.13 ng/ml



#15 AR-1232-5


R.T.: 5.454 min
Delta R.T.: 0.007 min
Response: 133523
Conc: 9.45 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 4.775 min
Delta R.T.: 0.000 min
Response: 18145
Conc: 0.98 ng/ml




#16 AR-1242-1

R.T.: 5.147 min
Delta R.T.: -0.003 min
Response: 52840
Conc: 1.15 ng/ml



#16 AR-1242-1


R.T.: 4.303 min
Delta R.T.: 0.004 min
Response: 8170
Conc: 0.20 ng/ml



#17 AR-1242-2


R.T.: 5.173 min
Delta R.T.: 0.001 min
Response: 77136
Conc: 1.10 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 4.319 min
Delta R.T.: 0.001 min
Response: 32285
Conc: 0.57 ng/ml




#18 AR-1242-3

R.T.: 5.248 min
Delta R.T.: 0.010 min
Response: 117638
Conc: 2.86 ng/ml



#18 AR-1242-3


R.T.: 4.501 min
Delta R.T.: -0.001 min
Response: 14888
Conc: 0.48 ng/ml



#19 AR-1242-4


R.T.: 5.345 min
Delta R.T.: 0.000 min
Response: 21190
Conc: 0.62 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 4.589 min
Delta R.T.: -0.004 min
Response: 35326
Conc: 1.17 ng/ml




#20 AR-1242-5

R.T.: 6.150 min
Delta R.T.: 0.005 min
Response: 328547
Conc: 8.98 ng/ml



#20 AR-1242-5


R.T.: 5.148 min
Delta R.T.: -0.004 min
Response: 188200
Conc: 4.73 ng/ml



#21 AR-1248-1


R.T.: 5.147 min
Delta R.T.: -0.002 min
Response: 52840
Conc: 1.50 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.303 min
Delta R.T.: 0.004 min
Response: 8170
Conc: 0.26 ng/ml




#22 AR-1248-2

R.T.: 5.454 min
Delta R.T.: 0.007 min
Response: 133523
Conc: 2.82 ng/ml



#22 AR-1248-2


R.T.: 4.547 min
Delta R.T.: -0.004 min
Response: 42647
Conc: 0.98 ng/ml



#23 AR-1248-3


R.T.: 5.658 min
Delta R.T.: -0.007 min
Response: 64286
Conc: 1.09 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 4.589 min
Delta R.T.: -0.004 min
Response: 35326
Conc: 0.77 ng/ml




#24 AR-1248-4

R.T.: 6.109 min
Delta R.T.: 0.005 min
Response: 188211
Conc: 2.97 ng/ml



#24 AR-1248-4


R.T.: 4.775 min
Delta R.T.: 0.000 min
Response: 18145
Conc: 0.34 ng/ml



#25 AR-1248-5


R.T.: 6.150 min
Delta R.T.: 0.005 min
Response: 328547
Conc: 5.27 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.196 min
Delta R.T.: 0.001 min
Response: 44482
Conc: 0.79 ng/ml




#26 AR-1254-1

R.T.: 6.079 min
Delta R.T.: 0.005 min
Response: 551752
Conc: 8.31 ng/ml



#26 AR-1254-1


R.T.: 5.148 min
Delta R.T.: -0.003 min
Response: 188200
Conc: 2.18 ng/ml



#27 AR-1254-2


R.T.: 6.315 min
Delta R.T.: 0.000 min
Response: 267877
Conc: 2.68 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 5.310 min
Delta R.T.: -0.002 min
Response: 72556
Conc: 0.98 ng/ml




#28 AR-1254-3

R.T.: 6.718 min
Delta R.T.: 0.004 min
Response: 285240
Conc: 2.79 ng/ml



#28 AR-1254-3


R.T.: 5.739 min
Delta R.T.: -0.005 min
Response: 87426
Conc: 0.76 ng/ml



#29 AR-1254-4


R.T.: 7.040 min
Delta R.T.: 0.012 min
Response: 244756
Conc: 3.37 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 5.990 min
Delta R.T.: -0.002 min
Response: 90334
Conc: 1.21 ng/ml




#30 AR-1254-5

R.T.: 7.493 min
Delta R.T.: 0.005 min
Response: 462722
Conc: 5.98 ng/ml



#30 AR-1254-5


R.T.: 6.445 min
Delta R.T.: 0.001 min
Response: 59067
Conc: 0.57 ng/ml



#31 AR-1260-1


R.T.: 6.887 min
Delta R.T.: -0.006 min
Response: 34792
Conc: 0.51 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 5.884 min
Delta R.T.: 0.000 min
Response: 28932
Conc: 0.41 ng/ml




#32 AR-1260-2

R.T.: 7.185 min
Delta R.T.: 0.008 min
Response: 111063
Conc: 1.43 ng/ml



#32 AR-1260-2


R.T.: 6.086 min
Delta R.T.: -0.005 min
Response: 57795
Conc: 0.65 ng/ml



#33 AR-1260-3


R.T.: 7.563 min
Delta R.T.: -0.008 min
Response: 119053
Conc: 1.89 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.247 min
Delta R.T.: -0.006 min
Response: 617481
Conc: 7.25 ng/ml




#34 AR-1260-4

R.T.: 7.810 min
Delta R.T.: -0.009 min
Response: 1225951
Conc: 16.71 ng/ml



#34 AR-1260-4


R.T.: 6.758 min
Delta R.T.: -0.007 min
Response: 36533
Conc: 0.54 ng/ml



#35 AR-1260-5


R.T.: 8.165 min
Delta R.T.: 0.004 min
Response: 310346
Conc: 2.22 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.030 min
Delta R.T.: -0.003 min
Response: 22340
Conc: 0.14 ng/ml




#36 AR-1262-1

R.T.: 7.563 min
Delta R.T.: -0.007 min
Response: 119053
Conc: 1.26 ng/ml



#36 AR-1262-1


R.T.: 6.485 min
Delta R.T.: 0.000 min
Response: 21340
Conc: 0.21 ng/ml



#37 AR-1262-2


R.T.: 8.165 min
Delta R.T.: 0.004 min
Response: 310346
Conc: 1.93 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 6.758 min
Delta R.T.: -0.006 min
Response: 36533
Conc: 0.40 ng/ml




#38 AR-1262-3

R.T.: 8.489 min
Delta R.T.: -0.001 min
Response: 186756
Conc: 1.73 ng/ml



#38 AR-1262-3


R.T.: 7.333 min
Delta R.T.: -0.004 min
Response: 46479
Conc: 0.64 ng/ml



#39 AR-1262-4


R.T.: 8.579 min
Delta R.T.: -0.004 min
Response: 95767
Conc: 1.87 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 7.409 min
Delta R.T.: 0.000 min
Response: 73777
Conc: 0.55 ng/ml




#40 AR-1262-5

R.T.: 9.265 min
Delta R.T.: 0.002 min
Response: 42137
Conc: 0.75 ng/ml



#40 AR-1262-5


R.T.: 7.946 min
Delta R.T.: -0.005 min
Response: 57308
Conc: 0.87 ng/ml



#41 AR-1268-1


R.T.: 8.489 min
Delta R.T.: -0.001 min
Response: 186756
Conc: 0.97 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.345 min
Delta R.T.: 0.005 min
Response: 41445
Conc: 0.19 ng/ml




#42 AR-1268-2

R.T.: 8.579 min
Delta R.T.: -0.010 min
Response: 95767
Conc: 0.55 ng/ml



#42 AR-1268-2


R.T.: 7.409 min
Delta R.T.: -0.002 min
Response: 73777
Conc: 0.38 ng/ml



#43 AR-1268-3


R.T.: 8.824 min
Delta R.T.: 0.000 min
Response: 76749
Conc: 0.50 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.627 min
Delta R.T.: -0.004 min
Response: 85262
Conc: 0.52 ng/ml




#44 AR-1268-4

R.T.: 9.265 min
Delta R.T.: 0.000 min
Response: 42137
Conc: 0.67 ng/ml



#44 AR-1268-4


R.T.: 7.946 min
Delta R.T.: -0.006 min
Response: 57308
Conc: 0.75 ng/ml



#45 AR-1268-5

R.T.: 9.692 min
Delta R.T.: 0.006 min
Response: 131155
Conc: 0.27 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.248 min
Delta R.T.: -0.007 min
Response: 251971
Conc: 0.49 ng/ml