

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032322\
 Data File : PP045016.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Mar 2022 11:13
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 23 13:22:28 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 17 04:32:40 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.900	3.155	100.9E6	73399487	55.380	55.697
2) SA Decachlor...	10.031	8.510	74344035	70177094	57.753	54.008
Target Compounds						
16) L4 AR-1242-1	5.152	4.296	24591874	21695545	535.932	543.107
17) L4 AR-1242-2	5.175	4.315	34999272	30670157	499.783	540.351
18) L4 AR-1242-3	5.242	4.500	21321723	17206804	519.136	551.797
19) L4 AR-1242-4	5.347	4.592	17751179	16997264	516.081	560.968
20) L4 AR-1242-5	6.148	5.149	19472566	22081724	532.501	555.094



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032322\
Data File : PP045016.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Mar 2022 11:13
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 23 13:22:28 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031722.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 17 04:32:40 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.900 min
Delta R.T.: 0.003 min
Response: 100905957
Conc: 55.38 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.155 min
Delta R.T.: 0.000 min
Response: 73399487
Conc: 55.70 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.031 min
Delta R.T.: 0.008 min
Response: 74344035
Conc: 57.75 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.510 min
Delta R.T.: -0.005 min
Response: 70177094
Conc: 54.01 ng/ml



#16 AR-1242-1


R.T.: 5.152 min
Delta R.T.: 0.003 min
Response: 24591874
Conc: 535.93 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.296 min
Delta R.T.: -0.002 min
Response: 21695545
Conc: 543.11 ng/ml




#17 AR-1242-2

R.T.: 5.175 min
Delta R.T.: 0.003 min
Response: 34999272
Conc: 499.78 ng/ml



#17 AR-1242-2


R.T.: 4.315 min
Delta R.T.: -0.002 min
Response: 30670157
Conc: 540.35 ng/ml



#18 AR-1242-3


R.T.: 5.242 min
Delta R.T.: 0.003 min
Response: 21321723
Conc: 519.14 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.500 min
Delta R.T.: -0.002 min
Response: 17206804
Conc: 551.80 ng/ml




#19 AR-1242-4

R.T.: 5.347 min
Delta R.T.: 0.002 min
Response: 17751179
Conc: 516.08 ng/ml



#19 AR-1242-4


R.T.: 4.592 min
Delta R.T.: -0.002 min
Response: 16997264
Conc: 560.97 ng/ml



#20 AR-1242-5

R.T.: 6.148 min
Delta R.T.: 0.004 min
Response: 19472566
Conc: 532.50 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.149 min
Delta R.T.: -0.003 min
Response: 22081724
Conc: 555.09 ng/ml