

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032322\
 Data File : PP045017.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Mar 2022 11:31
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 23 13:22:46 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 17 04:32:40 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.901	3.155	106.2E6	79186495	58.259	60.088
2) SA Decachlor...	10.032	8.512	77859875	74033240	60.485	56.976
Target Compounds						
21) L5 AR-1248-1	5.153	4.297	19587106	17983662	556.907	569.448
22) L5 AR-1248-2	5.450	4.549	25941904	24773107	548.400	566.460
23) L5 AR-1248-3	5.668	4.591	30466617	25920570	515.629	567.259
24) L5 AR-1248-4	6.108	4.773	33923236	30740967	534.558	569.048
25) L5 AR-1248-5	6.149	5.193	33513605	31241456	537.120	556.462


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032322\
Data File : PP045017.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Mar 2022 11:31
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 23 13:22:46 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031722.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 17 04:32:40 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.901 min
Delta R.T.: 0.004 min
Response: 106151487
Conc: 58.26 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.155 min
Delta R.T.: 0.000 min
Response: 79186495
Conc: 60.09 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.032 min
Delta R.T.: 0.009 min
Response: 77859875
Conc: 60.48 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.512 min
Delta R.T.: -0.004 min
Response: 74033240
Conc: 56.98 ng/ml



#21 AR-1248-1


R.T.: 5.153 min
Delta R.T.: 0.004 min
Response: 19587106
Conc: 556.91 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.297 min
Delta R.T.: -0.002 min
Response: 17983662
Conc: 569.45 ng/ml




#22 AR-1248-2

R.T.: 5.450 min
Delta R.T.: 0.003 min
Response: 25941904
Conc: 548.40 ng/ml



#22 AR-1248-2


R.T.: 4.549 min
Delta R.T.: -0.002 min
Response: 24773107
Conc: 566.46 ng/ml



#23 AR-1248-3


R.T.: 5.668 min
Delta R.T.: 0.004 min
Response: 30466617
Conc: 515.63 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.591 min
Delta R.T.: -0.002 min
Response: 25920570
Conc: 567.26 ng/ml




#24 AR-1248-4

R.T.: 6.108 min
Delta R.T.: 0.004 min
Response: 33923236
Conc: 534.56 ng/ml



#24 AR-1248-4


R.T.: 4.773 min
Delta R.T.: -0.002 min
Response: 30740967
Conc: 569.05 ng/ml



#25 AR-1248-5

R.T.: 6.149 min
Delta R.T.: 0.004 min
Response: 33513605
Conc: 537.12 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.193 min
Delta R.T.: -0.002 min
Response: 31241456
Conc: 556.46 ng/ml