

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032421\
 Data File : PP034331.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 24 Mar 2021 15:22
 Operator : DD\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 24 16:26:16 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031821.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Mar 20 01:42:16 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.848	4.261	3881900	2176976	46.209	49.194
2) SA Decachlor...	10.771	9.574	3934845	1436813	50.140	54.447
Target Compounds						
26) L6 AR-1254-1	7.066	6.385	1293846	863139	472.728	514.727
27) L6 AR-1254-2	7.295	6.542	2013013	747245	483.428	515.277
28) L6 AR-1254-3	7.676	6.962	2190402	1204958	504.328	526.154
29) L6 AR-1254-4	7.971	7.198	1507936	678209	519.415	567.181
30) L6 AR-1254-5	8.398	7.625	1733841	1038718	519.386	562.547



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032421\
Data File : PP034331.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 24 Mar 2021 15:22
Operator : DD\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 24 16:26:16 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031821.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Mar 20 01:42:16 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.848 min
Delta R.T.: -0.004 min
Response: 3881900
Conc: 46.21 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.261 min
Delta R.T.: -0.002 min
Response: 2176976
Conc: 49.19 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.771 min
Delta R.T.: -0.013 min
Response: 3934845
Conc: 50.14 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.574 min
Delta R.T.: -0.011 min
Response: 1436813
Conc: 54.45 ng/ml



#26 AR-1254-1


R.T.: 7.066 min
Delta R.T.: -0.006 min
Response: 1293846
Conc: 472.73 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 6.385 min
Delta R.T.: -0.005 min
Response: 863139
Conc: 514.73 ng/ml




#27 AR-1254-2

R.T.: 7.295 min
Delta R.T.: -0.006 min
Response: 2013013
Conc: 483.43 ng/ml



#27 AR-1254-2


R.T.: 6.542 min
Delta R.T.: -0.005 min
Response: 747245
Conc: 515.28 ng/ml



#28 AR-1254-3


R.T.: 7.676 min
Delta R.T.: -0.006 min
Response: 2190402
Conc: 504.33 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.962 min
Delta R.T.: -0.006 min
Response: 1204958
Conc: 526.15 ng/ml




#29 AR-1254-4

R.T.: 7.971 min
Delta R.T.: -0.007 min
Response: 1507936
Conc: 519.42 ng/ml



#29 AR-1254-4


R.T.: 7.198 min
Delta R.T.: -0.006 min
Response: 678209
Conc: 567.18 ng/ml



#30 AR-1254-5

R.T.: 8.398 min
Delta R.T.: -0.007 min
Response: 1733841
Conc: 519.39 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 7.625 min
Delta R.T.: -0.007 min
Response: 1038718
Conc: 562.55 ng/ml