

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032519\
 Data File : PP029491.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 25 Mar 2019 14:23
 Operator : EI\MA
 Sample : K2018-05
 Misc : hexane
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 25 16:47:39 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032119.M
 Quant Title : 8080.M
 QLast Update : Fri Mar 22 09:23:06 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.926	12.118	12346560	16807139	377.398	391.398
Target Compounds							
1) T	Dalapon	3.268	2.974	141580	8907154	2.112	162.791 #
8) T	DICHLORPROP	0.000	13.549	0	396170	N.D.	9.783 #
10) T	Pentachlo...	14.206	14.635	9403912	6953621	13.799	12.709
11) T	2,4,5-TP ...	15.059	15.275	478837	296998	1.969	1.435 #
12) T	2,4,5-T	15.429	15.866	3795047	9032685	16.754	54.487 #
13) T	2,4-DB	16.218	16.388	18571129	304860	893.104	12.852 #
14) T	DINOSEB	17.659	16.857	55642	1281410	0.330	8.482 #
15) T	Picloram	17.475	18.280	267147	3353999	0.815	14.072 #
16) T	DCPA	18.028	18.099	6375623	2345993	20.096	8.942 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032519\
Data File : PP029491.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 25 Mar 2019 14:23
Operator : EI\MA
Sample : K2018-05
Misc : hexane
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 25 16:47:39 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032119.M
Quant Title : 8080.M
QLast Update : Fri Mar 22 09:23:06 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 3.268 min
Delta R.T.: 0.027 min
Response: 141580
Conc: 2.11 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Dalapon

R.T.: 2.974 min
Delta R.T.: 0.017 min
Response: 8907154
Conc: 162.79 ng/ml




#4 2,4-DCAA

R.T.: 11.926 min
Delta R.T.: 0.000 min
Response: 12346560
Conc: 377.40 ng/ml



#4 2,4-DCAA


R.T.: 12.118 min
Delta R.T.: -0.003 min
Response: 16807139
Conc: 391.40 ng/ml



#8 DICHLORPROP


R.T.: 0.000 min
Exp R.T.: 13.438 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#8 DICHLORPROP

R.T.: 13.549 min
Delta R.T.: 0.004 min
Response: 396170
Conc: 9.78 ng/ml




#10 Pentachlorophenol

R.T.: 14.206 min
Delta R.T.: 0.000 min
Response: 9403912
Conc: 13.80 ng/ml



#10 Pentachlorophenol


R.T.: 14.635 min
Delta R.T.: -0.005 min
Response: 6953621
Conc: 12.71 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 15.059 min
Delta R.T.: -0.003 min
Response: 478837
Conc: 1.97 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 2,4,5-TP (SILVEX)

R.T.: 15.275 min
Delta R.T.: 0.065 min
Response: 296998
Conc: 1.43 ng/ml




#12 2,4,5-T

R.T.: 15.429 min
Delta R.T.: -0.021 min
Response: 3795047
Conc: 16.75 ng/ml



#12 2,4,5-T


R.T.: 15.866 min
Delta R.T.: 0.008 min
Response: 9032685
Conc: 54.49 ng/ml



#13 2,4-DB


R.T.: 16.218 min
Delta R.T.: 0.011 min
Response: 18571129
Conc: 893.10 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 2,4-DB

R.T.: 16.388 min
Delta R.T.: -0.067 min
Response: 304860
Conc: 12.85 ng/ml




#14 DINOSEB

R.T.: 17.659 min
Delta R.T.: -0.013 min
Response: 55642
Conc: 0.33 ng/ml



#14 DINOSEB


R.T.: 16.857 min
Delta R.T.: 0.013 min
Response: 1281410
Conc: 8.48 ng/ml



#15 Picloram


R.T.: 17.475 min
Delta R.T.: 0.015 min
Response: 267147
Conc: 0.82 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 Picloram

R.T.: 18.280 min
Delta R.T.: 0.038 min
Response: 3353999
Conc: 14.07 ng/ml



#16 DCPA

R.T.: 18.028 min
Delta R.T.: -0.027 min
Response: 6375623
Conc: 20.10 ng/ml



#16 DCPA

R.T.: 18.099 min
Delta R.T.: -0.017 min
Response: 2345993
Conc: 8.94 ng/ml