

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032521\
 Data File : PP034358.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 25 Mar 2021 12:55
 Operator : DD\AJ
 Sample : M1791-01
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 26 03:04:17 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031821.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Mar 20 01:42:16 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.848	4.260	1768362	983645	21.050	22.228
2)	SA Decachlor...	10.772	9.573	1915235	706481	24.405	26.772
Target Compounds							
9)	L2 AR-1221-2	5.198	4.614	27055	12378	39.312	34.904
27)	L6 AR-1254-2	7.297	0.000	13795	0	3.313	N.D. #
28)	L6 AR-1254-3	7.693	0.000	19626	0	4.519	N.D. #
29)	L6 AR-1254-4	7.974	0.000	9511	0	3.276	N.D. #
30)	L6 AR-1254-5	8.406	0.000	33229	0	9.954	N.D. #
35)	L7 AR-1260-5	9.031	8.167	19387	7629	2.832	2.457
37)	L8 AR-1262-2	9.031	8.167	19387	7629	2.558	2.364
40)	L8 AR-1262-5	10.065	0.000	973	0	0.343	N.D. #
44)	L9 AR-1268-4	10.065	0.000	973	0	0.313	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032521\
Data File : PP034358.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 25 Mar 2021 12:55
Operator : DD\AJ
Sample : M1791-01
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 26 03:04:17 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031821.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Mar 20 01:42:16 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.848 min
Delta R.T.: -0.004 min
Response: 1768362
Conc: 21.05 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.260 min
Delta R.T.: -0.003 min
Response: 983645
Conc: 22.23 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.772 min
Delta R.T.: -0.012 min
Response: 1915235
Conc: 24.41 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.573 min
Delta R.T.: -0.012 min
Response: 706481
Conc: 26.77 ng/ml



#9 AR-1221-2


R.T.: 5.198 min
Delta R.T.: -0.002 min
Response: 27055
Conc: 39.31 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 4.614 min
Delta R.T.: -0.005 min
Response: 12378
Conc: 34.90 ng/ml




#27 AR-1254-2

R.T.: 7.297 min
Delta R.T.: -0.005 min
Response: 13795
Conc: 3.31 ng/ml



#27 AR-1254-2


R.T.: 0.000 min
Exp R.T.: 6.547 min
Response: 0
Conc: N.D.



#28 AR-1254-3


R.T.: 7.693 min
Delta R.T.: 0.010 min
Response: 19626
Conc: 4.52 ng/ml

Instrument :
ECD_P
ClientSampleId :




#28 AR-1254-3

R.T.: 0.000 min
Exp R.T. : 6.968 min
Response: 0
Conc: N.D.




#29 AR-1254-4

R.T.: 7.974 min
Delta R.T.: -0.003 min
Response: 9511
Conc: 3.28 ng/ml



#29 AR-1254-4


R.T.: 0.000 min
Exp R.T. : 7.204 min
Response: 0
Conc: N.D.



#30 AR-1254-5


R.T.: 8.406 min
Delta R.T.: 0.001 min
Response: 33229
Conc: 9.95 ng/ml

Instrument :
ECD_P
ClientSampleId :




#30 AR-1254-5

R.T.: 0.000 min
Exp R.T. : 7.632 min
Response: 0
Conc: N.D.




#35 AR-1260-5

R.T.: 9.031 min
Delta R.T.: 0.003 min
Response: 19387
Conc: 2.83 ng/ml



#35 AR-1260-5


R.T.: 8.167 min
Delta R.T.: -0.010 min
Response: 7629
Conc: 2.46 ng/ml



#37 AR-1262-2


R.T.: 9.031 min
Delta R.T.: 0.003 min
Response: 19387
Conc: 2.56 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 8.167 min
Delta R.T.: -0.009 min
Response: 7629
Conc: 2.36 ng/ml




#40 AR-1262-5

R.T.: 10.065 min
Delta R.T.: 0.000 min
Response: 973
Conc: 0.34 ng/ml



#40 AR-1262-5


R.T.: 0.000 min
Exp R.T.: 9.024 min
Response: 0
Conc: N.D.



#44 AR-1268-4

R.T.: 10.065 min
Delta R.T.: 0.002 min
Response: 973
Conc: 0.31 ng/ml

Instrument :
ECD_P
ClientSampleId :



#44 AR-1268-4

R.T.: 0.000 min
Exp R.T. : 9.024 min
Response: 0
Conc: N.D.