

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032522\
 Data File : PP045183.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Mar 2022 13:27
 Operator : YP\AJ
 Sample : PB143647BL
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 26 21:04:34 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 17 04:32:40 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.898	3.153	40039764	31978546	21.975	24.266
2)	SA Decachlor...	10.022	8.507	28213074	29396562	21.917	22.624
Target Compounds							
3)	L1 AR-1016-1	5.149	4.302	87336	23045	1.488	0.455 #
4)	L1 AR-1016-2	5.175	4.317	112676	28765	1.260	0.399 #
5)	L1 AR-1016-3	5.236	4.509	128080	38574	2.461	0.980 #
6)	L1 AR-1016-4	5.352	4.554	52149	33415	1.219	1.061
7)	L1 AR-1016-5	5.666	4.775	24589	13119	0.560	0.329 #
8)	L2 AR-1221-1	4.126f	3.395	1195665	7923	58.255	0.534 #
9)	L2 AR-1221-2	4.248	3.459	40282	2467971	2.635	199.209 #
10)	L2 AR-1221-3	4.312	3.550	42895	24407	0.888	0.629 #
11)	L3 AR-1232-1	4.295	3.550	13161	24407	0.292	0.679 #
12)	L3 AR-1232-2	4.865	4.317	145231	28765	6.911	0.832 #
13)	L3 AR-1232-3	5.175	4.509	112676	38574	2.701	2.060
14)	L3 AR-1232-4	5.352	4.598	52149	17081	2.608	1.032 #
15)	L3 AR-1232-5	5.443	4.775	35612	13119	2.520	0.710 #
16)	L4 AR-1242-1	5.149	4.302	87336	23045	1.903	0.577 #
17)	L4 AR-1242-2	5.175	4.317	112676	28765	1.609	0.507 #
18)	L4 AR-1242-3	5.236	4.509	128080	38574	3.118	1.237 #
19)	L4 AR-1242-4	5.352	4.598	52149	17081	1.516	0.564 #
20)	L4 AR-1242-5	6.143	5.156	76012	25071	2.079	0.630 #
21)	L5 AR-1248-1	5.149	4.302	87336	23045	2.483	0.730 #
22)	L5 AR-1248-2	5.443	4.554	35612	33415	0.753	0.764
23)	L5 AR-1248-3	5.666	4.598	24589	17081	0.416	0.374
24)	L5 AR-1248-4	6.098	4.775	16935	13119	0.267	0.243
25)	L5 AR-1248-5	6.143	5.192	76012	21553	1.218	0.384 #
26)	L6 AR-1254-1	6.079	5.148	19335	105193	0.291	1.216 #
27)	L6 AR-1254-2	6.319	5.316	188620	52935	1.888	0.711 #
28)	L6 AR-1254-3	6.716	5.746	517916	21901	5.064	0.191 #
29)	L6 AR-1254-4	7.023	5.992	119286	13455	1.640	0.180 #
30)	L6 AR-1254-5	7.477	6.446	65165	27980	0.843	0.272 #
31)	L7 AR-1260-1	6.891	5.877	26692	48856	0.388	0.694 #
32)	L7 AR-1260-2	7.167	6.101	143776	52888	1.852	0.591 #
33)	L7 AR-1260-3	7.578	6.243	157267	249331	2.499	2.927
34)	L7 AR-1260-4	7.824	6.767	50455	9088	0.688	0.135 #
35)	L7 AR-1260-5	8.156	7.032	591988	10335	4.227	0.066 #
36)	L8 AR-1262-1	7.578	6.484	157267	20693	1.659	0.201 #
37)	L8 AR-1262-2	8.156	6.767	591988	9088	3.679	0.098 #
38)	L8 AR-1262-3	8.489	7.340	33719	7948	0.313	0.109 #
39)	L8 AR-1262-4	8.576	7.411	51231	18002	1.002	0.134 #
40)	L8 AR-1262-5	9.262	7.950	450437	17392	8.058	0.264 #
41)	L9 AR-1268-1	8.489	7.340	33719	7948	0.175	0.037 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032522\
 Data File : PP045183.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Mar 2022 13:27
 Operator : YP\AJ
 Sample : PB143647BL
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 26 21:04:34 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 17 04:32:40 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.592	7.411	25519	18002	0.146	0.092 #
43) L9	AR-1268-3	8.836	7.626	79932	85749	0.523	0.521
44) L9	AR-1268-4	9.262	7.950	450437	17392	7.197	0.227 #
45) L9	AR-1268-5	9.692	8.257	43382	51076	0.090	0.099



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032522\
Data File : PP045183.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Mar 2022 13:27
Operator : YP\AJ
Sample : PB143647BL
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 26 21:04:34 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031722.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 17 04:32:40 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.898 min
Delta R.T.: 0.001 min
Response: 40039764
Conc: 21.98 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.153 min
Delta R.T.: -0.002 min
Response: 31978546
Conc: 24.27 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.022 min
Delta R.T.: 0.000 min
Response: 28213074
Conc: 21.92 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.507 min
Delta R.T.: -0.008 min
Response: 29396562
Conc: 22.62 ng/ml



#3 AR-1016-1


R.T.: 5.149 min
Delta R.T.: -0.001 min
Response: 87336
Conc: 1.49 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 4.302 min
Delta R.T.: 0.003 min
Response: 23045
Conc: 0.45 ng/ml




#4 AR-1016-2

R.T.: 5.175 min
Delta R.T.: 0.002 min
Response: 112676
Conc: 1.26 ng/ml



#4 AR-1016-2


R.T.: 4.317 min
Delta R.T.: 0.000 min
Response: 28765
Conc: 0.40 ng/ml



#5 AR-1016-3


R.T.: 5.236 min
Delta R.T.: -0.002 min
Response: 128080
Conc: 2.46 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 4.509 min
Delta R.T.: 0.007 min
Response: 38574
Conc: 0.98 ng/ml




#6 AR-1016-4

R.T.: 5.352 min
Delta R.T.: 0.008 min
Response: 52149
Conc: 1.22 ng/ml



#6 AR-1016-4


R.T.: 4.554 min
Delta R.T.: 0.003 min
Response: 33415
Conc: 1.06 ng/ml



#7 AR-1016-5


R.T.: 5.666 min
Delta R.T.: 0.002 min
Response: 24589
Conc: 0.56 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 4.775 min
Delta R.T.: 0.000 min
Response: 13119
Conc: 0.33 ng/ml




#8 AR-1221-1

R.T.: 4.126 min
Delta R.T.: -0.020 min
Response: 1195665
Conc: 58.26 ng/ml



#8 AR-1221-1


R.T.: 3.395 min
Delta R.T.: 0.012 min
Response: 7923
Conc: 0.53 ng/ml



#9 AR-1221-2


R.T.: 4.248 min
Delta R.T.: 0.010 min
Response: 40282
Conc: 2.63 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.459 min
Delta R.T.: -0.013 min
Response: 2467971
Conc: 199.21 ng/ml




#10 AR-1221-3

R.T.: 4.312 min
Delta R.T.: -0.006 min
Response: 42895
Conc: 0.89 ng/ml



#10 AR-1221-3


R.T.: 3.550 min
Delta R.T.: 0.000 min
Response: 24407
Conc: 0.63 ng/ml



#11 AR-1232-1


R.T.: 4.295 min
Delta R.T.: -0.002 min
Response: 13161
Conc: 0.29 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 3.550 min
Delta R.T.: 0.000 min
Response: 24407
Conc: 0.68 ng/ml




#12 AR-1232-2

R.T.: 4.865 min
Delta R.T.: 0.005 min
Response: 145231
Conc: 6.91 ng/ml



#12 AR-1232-2


R.T.: 4.317 min
Delta R.T.: 0.000 min
Response: 28765
Conc: 0.83 ng/ml



#13 AR-1232-3


R.T.: 5.175 min
Delta R.T.: 0.002 min
Response: 112676
Conc: 2.70 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 4.509 min
Delta R.T.: 0.007 min
Response: 38574
Conc: 2.06 ng/ml




#14 AR-1232-4

R.T.: 5.352 min
Delta R.T.: 0.007 min
Response: 52149
Conc: 2.61 ng/ml



#14 AR-1232-4


R.T.: 4.598 min
Delta R.T.: 0.004 min
Response: 17081
Conc: 1.03 ng/ml



#15 AR-1232-5


R.T.: 5.443 min
Delta R.T.: -0.004 min
Response: 35612
Conc: 2.52 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 4.775 min
Delta R.T.: 0.000 min
Response: 13119
Conc: 0.71 ng/ml




#16 AR-1242-1

R.T.: 5.149 min
Delta R.T.: 0.000 min
Response: 87336
Conc: 1.90 ng/ml



#16 AR-1242-1


R.T.: 4.302 min
Delta R.T.: 0.003 min
Response: 23045
Conc: 0.58 ng/ml



#17 AR-1242-2


R.T.: 5.175 min
Delta R.T.: 0.003 min
Response: 112676
Conc: 1.61 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 4.317 min
Delta R.T.: 0.000 min
Response: 28765
Conc: 0.51 ng/ml




#18 AR-1242-3

R.T.: 5.236 min
Delta R.T.: -0.002 min
Response: 128080
Conc: 3.12 ng/ml



#18 AR-1242-3


R.T.: 4.509 min
Delta R.T.: 0.007 min
Response: 38574
Conc: 1.24 ng/ml



#19 AR-1242-4


R.T.: 5.352 min
Delta R.T.: 0.007 min
Response: 52149
Conc: 1.52 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 4.598 min
Delta R.T.: 0.004 min
Response: 17081
Conc: 0.56 ng/ml




#20 AR-1242-5

R.T.: 6.143 min
Delta R.T.: -0.002 min
Response: 76012
Conc: 2.08 ng/ml



#20 AR-1242-5


R.T.: 5.156 min
Delta R.T.: 0.003 min
Response: 25071
Conc: 0.63 ng/ml



#21 AR-1248-1


R.T.: 5.149 min
Delta R.T.: 0.000 min
Response: 87336
Conc: 2.48 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.302 min
Delta R.T.: 0.003 min
Response: 23045
Conc: 0.73 ng/ml




#22 AR-1248-2

R.T.: 5.443 min
Delta R.T.: -0.003 min
Response: 35612
Conc: 0.75 ng/ml



#22 AR-1248-2


R.T.: 4.554 min
Delta R.T.: 0.003 min
Response: 33415
Conc: 0.76 ng/ml



#23 AR-1248-3


R.T.: 5.666 min
Delta R.T.: 0.002 min
Response: 24589
Conc: 0.42 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 4.598 min
Delta R.T.: 0.004 min
Response: 17081
Conc: 0.37 ng/ml




#24 AR-1248-4

R.T.: 6.098 min
Delta R.T.: -0.006 min
Response: 16935
Conc: 0.27 ng/ml



#24 AR-1248-4


R.T.: 4.775 min
Delta R.T.: 0.000 min
Response: 13119
Conc: 0.24 ng/ml



#25 AR-1248-5


R.T.: 6.143 min
Delta R.T.: -0.002 min
Response: 76012
Conc: 1.22 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.192 min
Delta R.T.: -0.002 min
Response: 21553
Conc: 0.38 ng/ml




#26 AR-1254-1

R.T.: 6.079 min
Delta R.T.: 0.005 min
Response: 19335
Conc: 0.29 ng/ml



#26 AR-1254-1


R.T.: 5.148 min
Delta R.T.: -0.003 min
Response: 105193
Conc: 1.22 ng/ml



#27 AR-1254-2


R.T.: 6.319 min
Delta R.T.: 0.005 min
Response: 188620
Conc: 1.89 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 5.316 min
Delta R.T.: 0.004 min
Response: 52935
Conc: 0.71 ng/ml




#28 AR-1254-3

R.T.: 6.716 min
Delta R.T.: 0.002 min
Response: 517916
Conc: 5.06 ng/ml



#28 AR-1254-3


R.T.: 5.746 min
Delta R.T.: 0.002 min
Response: 21901
Conc: 0.19 ng/ml



#29 AR-1254-4


R.T.: 7.023 min
Delta R.T.: -0.005 min
Response: 119286
Conc: 1.64 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 5.992 min
Delta R.T.: -0.001 min
Response: 13455
Conc: 0.18 ng/ml




#30 AR-1254-5

R.T.: 7.477 min
Delta R.T.: -0.011 min
Response: 65165
Conc: 0.84 ng/ml



#30 AR-1254-5


R.T.: 6.446 min
Delta R.T.: 0.002 min
Response: 27980
Conc: 0.27 ng/ml



#31 AR-1260-1


R.T.: 6.891 min
Delta R.T.: -0.003 min
Response: 26692
Conc: 0.39 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 5.877 min
Delta R.T.: -0.006 min
Response: 48856
Conc: 0.69 ng/ml




#32 AR-1260-2

R.T.: 7.167 min
Delta R.T.: -0.009 min
Response: 143776
Conc: 1.85 ng/ml



#32 AR-1260-2


R.T.: 6.101 min
Delta R.T.: 0.011 min
Response: 52888
Conc: 0.59 ng/ml



#33 AR-1260-3


R.T.: 7.578 min
Delta R.T.: 0.007 min
Response: 157267
Conc: 2.50 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.243 min
Delta R.T.: -0.010 min
Response: 249331
Conc: 2.93 ng/ml




#34 AR-1260-4

R.T.: 7.824 min
Delta R.T.: 0.006 min
Response: 50455
Conc: 0.69 ng/ml



#34 AR-1260-4


R.T.: 6.767 min
Delta R.T.: 0.002 min
Response: 9088
Conc: 0.14 ng/ml



#35 AR-1260-5


R.T.: 8.156 min
Delta R.T.: -0.005 min
Response: 591988
Conc: 4.23 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.032 min
Delta R.T.: -0.001 min
Response: 10335
Conc: 0.07 ng/ml




#36 AR-1262-1

R.T.: 7.578 min
Delta R.T.: 0.008 min
Response: 157267
Conc: 1.66 ng/ml



#36 AR-1262-1


R.T.: 6.484 min
Delta R.T.: 0.000 min
Response: 20693
Conc: 0.20 ng/ml



#37 AR-1262-2


R.T.: 8.156 min
Delta R.T.: -0.004 min
Response: 591988
Conc: 3.68 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 6.767 min
Delta R.T.: 0.003 min
Response: 9088
Conc: 0.10 ng/ml




#38 AR-1262-3

R.T.: 8.489 min
Delta R.T.: -0.001 min
Response: 33719
Conc: 0.31 ng/ml



#38 AR-1262-3


R.T.: 7.340 min
Delta R.T.: 0.003 min
Response: 7948
Conc: 0.11 ng/ml



#39 AR-1262-4


R.T.: 8.576 min
Delta R.T.: -0.007 min
Response: 51231
Conc: 1.00 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 7.411 min
Delta R.T.: 0.003 min
Response: 18002
Conc: 0.13 ng/ml




#40 AR-1262-5

R.T.: 9.262 min
Delta R.T.: 0.000 min
Response: 450437
Conc: 8.06 ng/ml



#40 AR-1262-5


R.T.: 7.950 min
Delta R.T.: 0.000 min
Response: 17392
Conc: 0.26 ng/ml



#41 AR-1268-1


R.T.: 8.489 min
Delta R.T.: -0.002 min
Response: 33719
Conc: 0.18 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.340 min
Delta R.T.: 0.000 min
Response: 7948
Conc: 0.04 ng/ml




#42 AR-1268-2

R.T.: 8.592 min
Delta R.T.: 0.004 min
Response: 25519
Conc: 0.15 ng/ml



#42 AR-1268-2


R.T.: 7.411 min
Delta R.T.: 0.000 min
Response: 18002
Conc: 0.09 ng/ml



#43 AR-1268-3


R.T.: 8.836 min
Delta R.T.: 0.012 min
Response: 79932
Conc: 0.52 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.626 min
Delta R.T.: -0.005 min
Response: 85749
Conc: 0.52 ng/ml




#44 AR-1268-4

R.T.: 9.262 min
Delta R.T.: -0.002 min
Response: 450437
Conc: 7.20 ng/ml



#44 AR-1268-4


R.T.: 7.950 min
Delta R.T.: -0.002 min
Response: 17392
Conc: 0.23 ng/ml



#45 AR-1268-5

R.T.: 9.692 min
Delta R.T.: 0.006 min
Response: 43382
Conc: 0.09 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.257 min
Delta R.T.: 0.000 min
Response: 51076
Conc: 0.10 ng/ml