

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032522\
 Data File : PP045184.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Mar 2022 13:42
 Operator : YP\AJ
 Sample : PB143647BS
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 26 21:04:56 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 17 04:32:40 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.893	3.152	45589052	32886155	25.021	24.955
2)	SA Decachlor...	10.015	8.505	29930598	32493235	23.251	25.007
Target Compounds							
3)	L1 AR-1016-1	5.144	4.293	32601558	28266931	555.585	558.025
4)	L1 AR-1016-2	5.166	4.311	43881896	41189374	490.872	571.374
5)	L1 AR-1016-3	5.233	4.495	26716731	24309702	513.357	617.860
6)	L1 AR-1016-4	5.339	4.545	22376937	17311560	522.949	549.861
7)	L1 AR-1016-5	5.658	4.768	18137941	22675161	413.305	568.083 #
8)	L2 AR-1221-1	4.148	3.377	29523	2889608	1.438	194.675 #
9)	L2 AR-1221-2	4.248	3.467	95099	4889613	6.221	394.678 #
10)	L2 AR-1221-3	4.290f	3.546	18071324	15751497	374.262	405.798
11)	L3 AR-1232-1	4.290	3.546	18071324	15751497	401.611	438.253
12)	L3 AR-1232-2	4.854	4.311	24819434	41189374	1181.136	1191.050
13)	L3 AR-1232-3	5.166	4.495	43881896	24309702	1051.862	1297.941
14)	L3 AR-1232-4	5.339	4.587	22376937	22286049	1118.893	1346.229
15)	L3 AR-1232-5	5.440	4.768	17862703	22675161	1264.099	1226.339
16)	L4 AR-1242-1	5.144	4.293	32601558	28266931	710.487	707.609
17)	L4 AR-1242-2	5.166	4.311	43881896	41189374	626.625	725.679
18)	L4 AR-1242-3	5.233	4.495	26716731	24309702	650.492	779.576
19)	L4 AR-1242-4	5.339	4.587	22376937	22286049	650.566	735.516
20)	L4 AR-1242-5	6.140	5.144	2737602	17643803	74.863	443.532 #
21)	L5 AR-1248-1	5.144	4.293	32601558	28266931	926.939	895.064
22)	L5 AR-1248-2	5.440	4.545	17862703	17311560	377.610	395.845
23)	L5 AR-1248-3	5.658	4.587	18137941	22286049	306.974	487.720 #
24)	L5 AR-1248-4	6.098	4.768	5013158	22675161	78.997	419.741 #
25)	L5 AR-1248-5	6.140	5.187	2737602	2711737	43.875	48.301
26)	L6 AR-1254-1	6.068	5.144	15022563	17643803	226.245	204.031
27)	L6 AR-1254-2	6.307	5.304	16418756	17544125	164.333	235.765 #
28)	L6 AR-1254-3	6.707	5.752	8819449	32047018	86.229	278.913 #
29)	L6 AR-1254-4	7.020	5.984	5596670	3498525	76.946	46.925 #
30)	L6 AR-1254-5	7.481	6.435	43667488	58281582	564.593	565.996
31)	L7 AR-1260-1	6.887	5.875	34507640	41482420	501.231	589.148
32)	L7 AR-1260-2	7.170	6.082	40992650	50039092	527.988	559.540
33)	L7 AR-1260-3	7.566	6.244	24078046	47912834	382.639	562.560 #
34)	L7 AR-1260-4	7.813	6.756	30499054	33524942	415.769	498.831
35)	L7 AR-1260-5	8.155	7.024	55261877	77487248	394.634	495.897 #
36)	L8 AR-1262-1	7.566	6.476	24078046	36660368	253.959	356.087 #
37)	L8 AR-1262-2	8.155	6.756	55261877	33524942	343.398	363.208
38)	L8 AR-1262-3	8.499	7.330	25210989	15565103	233.720	213.880
39)	L8 AR-1262-4	0.000	7.399	0	55834056	N.D.	415.408 #
40)	L8 AR-1262-5	9.257	7.943	14794907	18038210	264.674	273.985
41)	L9 AR-1268-1	8.499	7.330	25210989	15565103	130.966	72.130 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032522\
 Data File : PP045184.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Mar 2022 13:42
 Operator : YP\AJ
 Sample : PB143647BS
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 26 21:04:56 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 17 04:32:40 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	0.000	7.399	0	55834056	N.D.	285.519 #
43) L9	AR-1268-3	8.815	7.621	1003220	1073692	6.569	6.518
44) L9	AR-1268-4	9.257	7.943	14794907	18038210	236.379	235.520
45) L9	AR-1268-5	9.675	8.246	5795469	6034091	12.027	11.709



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032522\
Data File : PP045184.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Mar 2022 13:42
Operator : YP\AJ
Sample : PB143647BS
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 26 21:04:56 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031722.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 17 04:32:40 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.893 min
Delta R.T.: -0.004 min
Response: 45589052
Conc: 25.02 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.152 min
Delta R.T.: -0.003 min
Response: 32886155
Conc: 24.95 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.015 min
Delta R.T.: -0.008 min
Response: 29930598
Conc: 23.25 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.505 min
Delta R.T.: -0.011 min
Response: 32493235
Conc: 25.01 ng/ml



#3 AR-1016-1


R.T.: 5.144 min
Delta R.T.: -0.007 min
Response: 32601558
Conc: 555.59 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 4.293 min
Delta R.T.: -0.006 min
Response: 28266931
Conc: 558.03 ng/ml




#4 AR-1016-2

R.T.: 5.166 min
Delta R.T.: -0.007 min
Response: 43881896
Conc: 490.87 ng/ml



#4 AR-1016-2


R.T.: 4.311 min
Delta R.T.: -0.007 min
Response: 41189374
Conc: 571.37 ng/ml



#5 AR-1016-3


R.T.: 5.233 min
Delta R.T.: -0.006 min
Response: 26716731
Conc: 513.36 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 4.495 min
Delta R.T.: -0.007 min
Response: 24309702
Conc: 617.86 ng/ml




#6 AR-1016-4

R.T.: 5.339 min
Delta R.T.: -0.006 min
Response: 22376937
Conc: 522.95 ng/ml



#6 AR-1016-4


R.T.: 4.545 min
Delta R.T.: -0.007 min
Response: 17311560
Conc: 549.86 ng/ml



#7 AR-1016-5


R.T.: 5.658 min
Delta R.T.: -0.006 min
Response: 18137941
Conc: 413.31 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 4.768 min
Delta R.T.: -0.007 min
Response: 22675161
Conc: 568.08 ng/ml




#8 AR-1221-1

R.T.: 4.148 min
Delta R.T.: 0.002 min
Response: 29523
Conc: 1.44 ng/ml



#8 AR-1221-1


R.T.: 3.377 min
Delta R.T.: -0.005 min
Response: 2889608
Conc: 194.68 ng/ml



#9 AR-1221-2


R.T.: 4.248 min
Delta R.T.: 0.010 min
Response: 95099
Conc: 6.22 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.467 min
Delta R.T.: -0.004 min
Response: 4889613
Conc: 394.68 ng/ml




#10 AR-1221-3

R.T.: 4.290 min
Delta R.T.: -0.028 min
Response: 18071324
Conc: 374.26 ng/ml



#10 AR-1221-3


R.T.: 3.546 min
Delta R.T.: -0.005 min
Response: 15751497
Conc: 405.80 ng/ml



#11 AR-1232-1


R.T.: 4.290 min
Delta R.T.: -0.006 min
Response: 18071324
Conc: 401.61 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 3.546 min
Delta R.T.: -0.004 min
Response: 15751497
Conc: 438.25 ng/ml




#12 AR-1232-2

R.T.: 4.854 min
Delta R.T.: -0.006 min
Response: 24819434
Conc: 1181.14 ng/ml



#12 AR-1232-2


R.T.: 4.311 min
Delta R.T.: -0.006 min
Response: 41189374
Conc: 1191.05 ng/ml



#13 AR-1232-3


R.T.: 5.166 min
Delta R.T.: -0.007 min
Response: 43881896
Conc: 1051.86 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 4.495 min
Delta R.T.: -0.007 min
Response: 24309702
Conc: 1297.94 ng/ml




#14 AR-1232-4

R.T.: 5.339 min
Delta R.T.: -0.006 min
Response: 22376937
Conc: 1118.89 ng/ml



#14 AR-1232-4


R.T.: 4.587 min
Delta R.T.: -0.007 min
Response: 22286049
Conc: 1346.23 ng/ml



#15 AR-1232-5


R.T.: 5.440 min
Delta R.T.: -0.007 min
Response: 17862703
Conc: 1264.10 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 4.768 min
Delta R.T.: -0.007 min
Response: 22675161
Conc: 1226.34 ng/ml




#16 AR-1242-1

R.T.: 5.144 min
Delta R.T.: -0.006 min
Response: 32601558
Conc: 710.49 ng/ml



#16 AR-1242-1


R.T.: 4.293 min
Delta R.T.: -0.006 min
Response: 28266931
Conc: 707.61 ng/ml



#17 AR-1242-2


R.T.: 5.166 min
Delta R.T.: -0.006 min
Response: 43881896
Conc: 626.62 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 4.311 min
Delta R.T.: -0.006 min
Response: 41189374
Conc: 725.68 ng/ml




#18 AR-1242-3

R.T.: 5.233 min
Delta R.T.: -0.006 min
Response: 26716731
Conc: 650.49 ng/ml



#18 AR-1242-3


R.T.: 4.495 min
Delta R.T.: -0.007 min
Response: 24309702
Conc: 779.58 ng/ml



#19 AR-1242-4


R.T.: 5.339 min
Delta R.T.: -0.006 min
Response: 22376937
Conc: 650.57 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 4.587 min
Delta R.T.: -0.006 min
Response: 22286049
Conc: 735.52 ng/ml




#20 AR-1242-5

R.T.: 6.140 min
Delta R.T.: -0.005 min
Response: 2737602
Conc: 74.86 ng/ml



#20 AR-1242-5


R.T.: 5.144 min
Delta R.T.: -0.009 min
Response: 17643803
Conc: 443.53 ng/ml



#21 AR-1248-1


R.T.: 5.144 min
Delta R.T.: -0.006 min
Response: 32601558
Conc: 926.94 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.293 min
Delta R.T.: -0.006 min
Response: 28266931
Conc: 895.06 ng/ml




#22 AR-1248-2

R.T.: 5.440 min
Delta R.T.: -0.006 min
Response: 17862703
Conc: 377.61 ng/ml



#22 AR-1248-2


R.T.: 4.545 min
Delta R.T.: -0.006 min
Response: 17311560
Conc: 395.84 ng/ml



#23 AR-1248-3


R.T.: 5.658 min
Delta R.T.: -0.006 min
Response: 18137941
Conc: 306.97 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 4.587 min
Delta R.T.: -0.006 min
Response: 22286049
Conc: 487.72 ng/ml




#24 AR-1248-4

R.T.: 6.098 min
Delta R.T.: -0.006 min
Response: 5013158
Conc: 79.00 ng/ml



#24 AR-1248-4


R.T.: 4.768 min
Delta R.T.: -0.007 min
Response: 22675161
Conc: 419.74 ng/ml



#25 AR-1248-5


R.T.: 6.140 min
Delta R.T.: -0.005 min
Response: 2737602
Conc: 43.88 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.187 min
Delta R.T.: -0.008 min
Response: 2711737
Conc: 48.30 ng/ml




#26 AR-1254-1

R.T.: 6.068 min
Delta R.T.: -0.007 min
Response: 15022563
Conc: 226.25 ng/ml



#26 AR-1254-1


R.T.: 5.144 min
Delta R.T.: -0.008 min
Response: 17643803
Conc: 204.03 ng/ml



#27 AR-1254-2


R.T.: 6.307 min
Delta R.T.: -0.007 min
Response: 16418756
Conc: 164.33 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 5.304 min
Delta R.T.: -0.008 min
Response: 17544125
Conc: 235.76 ng/ml




#28 AR-1254-3

R.T.: 6.707 min
Delta R.T.: -0.006 min
Response: 8819449
Conc: 86.23 ng/ml



#28 AR-1254-3


R.T.: 5.752 min
Delta R.T.: 0.008 min
Response: 32047018
Conc: 278.91 ng/ml



#29 AR-1254-4


R.T.: 7.020 min
Delta R.T.: -0.007 min
Response: 5596670
Conc: 76.95 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 5.984 min
Delta R.T.: -0.009 min
Response: 3498525
Conc: 46.93 ng/ml




#30 AR-1254-5

R.T.: 7.481 min
Delta R.T.: -0.008 min
Response: 43667488
Conc: 564.59 ng/ml



#30 AR-1254-5


R.T.: 6.435 min
Delta R.T.: -0.009 min
Response: 58281582
Conc: 566.00 ng/ml



#31 AR-1260-1


R.T.: 6.887 min
Delta R.T.: -0.006 min
Response: 34507640
Conc: 501.23 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 5.875 min
Delta R.T.: -0.009 min
Response: 41482420
Conc: 589.15 ng/ml




#32 AR-1260-2

R.T.: 7.170 min
Delta R.T.: -0.006 min
Response: 40992650
Conc: 527.99 ng/ml



#32 AR-1260-2


R.T.: 6.082 min
Delta R.T.: -0.009 min
Response: 50039092
Conc: 559.54 ng/ml



#33 AR-1260-3


R.T.: 7.566 min
Delta R.T.: -0.005 min
Response: 24078046
Conc: 382.64 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.244 min
Delta R.T.: -0.009 min
Response: 47912834
Conc: 562.56 ng/ml




#34 AR-1260-4

R.T.: 7.813 min
Delta R.T.: -0.006 min
Response: 30499054
Conc: 415.77 ng/ml



#34 AR-1260-4


R.T.: 6.756 min
Delta R.T.: -0.009 min
Response: 33524942
Conc: 498.83 ng/ml



#35 AR-1260-5


R.T.: 8.155 min
Delta R.T.: -0.006 min
Response: 55261877
Conc: 394.63 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.024 min
Delta R.T.: -0.009 min
Response: 77487248
Conc: 495.90 ng/ml




#36 AR-1262-1

R.T.: 7.566 min
Delta R.T.: -0.004 min
Response: 24078046
Conc: 253.96 ng/ml



#36 AR-1262-1


R.T.: 6.476 min
Delta R.T.: -0.008 min
Response: 36660368
Conc: 356.09 ng/ml



#37 AR-1262-2


R.T.: 8.155 min
Delta R.T.: -0.005 min
Response: 55261877
Conc: 343.40 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 6.756 min
Delta R.T.: -0.007 min
Response: 33524942
Conc: 363.21 ng/ml




#38 AR-1262-3

R.T.: 8.499 min
Delta R.T.: 0.009 min
Response: 25210989
Conc: 233.72 ng/ml



#38 AR-1262-3


R.T.: 7.330 min
Delta R.T.: -0.008 min
Response: 15565103
Conc: 213.88 ng/ml



#39 AR-1262-4


R.T.: 0.000 min
Exp R.T.: 8.583 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 7.399 min
Delta R.T.: -0.009 min
Response: 55834056
Conc: 415.41 ng/ml




#40 AR-1262-5

R.T.: 9.257 min
Delta R.T.: -0.006 min
Response: 14794907
Conc: 264.67 ng/ml



#40 AR-1262-5


R.T.: 7.943 min
Delta R.T.: -0.007 min
Response: 18038210
Conc: 273.99 ng/ml



#41 AR-1268-1


R.T.: 8.499 min
Delta R.T.: 0.008 min
Response: 25210989
Conc: 130.97 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.330 min
Delta R.T.: -0.010 min
Response: 15565103
Conc: 72.13 ng/ml




#42 AR-1268-2

R.T.: 0.000 min
Exp R.T. : 8.589 min
Response: 0
Conc: N.D.



#42 AR-1268-2


R.T.: 7.399 min
Delta R.T.: -0.012 min
Response: 55834056
Conc: 285.52 ng/ml



#43 AR-1268-3


R.T.: 8.815 min
Delta R.T.: -0.009 min
Response: 1003220
Conc: 6.57 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.621 min
Delta R.T.: -0.010 min
Response: 1073692
Conc: 6.52 ng/ml




#44 AR-1268-4

R.T.: 9.257 min
Delta R.T.: -0.008 min
Response: 14794907
Conc: 236.38 ng/ml



#44 AR-1268-4


R.T.: 7.943 min
Delta R.T.: -0.009 min
Response: 18038210
Conc: 235.52 ng/ml



#45 AR-1268-5

R.T.: 9.675 min
Delta R.T.: -0.010 min
Response: 5795469
Conc: 12.03 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.246 min
Delta R.T.: -0.010 min
Response: 6034091
Conc: 11.71 ng/ml