

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032522\
 Data File : PP045260.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Mar 2022 19:29
 Operator : YP\AJ
 Sample : N2133-07 10X
 Misc :
 ALS Vial : 68 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 1932

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 03/28/2022
 Supervised By :mohammad ahmed 03/29/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 28 01:52:07 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 17 04:32:40 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.889	3.150	3082059	2443885	1.692m	1.854m
2) SA Decachlor...	10.009	8.502	1821761	1889000	1.415m	1.454

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032522\
Data File : PP045260.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2022 19:29
Operator : YP\AJ
Sample : N2133-07 10X
Misc :
ALS Vial : 68 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
1932


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/28/2022
Supervised By :mohammad ahmed 03/29/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 28 01:52:07 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031722.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 17 04:32:40 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.889 min
Delta R.T.: -0.008 min
Response: 3082059
Conc: 1.69 ng/ml

Instrument :
ECD_P
ClientSampleId :
1932


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/28/2022
Supervised By :mohammad ahmed 03/29/2022




#1 Tetrachloro-m-xylene

R.T.: 3.150 min
Delta R.T.: -0.005 min
Response: 2443885
Conc: 1.85 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.009 min
Delta R.T.: -0.014 min
Response: 1821761
Conc: 1.42 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.502 min
Delta R.T.: -0.013 min
Response: 1889000
Conc: 1.45 ng/ml