

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032621\
 Data File : PP034441.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 26 Mar 2021 23:16
 Operator : DD\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 27 00:22:26 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031821.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Mar 20 01:42:16 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.847	4.259	3401608	1951844	40.492	44.107
2)	SA Decachlor...	10.771	9.571	3600472	1334054	45.879	50.553
Target Compounds							
21)	L5 AR-1248-1	6.160	5.513	634136	317658	424.134	461.784
22)	L5 AR-1248-2	6.454	5.775	869114	454288	414.172	454.929
23)	L5 AR-1248-3	6.669	5.820	1034677	485139	406.287	460.035
24)	L5 AR-1248-4	7.096	6.006	1056327	541929	408.841	441.605
25)	L5 AR-1248-5	7.135	6.426	1048181	511497	403.912	506.607 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032621\
Data File : PP034441.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 26 Mar 2021 23:16
Operator : DD\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 00:22:26 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031821.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Mar 20 01:42:16 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.847 min
Delta R.T.: -0.005 min
Response: 3401608
Conc: 40.49 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.259 min
Delta R.T.: -0.004 min
Response: 1951844
Conc: 44.11 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.771 min
Delta R.T.: -0.014 min
Response: 3600472
Conc: 45.88 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.571 min
Delta R.T.: -0.014 min
Response: 1334054
Conc: 50.55 ng/ml



#21 AR-1248-1


R.T.: 6.160 min
Delta R.T.: -0.006 min
Response: 634136
Conc: 424.13 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 5.513 min
Delta R.T.: -0.006 min
Response: 317658
Conc: 461.78 ng/ml




#22 AR-1248-2

R.T.: 6.454 min
Delta R.T.: -0.007 min
Response: 869114
Conc: 414.17 ng/ml



#22 AR-1248-2


R.T.: 5.775 min
Delta R.T.: -0.006 min
Response: 454288
Conc: 454.93 ng/ml



#23 AR-1248-3


R.T.: 6.669 min
Delta R.T.: -0.006 min
Response: 1034677
Conc: 406.29 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.820 min
Delta R.T.: -0.007 min
Response: 485139
Conc: 460.04 ng/ml




#24 AR-1248-4

R.T.: 7.096 min
Delta R.T.: -0.007 min
Response: 1056327
Conc: 408.84 ng/ml



#24 AR-1248-4


R.T.: 6.006 min
Delta R.T.: -0.006 min
Response: 541929
Conc: 441.60 ng/ml



#25 AR-1248-5

R.T.: 7.135 min
Delta R.T.: -0.007 min
Response: 1048181
Conc: 403.91 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 6.426 min
Delta R.T.: -0.007 min
Response: 511497
Conc: 506.61 ng/ml