

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032624\
Data File : PP063801.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Mar 2024 11:45
Operator : YP\AJ
Sample : PB159770BL
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PB159770BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 26 11:59:44 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031924.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 19 06:23:02 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.487	3.638	36273962	39931384	15.202	16.725
2) SA Decachlor...	10.356	8.672	42538951	38231478	16.754	14.147
Target Compounds						
9) L2 AR-1221-2	0.000	3.939	0	502313	N.D.	24.510 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032624\
Data File : PP063801.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Mar 2024 11:45
Operator : YP\AJ
Sample : PB159770BL
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PB159770BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 26 11:59:44 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031924.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 19 06:23:02 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.487 min
Delta R.T.: -0.002 min
Response: 36273962
Conc: 15.20 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB159770BL




#1 Tetrachloro-m-xylene

R.T.: 3.638 min
Delta R.T.: 0.000 min
Response: 39931384
Conc: 16.72 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.356 min
Delta R.T.: -0.010 min
Response: 42538951
Conc: 16.75 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.672 min
Delta R.T.: -0.012 min
Response: 38231478
Conc: 14.15 ng/ml



#9 AR-1221-2

R.T.: 0.000 min
Exp R.T.: 4.789 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :
PB159770BL



#9 AR-1221-2

R.T.: 3.939 min
Delta R.T.: -0.002 min
Response: 502313
Conc: 24.51 ng/ml