

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032624\
 Data File : PP063810.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Mar 2024 14:31
 Operator : YP\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 26 14:51:54 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031924.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Mar 19 06:23:02 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.486	3.638	42309506	43360932	17.731	18.161
2)	SA Decachlor...	10.351	8.671	49896434	47258960	19.652	17.487
Target Compounds							
9)	L2 AR-1221-2	0.000	3.939	0	615786	N.D.	30.047 #
27)	L6 AR-1254-2	6.771	0.000	2550887	0	18.607	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032624\
Data File : PP063810.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Mar 2024 14:31
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 26 14:51:54 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031924.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 19 06:23:02 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.486 min
Delta R.T.: -0.003 min
Response: 42309506
Conc: 17.73 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.638 min
Delta R.T.: -0.002 min
Response: 43360932
Conc: 18.16 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.351 min
Delta R.T.: -0.014 min
Response: 49896434
Conc: 19.65 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.671 min
Delta R.T.: -0.013 min
Response: 47258960
Conc: 17.49 ng/ml



#9 AR-1221-2


R.T.: 0.000 min
Exp R.T.: 4.789 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :
I.BLK




#9 AR-1221-2

R.T.: 3.939 min
Delta R.T.: -0.003 min
Response: 615786
Conc: 30.05 ng/ml



#27 AR-1254-2

R.T.: 6.771 min
Delta R.T.: 0.014 min
Response: 2550887
Conc: 18.61 ng/ml



#27 AR-1254-2

R.T.: 0.000 min
Exp R.T.: 5.687 min
Response: 0
Conc: N.D.