

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_P\Data\PP032819\  
 Data File : PP029617.D  
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH  
 Acq On : 28 Mar 2019 21:49  
 Operator : EI\MA  
 Sample : K2107-06  
 Misc : hexane  
 ALS Vial : 19 Sample Multiplier: 1

Instrument :  
 ECD\_P  
 ClientSampleId :

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Mar 29 01:07:34 2019  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_P\methods\PP032119.M  
 Quant Title : 8080.M  
 QLast Update : Fri Mar 22 09:23:06 2019  
 Response via : Initial Calibration  
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl  
 Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2  
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

|                             | Compound     | RT#1   | RT#2   | Resp#1   | Resp#2   | ng/ml   | ng/ml   |
|-----------------------------|--------------|--------|--------|----------|----------|---------|---------|
| -----                       |              |        |        |          |          |         |         |
| System Monitoring Compounds |              |        |        |          |          |         |         |
| 4) S                        | 2,4-DCAA     | 11.924 | 12.118 | 21264299 | 23521558 | 649.986 | 547.760 |
| Target Compounds            |              |        |        |          |          |         |         |
| 3) T                        | 4-Nitroph... | 11.421 | 0.000  | 147950   | 0        | 2.258   | N.D. #  |
| 5) T                        | DICAMBA      | 12.223 | 0.000  | 541576   | 0        | 3.665   | N.D. #  |
| 8) T                        | DICHLORPROP  | 13.374 | 13.552 | -239314  | 1195634  | N.D.    | 29.524  |
| 11) T                       | 2,4,5-TP ... | 0.000  | 15.144 | 0        | 9446     | N.D.    | 0.046 # |
| 12) T                       | 2,4,5-T      | 0.000  | 15.852 | 0        | 1409521  | N.D.    | 8.503 # |
| 13) T                       | 2,4-DB       | 16.211 | 0.000  | 884712   | 0        | 42.547  | N.D. #  |
| 14) T                       | DINOSEB      | 17.696 | 0.000  | 125547   | 0        | 0.745   | N.D. #  |
| 15) T                       | Picloram     | 17.463 | 0.000  | 220754   | 0        | 0.674   | N.D. #  |
| 16) T                       | DCPA         | 18.046 | 0.000  | 754967   | 0        | 2.380   | N.D. #  |
| -----                       |              |        |        |          |          |         |         |


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_P\Data\PP032819\  
Data File : PP029617.D  
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH  
Acq On : 28 Mar 2019 21:49  
Operator : EI\MA  
Sample : K2107-06  
Misc : hexane  
ALS Vial : 19 Sample Multiplier: 1

Instrument :  
ECD\_P  
ClientSampleId :

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Mar 29 01:07:34 2019  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_P\methods\PP032119.M  
Quant Title : 8080.M  
QLast Update : Fri Mar 22 09:23:06 2019  
Response via : Initial Calibration  
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl  
Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2  
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#3 4-Nitrophenol


R.T.: 11.421 min  
Delta R.T.: -0.070 min  
Response: 147950  
Conc: 2.26 ng/ml

Instrument :  
ECD\_P  
ClientSampleId :




#3 4-Nitrophenol

R.T.: 0.000 min  
Exp R.T. : 11.266 min  
Response: 0  
Conc: N.D.




#4 2,4-DCAA

R.T.: 11.924 min  
Delta R.T.: 0.000 min  
Response: 21264299  
Conc: 649.99 ng/ml



#4 2,4-DCAA


R.T.: 12.118 min  
Delta R.T.: -0.003 min  
Response: 23521558  
Conc: 547.76 ng/ml



#5 DICAMBA


R.T.: 12.223 min  
Delta R.T.: -0.024 min  
Response: 541576  
Conc: 3.67 ng/ml

Instrument :  
ECD\_P  
ClientSampleId :




#5 DICAMBA

R.T.: 0.000 min  
Exp R.T.: 12.433 min  
Response: 0  
Conc: N.D.




#8 DICHLORPROP

R.T.: 13.374 min  
Delta R.T.: -0.064 min  
Response: -239314  
Conc: N.D.



#8 DICHLORPROP


R.T.: 13.552 min  
Delta R.T.: 0.007 min  
Response: 1195634  
Conc: 29.52 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 0.000 min  
Exp R.T.: 15.062 min  
Response: 0  
Conc: N.D.

Instrument :  
ECD\_P  
ClientSampleId :




#11 2,4,5-TP (SILVEX)

R.T.: 15.144 min  
Delta R.T.: -0.066 min  
Response: 9446  
Conc: 0.05 ng/ml




#12 2,4,5-T

R.T.: 0.000 min  
Exp R.T.: 15.451 min  
Response: 0  
Conc: N.D.



#12 2,4,5-T


R.T.: 15.852 min  
Delta R.T.: -0.006 min  
Response: 1409521  
Conc: 8.50 ng/ml



#13 2,4-DB


R.T.: 16.211 min  
Delta R.T.: 0.004 min  
Response: 884712  
Conc: 42.55 ng/ml

Instrument :  
ECD\_P  
ClientSampleId :




#13 2,4-DB

R.T.: 0.000 min  
Exp R.T. : 16.455 min  
Response: 0  
Conc: N.D.




#14 DINOSEB

R.T.: 17.696 min  
Delta R.T.: 0.024 min  
Response: 125547  
Conc: 0.74 ng/ml



#14 DINOSEB


R.T.: 0.000 min  
Exp R.T. : 16.845 min  
Response: 0  
Conc: N.D.



#15 Picloram


R.T.: 17.463 min  
Delta R.T.: 0.004 min  
Response: 220754  
Conc: 0.67 ng/ml

Instrument :  
ECD\_P  
ClientSampleId :




#15 Picloram

R.T.: 0.000 min  
Exp R.T. : 18.242 min  
Response: 0  
Conc: N.D.



#16 DCPA

R.T.: 18.046 min  
Delta R.T.: -0.009 min  
Response: 754967  
Conc: 2.38 ng/ml



#16 DCPA

R.T.: 0.000 min  
Exp R.T. : 18.116 min  
Response: 0  
Conc: N.D.