

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032825\
 Data File : PP070952.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Mar 2025 08:45
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 28 10:57:29 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Mar 28 03:13:46 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.512	3.816	81135595	56195997	55.512	54.210
2) SA Decachlor...	10.228	8.857	60307408	42968788	57.293	54.515
Target Compounds						
16) L4 AR-1242-1	5.665	4.903	22558462	18225314	512.959	532.969
17) L4 AR-1242-2	5.687	4.921	34430090	25860153	535.324	524.997
18) L4 AR-1242-3	5.749	5.099	21076129	14712338	513.558	552.176
19) L4 AR-1242-4	5.846	5.182	17261385	14565620	565.514	552.183
20) L4 AR-1242-5	6.577	5.707	19738595	17101250	540.629	552.396



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP032825\
Data File : PP070952.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Mar 2025 08:45
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 28 10:57:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP032725.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 28 03:13:46 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.512 min
Delta R.T.: 0.000 min
Response: 81135595
Conc: 55.51 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.816 min
Delta R.T.: 0.000 min
Response: 56195997
Conc: 54.21 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.228 min
Delta R.T.: 0.002 min
Response: 60307408
Conc: 57.29 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.857 min
Delta R.T.: -0.002 min
Response: 42968788
Conc: 54.52 ng/ml



#16 AR-1242-1


R.T.: 5.665 min
Delta R.T.: 0.000 min
Response: 22558462
Conc: 512.96 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.903 min
Delta R.T.: 0.000 min
Response: 18225314
Conc: 532.97 ng/ml




#17 AR-1242-2

R.T.: 5.687 min
Delta R.T.: 0.001 min
Response: 34430090
Conc: 535.32 ng/ml



#17 AR-1242-2


R.T.: 4.921 min
Delta R.T.: 0.000 min
Response: 25860153
Conc: 525.00 ng/ml



#18 AR-1242-3


R.T.: 5.749 min
Delta R.T.: 0.001 min
Response: 21076129
Conc: 513.56 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 5.099 min
Delta R.T.: 0.000 min
Response: 14712338
Conc: 552.18 ng/ml




#19 AR-1242-4

R.T.: 5.846 min
Delta R.T.: 0.001 min
Response: 17261385
Conc: 565.51 ng/ml



#19 AR-1242-4


R.T.: 5.182 min
Delta R.T.: 0.000 min
Response: 14565620
Conc: 552.18 ng/ml



#20 AR-1242-5

R.T.: 6.577 min
Delta R.T.: 0.001 min
Response: 19738595
Conc: 540.63 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.707 min
Delta R.T.: 0.000 min
Response: 17101250
Conc: 552.40 ng/ml