

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP033021\
 Data File : PP034477.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 30 Mar 2021 22:02
 Operator : DD\AJ
 Sample : AR1268ICC500
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 31 05:49:49 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP033021.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 31 05:48:47 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.846	4.259	4021530	2306820	50.000	50.000
2) SA Decachlor...	10.765	9.566	6451606	2288451	50.000	50.000
Target Compounds						
41) L9 AR-1268-1	9.316	8.450	4977154	2115638	500.000	500.000
42) L9 AR-1268-2	9.409	8.515	4798548	2048267	500.000	500.000
43) L9 AR-1268-3	9.626	8.721	4348136	1771516	500.000	500.000
44) L9 AR-1268-4	10.050	9.010	1651814	630223	500.000	500.000
45) L9 AR-1268-5	10.446	9.307	14451868	5123122	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP033021\
Data File : PP034477.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 30 Mar 2021 22:02
Operator : DD\AJ
Sample : AR1268ICC500
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 31 05:49:49 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP033021.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 31 05:48:47 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.846 min
Delta R.T.: 0.000 min
Response: 4021530
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC500




#1 Tetrachloro-m-xylene

R.T.: 4.259 min
Delta R.T.: 0.000 min
Response: 2306820
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.765 min
Delta R.T.: 0.000 min
Response: 6451606
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.566 min
Delta R.T.: 0.000 min
Response: 2288451
Conc: 50.00 ng/ml



#41 AR-1268-1


R.T.: 9.316 min
Delta R.T.: 0.000 min
Response: 4977154
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC500




#41 AR-1268-1

R.T.: 8.450 min
Delta R.T.: 0.000 min
Response: 2115638
Conc: 500.00 ng/ml




#42 AR-1268-2

R.T.: 9.409 min
Delta R.T.: 0.000 min
Response: 4798548
Conc: 500.00 ng/ml



#42 AR-1268-2


R.T.: 8.515 min
Delta R.T.: 0.000 min
Response: 2048267
Conc: 500.00 ng/ml



#43 AR-1268-3


R.T.: 9.626 min
Delta R.T.: 0.000 min
Response: 4348136
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC500




#43 AR-1268-3

R.T.: 8.721 min
Delta R.T.: 0.000 min
Response: 1771516
Conc: 500.00 ng/ml




#44 AR-1268-4

R.T.: 10.050 min
Delta R.T.: 0.000 min
Response: 1651814
Conc: 500.00 ng/ml



#44 AR-1268-4


R.T.: 9.010 min
Delta R.T.: 0.000 min
Response: 630223
Conc: 500.00 ng/ml



#45 AR-1268-5

R.T.: 10.446 min
Delta R.T.: 0.000 min
Response: 14451868
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC500



#45 AR-1268-5

R.T.: 9.307 min
Delta R.T.: 0.000 min
Response: 5123122
Conc: 500.00 ng/ml