

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP033022\
 Data File : PP045359.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Mar 2022 17:54
 Operator : YP\AJ
 Sample : AR1254ICC500
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 30 19:05:53 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP033022.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 30 19:04:46 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.887	3.146	127.3E6	83784668	50.000	50.000
2)	SA Decachlor...	10.003	8.498	74054122	77960180	50.000	50.000
Target Compounds							
26)	L6 AR-1254-1	6.062	5.139	36008600	46152726	500.000	500.000
27)	L6 AR-1254-2	6.301	5.300	53629130	39400714	500.000	500.000
28)	L6 AR-1254-3	6.700	5.730	54571764	60350724	500.000	500.000
29)	L6 AR-1254-4	7.014	5.979	39269268	40322330	500.000	500.000
30)	L6 AR-1254-5	7.475	6.430	40358289	53987672	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP033022\
Data File : PP045359.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Mar 2022 17:54
Operator : YP\AJ
Sample : AR1254ICC500
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 30 19:05:53 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP033022.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 30 19:04:46 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.887 min
Delta R.T.: 0.000 min
Response: 127275072
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.146 min
Delta R.T.: 0.000 min
Response: 83784668
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.003 min
Delta R.T.: 0.000 min
Response: 74054122
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.498 min
Delta R.T.: 0.000 min
Response: 77960180
Conc: 50.00 ng/ml



#26 AR-1254-1


R.T.: 6.062 min
Delta R.T.: 0.000 min
Response: 36008600
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500




#26 AR-1254-1

R.T.: 5.139 min
Delta R.T.: 0.000 min
Response: 46152726
Conc: 500.00 ng/ml




#27 AR-1254-2

R.T.: 6.301 min
Delta R.T.: 0.000 min
Response: 53629130
Conc: 500.00 ng/ml



#27 AR-1254-2


R.T.: 5.300 min
Delta R.T.: 0.000 min
Response: 39400714
Conc: 500.00 ng/ml



#28 AR-1254-3


R.T.: 6.700 min
Delta R.T.: 0.000 min
Response: 54571764
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500




#28 AR-1254-3

R.T.: 5.730 min
Delta R.T.: 0.000 min
Response: 60350724
Conc: 500.00 ng/ml




#29 AR-1254-4

R.T.: 7.014 min
Delta R.T.: 0.000 min
Response: 39269268
Conc: 500.00 ng/ml



#29 AR-1254-4


R.T.: 5.979 min
Delta R.T.: 0.000 min
Response: 40322330
Conc: 500.00 ng/ml



#30 AR-1254-5

R.T.: 7.475 min
Delta R.T.: 0.000 min
Response: 40358289
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500



#30 AR-1254-5

R.T.: 6.430 min
Delta R.T.: 0.000 min
Response: 53987672
Conc: 500.00 ng/ml