

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP033122\
 Data File : PP045398.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Mar 2022 20:27
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 01 01:18:50 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP033022.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 31 07:54:52 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.886	3.146	137.2E6	89469173	56.846	56.542
2) SA Decachlor...	10.000	8.495	77012754	80163458	53.749	52.863
Target Compounds						
21) L5 AR-1248-1	5.135	4.286	22926840	19433740	500.922	509.838
22) L5 AR-1248-2	5.432	4.538	31063874	26276607	504.927	510.133
23) L5 AR-1248-3	5.650	4.581	36098070	27627592	504.767	513.145
24) L5 AR-1248-4	6.089	4.762	38141665	32458709	507.344	504.453
25) L5 AR-1248-5	6.130	5.181	38217331	33273454	509.236	511.493



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP033122\
Data File : PP045398.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Mar 2022 20:27
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 01 01:18:50 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP033022.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 31 07:54:52 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.886 min
Delta R.T.: -0.001 min
Response: 137157785
Conc: 56.85 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.146 min
Delta R.T.: 0.000 min
Response: 89469173
Conc: 56.54 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.000 min
Delta R.T.: -0.004 min
Response: 77012754
Conc: 53.75 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.495 min
Delta R.T.: -0.002 min
Response: 80163458
Conc: 52.86 ng/ml



#21 AR-1248-1


R.T.: 5.135 min
Delta R.T.: -0.002 min
Response: 22926840
Conc: 500.92 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.286 min
Delta R.T.: 0.000 min
Response: 19433740
Conc: 509.84 ng/ml




#22 AR-1248-2

R.T.: 5.432 min
Delta R.T.: -0.002 min
Response: 31063874
Conc: 504.93 ng/ml



#22 AR-1248-2


R.T.: 4.538 min
Delta R.T.: -0.001 min
Response: 26276607
Conc: 510.13 ng/ml



#23 AR-1248-3


R.T.: 5.650 min
Delta R.T.: -0.002 min
Response: 36098070
Conc: 504.77 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.581 min
Delta R.T.: -0.001 min
Response: 27627592
Conc: 513.14 ng/ml




#24 AR-1248-4

R.T.: 6.089 min
Delta R.T.: -0.002 min
Response: 38141665
Conc: 507.34 ng/ml



#24 AR-1248-4


R.T.: 4.762 min
Delta R.T.: -0.001 min
Response: 32458709
Conc: 504.45 ng/ml



#25 AR-1248-5

R.T.: 6.130 min
Delta R.T.: -0.002 min
Response: 38217331
Conc: 509.24 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.181 min
Delta R.T.: -0.002 min
Response: 33273454
Conc: 511.49 ng/ml