

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040121\
 Data File : PP034563.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 01 Apr 2021 21:57
 Operator : DD\AJ
 Sample : M1854-01MS
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 02 03:00:19 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP033021.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 31 06:02:19 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.844	4.256	1750026	916956	22.941	20.412
2)	SA Decachlor...	10.764	9.562	1629986	557848	21.174	19.757
Target Compounds							
3)	L1 AR-1016-1	6.156	5.509	1189379	550624	505.520	464.142
4)	L1 AR-1016-2	6.179	5.529	1777956	812610	500.591	457.646
5)	L1 AR-1016-3	6.244	5.722	1129211	456719	528.831	488.796
6)	L1 AR-1016-4	6.351	5.770	930913	378326	522.740	520.902
7)	L1 AR-1016-5	6.665	6.000	941362	491270	545.611	537.109
8)	L2 AR-1221-1	5.091	4.510	140806	67834	192.616	149.471
9)	L2 AR-1221-2	5.191	4.611	159904	90806	280.714	260.285
10)	L2 AR-1221-3	5.277	4.700	651543	364879	358.005	347.827
11)	L3 AR-1232-1	5.277	4.700	651543	364879	438.688	431.942
12)	L3 AR-1232-2	5.862	5.529	1035458	812610	1364.503	1079.797
13)	L3 AR-1232-3	6.179	5.722	1777956	456719	1212.661	1150.659
14)	L3 AR-1232-4	6.351	5.816	930913	468356	1295.972	1361.650
15)	L3 AR-1232-5	6.449	6.000	750245	491270	1594.415	1382.070
16)	L4 AR-1242-1	6.156	5.509	1189379	550624	635.450	587.690
17)	L4 AR-1242-2	6.179	5.529	1777956	812610	630.122	571.612
18)	L4 AR-1242-3	6.244	5.722	1129211	456719	663.007	611.697
19)	L4 AR-1242-4	6.351	5.816	930913	468356	665.676	636.540
20)	L4 AR-1242-5	7.131	6.377	115407	371156	78.373	452.651 #
21)	L5 AR-1248-1	6.156	5.509	1189379	550624	836.158	754.772
22)	L5 AR-1248-2	6.449	5.770	750245	378326	379.122	368.558
23)	L5 AR-1248-3	6.665	5.816	941362	468356	396.014	427.335
24)	L5 AR-1248-4	7.091	6.000	140412	491270	55.023	389.240 #
25)	L5 AR-1248-5	7.131	6.421	115407	53090	45.290	46.977
26)	L6 AR-1254-1	7.061	6.377	699798	371156	254.516	201.179
27)	L6 AR-1254-2	7.289	6.534	769087	344775	180.871	215.499
28)	L6 AR-1254-3	7.670	6.971	397859	595329	87.170	238.536 #
29)	L6 AR-1254-4	7.963	7.190	245231	55995	79.315	42.596 #
30)	L6 AR-1254-5	8.393	7.616	2460049	1007216	692.224	495.572 #
31)	L7 AR-1260-1	7.837	7.089	1753459	833749	564.851	542.925
32)	L7 AR-1260-2	8.102	7.283	1956033	924981	520.856	512.059
33)	L7 AR-1260-3	8.467	7.439	1417821	903567	469.054	520.823
34)	L7 AR-1260-4	8.696	7.918	1685859	639870	482.884	446.287
35)	L7 AR-1260-5	9.015	8.162	3124741	1371403	436.319	401.470
36)	L8 AR-1262-1	8.467	7.616	1417821	1007216	309.855	937.818 #
37)	L8 AR-1262-2	9.015	8.162	3124741	1371403	400.489	399.708
38)	L8 AR-1262-3	9.337f	8.446	2031689	305148	377.532	203.653 #
39)	L8 AR-1262-4	9.387f	8.509	1347242	1063007	303.373	385.969 #
40)	L8 AR-1262-5	10.049	9.007	815193	308740	283.091	272.064
41)	L9 AR-1268-1	9.337f	8.446	2031689	305148	202.936	70.625 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040121\
 Data File : PP034563.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 01 Apr 2021 21:57
 Operator : DD\AJ
 Sample : M1854-01MS
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 02 03:00:19 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP033021.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 31 06:02:19 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	9.387f	8.509	1347242	1063007	140.848	256.829 #
43) L9	AR-1268-3	9.626	8.717	74991	54020	8.539	14.848 #
44) L9	AR-1268-4	10.049	9.007	815193	308740	255.196	239.654
45) L9	AR-1268-5	10.445	9.303	305368	109478	10.694	10.541



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040121\
Data File : PP034563.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 01 Apr 2021 21:57
Operator : DD\AJ
Sample : M1854-01MS
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 02 03:00:19 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP033021.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 31 06:02:19 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.844 min
Delta R.T.: -0.002 min
Response: 1750026
Conc: 22.94 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.256 min
Delta R.T.: -0.004 min
Response: 916956
Conc: 20.41 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.764 min
Delta R.T.: 0.000 min
Response: 1629986
Conc: 21.17 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.562 min
Delta R.T.: -0.004 min
Response: 557848
Conc: 19.76 ng/ml



#3 AR-1016-1


R.T.: 6.156 min
Delta R.T.: -0.001 min
Response: 1189379
Conc: 505.52 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 5.509 min
Delta R.T.: -0.003 min
Response: 550624
Conc: 464.14 ng/ml




#4 AR-1016-2

R.T.: 6.179 min
Delta R.T.: -0.002 min
Response: 1777956
Conc: 500.59 ng/ml



#4 AR-1016-2


R.T.: 5.529 min
Delta R.T.: -0.003 min
Response: 812610
Conc: 457.65 ng/ml



#5 AR-1016-3


R.T.: 6.244 min
Delta R.T.: -0.001 min
Response: 1129211
Conc: 528.83 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 5.722 min
Delta R.T.: -0.003 min
Response: 456719
Conc: 488.80 ng/ml




#6 AR-1016-4

R.T.: 6.351 min
Delta R.T.: -0.002 min
Response: 930913
Conc: 522.74 ng/ml



#6 AR-1016-4


R.T.: 5.770 min
Delta R.T.: -0.004 min
Response: 378326
Conc: 520.90 ng/ml



#7 AR-1016-5


R.T.: 6.665 min
Delta R.T.: -0.001 min
Response: 941362
Conc: 545.61 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 6.000 min
Delta R.T.: -0.003 min
Response: 491270
Conc: 537.11 ng/ml




#8 AR-1221-1

R.T.: 5.091 min
Delta R.T.: -0.004 min
Response: 140806
Conc: 192.62 ng/ml



#8 AR-1221-1


R.T.: 4.510 min
Delta R.T.: -0.004 min
Response: 67834
Conc: 149.47 ng/ml



#9 AR-1221-2


R.T.: 5.191 min
Delta R.T.: -0.002 min
Response: 159904
Conc: 280.71 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 4.611 min
Delta R.T.: -0.003 min
Response: 90806
Conc: 260.28 ng/ml




#10 AR-1221-3

R.T.: 5.277 min
Delta R.T.: -0.002 min
Response: 651543
Conc: 358.01 ng/ml



#10 AR-1221-3


R.T.: 4.700 min
Delta R.T.: -0.003 min
Response: 364879
Conc: 347.83 ng/ml



#11 AR-1232-1


R.T.: 5.277 min
Delta R.T.: -0.001 min
Response: 651543
Conc: 438.69 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 4.700 min
Delta R.T.: -0.004 min
Response: 364879
Conc: 431.94 ng/ml




#12 AR-1232-2

R.T.: 5.862 min
Delta R.T.: -0.002 min
Response: 1035458
Conc: 1364.50 ng/ml



#12 AR-1232-2


R.T.: 5.529 min
Delta R.T.: -0.003 min
Response: 812610
Conc: 1079.80 ng/ml



#13 AR-1232-3


R.T.: 6.179 min
Delta R.T.: -0.002 min
Response: 1777956
Conc: 1212.66 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 5.722 min
Delta R.T.: -0.002 min
Response: 456719
Conc: 1150.66 ng/ml




#14 AR-1232-4

R.T.: 6.351 min
Delta R.T.: -0.002 min
Response: 930913
Conc: 1295.97 ng/ml



#14 AR-1232-4


R.T.: 5.816 min
Delta R.T.: -0.003 min
Response: 468356
Conc: 1361.65 ng/ml



#15 AR-1232-5


R.T.: 6.449 min
Delta R.T.: -0.001 min
Response: 750245
Conc: 1594.42 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 6.000 min
Delta R.T.: -0.003 min
Response: 491270
Conc: 1382.07 ng/ml




#16 AR-1242-1

R.T.: 6.156 min
Delta R.T.: -0.002 min
Response: 1189379
Conc: 635.45 ng/ml



#16 AR-1242-1


R.T.: 5.509 min
Delta R.T.: -0.004 min
Response: 550624
Conc: 587.69 ng/ml



#17 AR-1242-2


R.T.: 6.179 min
Delta R.T.: -0.003 min
Response: 1777956
Conc: 630.12 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 5.529 min
Delta R.T.: -0.003 min
Response: 812610
Conc: 571.61 ng/ml




#18 AR-1242-3

R.T.: 6.244 min
Delta R.T.: -0.003 min
Response: 1129211
Conc: 663.01 ng/ml



#18 AR-1242-3


R.T.: 5.722 min
Delta R.T.: -0.003 min
Response: 456719
Conc: 611.70 ng/ml



#19 AR-1242-4


R.T.: 6.351 min
Delta R.T.: -0.003 min
Response: 930913
Conc: 665.68 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 5.816 min
Delta R.T.: -0.004 min
Response: 468356
Conc: 636.54 ng/ml




#20 AR-1242-5

R.T.: 7.131 min
Delta R.T.: -0.003 min
Response: 115407
Conc: 78.37 ng/ml



#20 AR-1242-5


R.T.: 6.377 min
Delta R.T.: -0.004 min
Response: 371156
Conc: 452.65 ng/ml



#21 AR-1248-1


R.T.: 6.156 min
Delta R.T.: -0.002 min
Response: 1189379
Conc: 836.16 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 5.509 min
Delta R.T.: -0.004 min
Response: 550624
Conc: 754.77 ng/ml




#22 AR-1248-2

R.T.: 6.449 min
Delta R.T.: -0.002 min
Response: 750245
Conc: 379.12 ng/ml



#22 AR-1248-2


R.T.: 5.770 min
Delta R.T.: -0.004 min
Response: 378326
Conc: 368.56 ng/ml



#23 AR-1248-3


R.T.: 6.665 min
Delta R.T.: -0.001 min
Response: 941362
Conc: 396.01 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 5.816 min
Delta R.T.: -0.003 min
Response: 468356
Conc: 427.34 ng/ml




#24 AR-1248-4

R.T.: 7.091 min
Delta R.T.: -0.003 min
Response: 140412
Conc: 55.02 ng/ml



#24 AR-1248-4


R.T.: 6.000 min
Delta R.T.: -0.004 min
Response: 491270
Conc: 389.24 ng/ml



#25 AR-1248-5


R.T.: 7.131 min
Delta R.T.: -0.002 min
Response: 115407
Conc: 45.29 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 6.421 min
Delta R.T.: -0.004 min
Response: 53090
Conc: 46.98 ng/ml




#26 AR-1254-1

R.T.: 7.061 min
Delta R.T.: -0.002 min
Response: 699798
Conc: 254.52 ng/ml



#26 AR-1254-1


R.T.: 6.377 min
Delta R.T.: -0.004 min
Response: 371156
Conc: 201.18 ng/ml



#27 AR-1254-2


R.T.: 7.289 min
Delta R.T.: -0.004 min
Response: 769087
Conc: 180.87 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 6.534 min
Delta R.T.: -0.004 min
Response: 344775
Conc: 215.50 ng/ml




#28 AR-1254-3

R.T.: 7.670 min
Delta R.T.: -0.003 min
Response: 397859
Conc: 87.17 ng/ml



#28 AR-1254-3


R.T.: 6.971 min
Delta R.T.: 0.012 min
Response: 595329
Conc: 238.54 ng/ml



#29 AR-1254-4


R.T.: 7.963 min
Delta R.T.: -0.005 min
Response: 245231
Conc: 79.32 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 7.190 min
Delta R.T.: -0.004 min
Response: 55995
Conc: 42.60 ng/ml




#30 AR-1254-5

R.T.: 8.393 min
Delta R.T.: -0.002 min
Response: 2460049
Conc: 692.22 ng/ml



#30 AR-1254-5


R.T.: 7.616 min
Delta R.T.: -0.005 min
Response: 1007216
Conc: 495.57 ng/ml



#31 AR-1260-1


R.T.: 7.837 min
Delta R.T.: 0.000 min
Response: 1753459
Conc: 564.85 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 7.089 min
Delta R.T.: -0.004 min
Response: 833749
Conc: 542.93 ng/ml




#32 AR-1260-2

R.T.: 8.102 min
Delta R.T.: 0.000 min
Response: 1956033
Conc: 520.86 ng/ml



#32 AR-1260-2


R.T.: 7.283 min
Delta R.T.: -0.003 min
Response: 924981
Conc: 512.06 ng/ml



#33 AR-1260-3


R.T.: 8.467 min
Delta R.T.: 0.000 min
Response: 1417821
Conc: 469.05 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 7.439 min
Delta R.T.: -0.004 min
Response: 903567
Conc: 520.82 ng/ml




#34 AR-1260-4

R.T.: 8.696 min
Delta R.T.: 0.000 min
Response: 1685859
Conc: 482.88 ng/ml



#34 AR-1260-4


R.T.: 7.918 min
Delta R.T.: -0.004 min
Response: 639870
Conc: 446.29 ng/ml



#35 AR-1260-5


R.T.: 9.015 min
Delta R.T.: 0.000 min
Response: 3124741
Conc: 436.32 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 8.162 min
Delta R.T.: -0.004 min
Response: 1371403
Conc: 401.47 ng/ml




#36 AR-1262-1

R.T.: 8.467 min
Delta R.T.: -0.001 min
Response: 1417821
Conc: 309.86 ng/ml



#36 AR-1262-1


R.T.: 7.616 min
Delta R.T.: -0.004 min
Response: 1007216
Conc: 937.82 ng/ml



#37 AR-1262-2


R.T.: 9.015 min
Delta R.T.: -0.002 min
Response: 3124741
Conc: 400.49 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 8.162 min
Delta R.T.: -0.004 min
Response: 1371403
Conc: 399.71 ng/ml




#38 AR-1262-3

R.T.: 9.337 min
Delta R.T.: 0.019 min
Response: 2031689
Conc: 377.53 ng/ml



#38 AR-1262-3


R.T.: 8.446 min
Delta R.T.: -0.004 min
Response: 305148
Conc: 203.65 ng/ml



#39 AR-1262-4


R.T.: 9.387 min
Delta R.T.: -0.021 min
Response: 1347242
Conc: 303.37 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 8.509 min
Delta R.T.: -0.005 min
Response: 1063007
Conc: 385.97 ng/ml




#40 AR-1262-5

R.T.: 10.049 min
Delta R.T.: 0.000 min
Response: 815193
Conc: 283.09 ng/ml



#40 AR-1262-5


R.T.: 9.007 min
Delta R.T.: -0.003 min
Response: 308740
Conc: 272.06 ng/ml



#41 AR-1268-1


R.T.: 9.337 min
Delta R.T.: 0.021 min
Response: 2031689
Conc: 202.94 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 8.446 min
Delta R.T.: -0.003 min
Response: 305148
Conc: 70.62 ng/ml




#42 AR-1268-2

R.T.: 9.387 min
Delta R.T.: -0.022 min
Response: 1347242
Conc: 140.85 ng/ml



#42 AR-1268-2


R.T.: 8.509 min
Delta R.T.: -0.005 min
Response: 1063007
Conc: 256.83 ng/ml



#43 AR-1268-3


R.T.: 9.626 min
Delta R.T.: 0.000 min
Response: 74991
Conc: 8.54 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 8.717 min
Delta R.T.: -0.004 min
Response: 54020
Conc: 14.85 ng/ml




#44 AR-1268-4

R.T.: 10.049 min
Delta R.T.: 0.000 min
Response: 815193
Conc: 255.20 ng/ml



#44 AR-1268-4


R.T.: 9.007 min
Delta R.T.: -0.002 min
Response: 308740
Conc: 239.65 ng/ml



#45 AR-1268-5

R.T.: 10.445 min
Delta R.T.: 0.000 min
Response: 305368
Conc: 10.69 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 9.303 min
Delta R.T.: -0.003 min
Response: 109478
Conc: 10.54 ng/ml