Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040121\
Data File : PP034552.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 01 Apr 2021 18:51
Operator : DD\AJ
Sample : PB135449BSD
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PB135449BSD

Manual Integrations APPROVED
Ankita
4/2/2021 11:29:28 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 02 02:51:31 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP033021.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 31 06:02:19 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.845	4.257	1696599	884496	22.240	19.689
2)	SA Decachlor...	10.766	9.563	1763592	603711	22.910	21.381
	Target Compounds						
3)	L1 AR-1016-1	6.157	5.510	1162374	529464	494.042	446.305
4)	L1 AR-1016-2	6.180	5.531	1767831	787545	497.741	443.529
5)	L1 AR-1016-3	6.246	5.723	1134933	436690	531.511	467.360
6)	L1 AR-1016-4	6.352	5.771	918728	362080	515.897	498.534
7)	L1 AR-1016-5	6.666	6.002	912356	411214	528.799	449.583
31)	L7 AR-1260-1	7.837	7.089	1756568	823495	565.853m	536.248m
32)	L7 AR-1260-2	8.102	7.283	1995367	975447	531.330m	539.997m
33)	L7 AR-1260-3	8.468	7.440	1536231	900806	508.228	519.232m
34)	L7 AR-1260-4	8.698	7.919	1756102	639631	503.004	446.121
35)	L7 AR-1260-5	9.015	8.163	3233473	1416219	451.502	414.590

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040121\
Data File : PP034552.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 01 Apr 2021 18:51
Operator : DD\AJ
Sample : PB135449BSD
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PB135449BSD

Manual Integrations APPROVED
Ankita
4/2/2021 11:29:28 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 02 02:51:31 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP033021.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 31 06:02:19 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm