

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040121\
 Data File : PP034586.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 02 Apr 2021 5:32
 Operator : DD\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 02 06:47:13 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP033021.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 31 06:02:19 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.843	4.255	4237835	2335176	55.553	51.983
2)	SA Decachlor...	10.763	9.560	4114343	1408588	53.447	49.886
Target Compounds							
16)	L4 AR-1242-1	6.155	5.508	923132	438556	493.202	468.078
17)	L4 AR-1242-2	6.178	5.528	1385775	630304	491.130	443.373
18)	L4 AR-1242-3	6.244	5.721	891334	353521	523.339	473.481
19)	L4 AR-1242-4	6.351	5.814	724595	368272	518.142	500.517
20)	L4 AR-1242-5	7.130	6.376	744549	395831	505.625	482.744



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040121\
Data File : PP034586.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 02 Apr 2021 5:32
Operator : DD\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 02 06:47:13 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP033021.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 31 06:02:19 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.843 min
Delta R.T.: -0.003 min
Response: 4237835
Conc: 55.55 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.255 min
Delta R.T.: -0.004 min
Response: 2335176
Conc: 51.98 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.763 min
Delta R.T.: -0.002 min
Response: 4114343
Conc: 53.45 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.560 min
Delta R.T.: -0.006 min
Response: 1408588
Conc: 49.89 ng/ml



#16 AR-1242-1


R.T.: 6.155 min
Delta R.T.: -0.003 min
Response: 923132
Conc: 493.20 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 5.508 min
Delta R.T.: -0.005 min
Response: 438556
Conc: 468.08 ng/ml




#17 AR-1242-2

R.T.: 6.178 min
Delta R.T.: -0.004 min
Response: 1385775
Conc: 491.13 ng/ml



#17 AR-1242-2


R.T.: 5.528 min
Delta R.T.: -0.004 min
Response: 630304
Conc: 443.37 ng/ml



#18 AR-1242-3


R.T.: 6.244 min
Delta R.T.: -0.003 min
Response: 891334
Conc: 523.34 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 5.721 min
Delta R.T.: -0.005 min
Response: 353521
Conc: 473.48 ng/ml




#19 AR-1242-4

R.T.: 6.351 min
Delta R.T.: -0.003 min
Response: 724595
Conc: 518.14 ng/ml



#19 AR-1242-4


R.T.: 5.814 min
Delta R.T.: -0.005 min
Response: 368272
Conc: 500.52 ng/ml



#20 AR-1242-5

R.T.: 7.130 min
Delta R.T.: -0.003 min
Response: 744549
Conc: 505.63 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 6.376 min
Delta R.T.: -0.005 min
Response: 395831
Conc: 482.74 ng/ml