

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040124\
 Data File : PP063980.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Apr 2024 11:21
 Operator : YP\AJ
 Sample : P1921-05DL 50X
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 S-20DL

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/02/2024
 Supervised By :Ankita Jodhani 04/02/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 01 13:43:11 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031924.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Mar 19 06:23:02 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds							
16)	L4	AR-1242-1	5.669	4.728	68650918	62989835	1138.163 1068.533
17)	L4	AR-1242-2	5.692	4.747	93047501	88056435	1083.560 1076.005
18)	L4	AR-1242-3	5.754	4.923	58887078	41177378	1047.681 917.454
19)	L4	AR-1242-4	5.854	5.008	44768783	33230282	994.973 669.509 #
20)	L4	AR-1242-5	6.597	5.533	15933141	20469485	355.778 350.754
31)	L7	AR-1260-1	7.287	6.218	6145305	6607366	48.656 46.235
32)	L7	AR-1260-2	7.545	6.407	6885349	6463902	48.230m 38.238
33)	L7	AR-1260-3	7.908	6.560	3525440	5372652	35.123 33.427
34)	L7	AR-1260-4	8.135	7.035	3525260	3270776	29.304m 26.834
35)	L7	AR-1260-5	8.466	7.277	5354359	6043132	25.793 22.650

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040124\
Data File : PP063980.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Apr 2024 11:21
Operator : YP\AJ
Sample : P1921-05DL 50X
Misc :
ALS Vial : 12 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
S-20DL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/02/2024
Supervised By :Ankita Jodhani 04/02/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 01 13:43:11 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031924.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 19 06:23:02 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T. : 4.489 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :
S-20DL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/02/2024
Supervised By :Ankita Jodhani 04/02/2024




#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T. : 3.639 min
Response: 0
Conc: N.D.




#2 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 10.365 min
Response: 0
Conc: N.D.



#2 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 8.684 min
Response: 0
Conc: N.D.




#16 AR-1242-1

R.T.: 5.669 min
Delta R.T.: -0.005 min
Response: 68650918
Conc: 1138.16 ng/ml

Instrument :
ECD_P
ClientSampleId :
S-20DL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/02/2024
Supervised By :Ankita Jodhani 04/02/2024




#16 AR-1242-1

R.T.: 4.728 min
Delta R.T.: -0.005 min
Response: 62989835
Conc: 1068.53 ng/ml




#17 AR-1242-2

R.T.: 5.692 min
Delta R.T.: -0.005 min
Response: 93047501
Conc: 1083.56 ng/ml



#17 AR-1242-2

R.T.: 4.747 min
Delta R.T.: -0.005 min
Response: 88056435
Conc: 1076.00 ng/ml




#18 AR-1242-3

R.T.: 5.754 min
Delta R.T.: -0.006 min
Response: 58887078
Conc: 1047.68 ng/ml

Instrument :
ECD_P
ClientSampleId :
S-20DL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/02/2024
Supervised By :Ankita Jodhani 04/02/2024




#18 AR-1242-3

R.T.: 4.923 min
Delta R.T.: -0.006 min
Response: 41177378
Conc: 917.45 ng/ml




#19 AR-1242-4

R.T.: 5.854 min
Delta R.T.: -0.005 min
Response: 44768783
Conc: 994.97 ng/ml



#19 AR-1242-4

R.T.: 5.008 min
Delta R.T.: -0.005 min
Response: 33230282
Conc: 669.51 ng/ml




#20 AR-1242-5

R.T.: 6.597 min
Delta R.T.: -0.006 min
Response: 15933141
Conc: 355.78 ng/ml

Instrument :
ECD_P
ClientSampleId :
S-20DL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/02/2024
Supervised By :Ankita Jodhani 04/02/2024




#20 AR-1242-5

R.T.: 5.533 min
Delta R.T.: -0.007 min
Response: 20469485
Conc: 350.75 ng/ml




#31 AR-1260-1

R.T.: 7.287 min
Delta R.T.: -0.006 min
Response: 6145305
Conc: 48.66 ng/ml



#31 AR-1260-1

R.T.: 6.218 min
Delta R.T.: -0.008 min
Response: 6607366
Conc: 46.24 ng/ml




#32 AR-1260-2

R.T.: 7.545 min
Delta R.T.: -0.007 min
Response: 6885349
Conc: 48.23 ng/ml

Instrument :
ECD_P
ClientSampleId :
S-20DL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/02/2024
Supervised By :Ankita Jodhani 04/02/2024




#32 AR-1260-2

R.T.: 6.407 min
Delta R.T.: -0.008 min
Response: 6463902
Conc: 38.24 ng/ml




#33 AR-1260-3

R.T.: 7.908 min
Delta R.T.: -0.007 min
Response: 3525440
Conc: 35.12 ng/ml



#33 AR-1260-3

R.T.: 6.560 min
Delta R.T.: -0.009 min
Response: 5372652
Conc: 33.43 ng/ml




#34 AR-1260-4

R.T.: 8.135 min
Delta R.T.: -0.009 min
Response: 3525260
Conc: 29.30 ng/ml

Instrument :
ECD_P
ClientSampleId :
S-20DL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/02/2024
Supervised By :Ankita Jodhani 04/02/2024




#34 AR-1260-4

R.T.: 7.035 min
Delta R.T.: -0.010 min
Response: 3270776
Conc: 26.83 ng/ml



#35 AR-1260-5

R.T.: 8.466 min
Delta R.T.: -0.008 min
Response: 5354359
Conc: 25.79 ng/ml



#35 AR-1260-5

R.T.: 7.277 min
Delta R.T.: -0.009 min
Response: 6043132
Conc: 22.65 ng/ml