

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040224\
 Data File : PP064016.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Apr 2024 09:50
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 02 16:36:06 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031924.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Mar 19 06:23:02 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.486	3.637	112.6E6	117.1E6	47.201	49.057
2) SA Decachlor...	10.354	8.669	125.4E6	114.5E6	49.409	42.376
Target Compounds						
21) L5 AR-1248-1	5.668	4.727	24369498	23217890	521.627	521.259
22) L5 AR-1248-2	5.944	4.965	38065566	34835961	524.478	490.829
23) L5 AR-1248-3	6.150	5.006	41947074	37413985	519.505	505.593
24) L5 AR-1248-4	6.557	5.178	44238221	41798307	540.977	482.703
25) L5 AR-1248-5	6.597	5.573	45257373	40052488	554.920	533.480



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP040224\
Data File : PP064016.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Apr 2024 09:50
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 02 16:36:06 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP031924.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 19 06:23:02 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.486 min
Delta R.T.: -0.003 min
Response: 112629264
Conc: 47.20 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.637 min
Delta R.T.: -0.003 min
Response: 117125848
Conc: 49.06 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.354 min
Delta R.T.: -0.011 min
Response: 125449157
Conc: 49.41 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.669 min
Delta R.T.: -0.014 min
Response: 114520383
Conc: 42.38 ng/ml



#21 AR-1248-1


R.T.: 5.668 min
Delta R.T.: -0.005 min
Response: 24369498
Conc: 521.63 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.727 min
Delta R.T.: -0.005 min
Response: 23217890
Conc: 521.26 ng/ml




#22 AR-1248-2

R.T.: 5.944 min
Delta R.T.: -0.006 min
Response: 38065566
Conc: 524.48 ng/ml



#22 AR-1248-2


R.T.: 4.965 min
Delta R.T.: -0.006 min
Response: 34835961
Conc: 490.83 ng/ml



#23 AR-1248-3


R.T.: 6.150 min
Delta R.T.: -0.006 min
Response: 41947074
Conc: 519.51 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.006 min
Delta R.T.: -0.007 min
Response: 37413985
Conc: 505.59 ng/ml




#24 AR-1248-4

R.T.: 6.557 min
Delta R.T.: -0.006 min
Response: 44238221
Conc: 540.98 ng/ml



#24 AR-1248-4


R.T.: 5.178 min
Delta R.T.: -0.007 min
Response: 41798307
Conc: 482.70 ng/ml



#25 AR-1248-5

R.T.: 6.597 min
Delta R.T.: -0.005 min
Response: 45257373
Conc: 554.92 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.573 min
Delta R.T.: -0.007 min
Response: 40052488
Conc: 533.48 ng/ml